



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
EPA Sample No.: SSTDICCC040 Date Analyzed: 01/31/2023
Lab File ID: BP013673.D Time Analyzed: 13:38
Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	222064	8.14	832306	10.97	549003	14.78
UPPER LIMIT	444128	8.64	1664610	11.469	1098010	15.281
LOWER LIMIT	111032	7.64	416153	10.469	274502	14.281
EPA SAMPLE NO.						
01 PB150378BS	297815	8.14	1192600	10.97	801024	14.78

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.