

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011923\
 Data File : BF132148.D
 Acq On : 19 Jan 2023 15:44
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF011923

Quant Time: Jan 20 00:55:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 00:50:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	39.301	1.7	95	0.00
3	Pyridine	40.000	39.841	0.4	95	0.00
4	n-Nitrosodimethylamine	40.000	40.425	-1.1	98	0.00
5 S	2-Fluorophenol	80.000	79.645	0.4	98	0.00
6	Aniline	40.000	39.256	1.9	97	0.00
7 S	Phenol-d6	80.000	79.486	0.6	98	0.00
8	2-Chlorophenol	40.000	40.571	-1.4	99	0.00
9	Benzaldehyde	40.000	40.632	-1.6	100	0.00
10 C	Phenol	40.000	39.701	0.7	98	0.00
11	bis(2-Chloroethyl)ether	40.000	39.259	1.9	97	0.00
12	1,3-Dichlorobenzene	40.000	39.552	1.1	98	0.00
13 C	1,4-Dichlorobenzene	40.000	39.996	0.0	98	0.00
14	1,2-Dichlorobenzene	40.000	39.349	1.6	98	0.00
15	Benzyl Alcohol	40.000	41.551	-3.9	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.664	0.8	97	0.00
17	2-Methylphenol	40.000	40.067	-0.2	98	0.00
18	Hexachloroethane	40.000	41.373	-3.4	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.753	3.1	97	0.00
20	3+4-Methylphenols	40.000	40.648	-1.6	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	38.155	4.6	97	0.00
23 S	Nitrobenzene-d5	80.000	82.897	-3.6	100	0.00
24	Nitrobenzene	40.000	41.232	-3.1	99	0.00
25	Isophorone	40.000	39.224	1.9	98	0.00
26 C	2-Nitrophenol	40.000	44.494	-11.2	111	0.00
27	2,4-Dimethylphenol	40.000	40.276	-0.7	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.240	1.9	99	0.00
29 C	2,4-Dichlorophenol	40.000	40.736	-1.8	98	0.00
30	1,2,4-Trichlorobenzene	40.000	39.345	1.6	98	0.00
31	Naphthalene	40.000	38.858	2.9	98	0.00
32	Benzoic acid	40.000	42.386	-6.0	115	0.00
33	4-Chloroaniline	40.000	39.292	1.8	97	0.00
34 C	Hexachlorobutadiene	40.000	40.159	-0.4	100	0.00
35	Caprolactam	40.000	41.792	-4.5	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.638	-1.6	98	0.00
37	2-Methylnaphthalene	40.000	38.851	2.9	98	0.00
38	1-Methylnaphthalene	40.000	38.619	3.5	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.204	2.0	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.778	-6.9	102	0.00
42 S	2,4,6-Tribromophenol	80.000	83.991	-5.0	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.471	-3.7	98	0.00
44	2,4,5-Trichlorophenol	40.000	41.656	-4.1	99	0.00
45 S	2-Fluorobiphenyl	80.000	73.696	7.9	99	0.00
46	1,1'-Biphenyl	40.000	39.026	2.4	98	0.00
47	2-Chloronaphthalene	40.000	38.849	2.9	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011923\
 Data File : BF132148.D
 Acq On : 19 Jan 2023 15:44
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF011923

Quant Time: Jan 20 00:55:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 00:50:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.882	-2.2	102	0.00
49	Acenaphthylene	40.000	39.175	2.1	98	0.00
50	Dimethylphthalate	40.000	39.729	0.7	98	0.00
51	2,6-Dinitrotoluene	40.000	44.109	-10.3	101	0.00
52 C	Acenaphthene	40.000	38.803	3.0	98	0.00
53	3-Nitroaniline	40.000	43.470	-8.7	100	0.00
54 P	2,4-Dinitrophenol	40.000	43.353	-8.4	117	0.00
55	Dibenzofuran	40.000	38.579	3.6	97	0.00
56 P	4-Nitrophenol	40.000	44.675	-11.7	103	0.00
57	2,4-Dinitrotoluene	40.000	45.732	-14.3	107	0.00
58	Fluorene	40.000	38.193	4.5	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.074	-5.2	98	0.00
60	Diethylphthalate	40.000	39.637	0.9	97	0.00
61	4-Chlorophenyl-phenylether	40.000	38.934	2.7	98	0.00
62	4-Nitroaniline	40.000	40.810	-2.0	99	0.00
63	Azobenzene	40.000	38.693	3.3	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.985	-7.5	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.352	1.6	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.565	1.1	96	0.00
68	Hexachlorobenzene	40.000	39.667	0.8	97	0.00
69	Atrazine	40.000	40.505	-1.3	96	0.00
70 C	Pentachlorophenol	40.000	44.053	-10.1	100	0.00
71	Phenanthrene	40.000	38.938	2.7	97	0.00
72	Anthracene	40.000	38.898	2.8	97	0.00
73	Carbazole	40.000	38.127	4.7	95	0.00
74	Di-n-butylphthalate	40.000	40.307	-0.8	97	0.00
75 C	Fluoranthene	40.000	37.600	6.0	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	40.675	-1.7	83	0.00
78	Pyrene	40.000	40.977	-2.4	95	0.00
79 S	Terphenyl-d14	80.000	79.705	0.4	95	0.00
80	Butylbenzylphthalate	40.000	41.476	-3.7	98	0.00
81	Benzo(a)anthracene	40.000	39.867	0.3	94	0.00
82	3,3'-Dichlorobenzidine	40.000	44.466	-11.2	97	0.00
83	Chrysene	40.000	39.446	1.4	93	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.877	-2.2	98	0.00
85 c	Di-n-octyl phthalate	40.000	42.292	-5.7	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.080	-2.7	101	0.00
88	Benzo(b)fluoranthene	40.000	41.092	-2.7	100	0.00
89	Benzo(k)fluoranthene	40.000	38.618	3.5	97	0.00
90 C	Benzo(a)pyrene	40.000	41.276	-3.2	99	0.00
91	Dibenzo(a,h)anthracene	40.000	44.798	-12.0	101	0.00
92	Benzo(g,h,i)perylene	40.000	40.647	-1.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011923\
Data File : BF132148.D
Acq On : 19 Jan 2023 15:44
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF011923

Quant Time: Jan 20 00:55:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 00:50:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0