

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122722\
 Data File : BG056134.D
 Acq On : 27 Dec 2022 20:24
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG122722

Quant Time: Dec 28 05:22:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 05:20:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	39.828	0.4	92	0.00
3	Pyridine	40.000	41.978	-4.9	99	0.00
4	n-Nitrosodimethylamine	40.000	42.257	-5.6	98	0.00
5 S	2-Fluorophenol	80.000	80.566	-0.7	98	0.00
6	Aniline	40.000	40.965	-2.4	98	0.00
7 S	Phenol-d6	80.000	82.973	-3.7	97	0.00
8	2-Chlorophenol	40.000	41.713	-4.3	100	0.00
9	Benzaldehyde	40.000	40.869	-2.2	102	0.00
10 C	Phenol	40.000	41.453	-3.6	98	0.00
11	bis(2-Chloroethyl)ether	40.000	39.116	2.2	94	0.00
12	1,3-Dichlorobenzene	40.000	40.105	-0.3	98	0.00
13 C	1,4-Dichlorobenzene	40.000	41.260	-3.1	99	0.00
14	1,2-Dichlorobenzene	40.000	40.147	-0.4	97	0.00
15	Benzyl Alcohol	40.000	40.968	-2.4	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.611	3.5	94	0.00
17	2-Methylphenol	40.000	42.001	-5.0	99	0.00
18	Hexachloroethane	40.000	40.766	-1.9	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.703	-1.8	98	0.00
20	3+4-Methylphenols	40.000	41.931	-4.8	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	40.484	-1.2	97	0.00
23 S	Nitrobenzene-d5	80.000	80.376	-0.5	97	0.00
24	Nitrobenzene	40.000	40.357	-0.9	98	0.00
25	Isophorone	40.000	39.682	0.8	96	0.00
26 C	2-Nitrophenol	40.000	41.974	-4.9	102	0.00
27	2,4-Dimethylphenol	40.000	40.598	-1.5	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.971	0.1	98	0.00
29 C	2,4-Dichlorophenol	40.000	41.547	-3.9	97	0.00
30	1,2,4-Trichlorobenzene	40.000	41.538	-3.8	101	0.00
31	Naphthalene	40.000	41.269	-3.2	101	0.00
32	Benzoic acid	40.000	41.809	-4.5	97	0.00
33	4-Chloroaniline	40.000	40.590	-1.5	98	0.00
34 C	Hexachlorobutadiene	40.000	40.379	-0.9	96	0.00
35	Caprolactam	40.000	42.755	-6.9	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.164	-5.4	103	0.00
37	2-Methylnaphthalene	40.000	40.605	-1.5	98	0.00
38	1-Methylnaphthalene	40.000	40.438	-1.1	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.888	-2.2	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.714	-4.3	97	0.00
42 S	2,4,6-Tribromophenol	80.000	83.440	-4.3	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.419	-3.5	99	0.00
44	2,4,5-Trichlorophenol	40.000	41.556	-3.9	102	0.00
45 S	2-Fluorobiphenyl	80.000	80.357	-0.4	99	0.00
46	1,1'-Biphenyl	40.000	40.745	-1.9	100	0.00
47	2-Chloronaphthalene	40.000	40.951	-2.4	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122722\
 Data File : BG056134.D
 Acq On : 27 Dec 2022 20:24
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG122722

Quant Time: Dec 28 05:22:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 05:20:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.149	-5.4	103	0.00
49	Acenaphthylene	40.000	40.686	-1.7	100	0.00
50	Dimethylphthalate	40.000	40.982	-2.5	101	0.00
51	2,6-Dinitrotoluene	40.000	41.114	-2.8	103	0.00
52 C	Acenaphthene	40.000	41.654	-4.1	101	0.00
53	3-Nitroaniline	40.000	42.042	-5.1	103	0.00
54 P	2,4-Dinitrophenol	40.000	43.334	-8.3	102	0.00
55	Dibenzofuran	40.000	41.180	-2.9	100	0.00
56 P	4-Nitrophenol	40.000	42.310	-5.8	102	0.00
57	2,4-Dinitrotoluene	40.000	42.213	-5.5	102	0.00
58	Fluorene	40.000	40.509	-1.3	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.844	-4.6	103	0.00
60	Diethylphthalate	40.000	41.116	-2.8	101	0.00
61	4-Chlorophenyl-phenylether	40.000	40.928	-2.3	100	0.00
62	4-Nitroaniline	40.000	41.527	-3.8	100	0.00
63	Azobenzene	40.000	40.042	-0.1	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.663	-6.7	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.146	-0.4	102	0.00
67	4-Bromophenyl-phenylether	40.000	39.928	0.2	101	0.00
68	Hexachlorobenzene	40.000	40.388	-1.0	102	0.00
69	Atrazine	40.000	40.223	-0.6	100	0.00
70 C	Pentachlorophenol	40.000	41.702	-4.3	103	0.00
71	Phenanthrene	40.000	40.305	-0.8	102	0.00
72	Anthracene	40.000	40.446	-1.1	102	0.00
73	Carbazole	40.000	40.029	-0.1	101	0.00
74	Di-n-butylphthalate	40.000	41.078	-2.7	103	0.00
75 C	Fluoranthene	40.000	40.173	-0.4	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzidine	40.000	38.970	2.6	95	0.00
78	Pyrene	40.000	39.693	0.8	102	0.00
79 S	Terphenyl-d14	80.000	78.859	1.4	101	0.00
80	Butylbenzylphthalate	40.000	40.426	-1.1	102	0.00
81	Benzo(a)anthracene	40.000	40.053	-0.1	101	0.00
82	3,3'-Dichlorobenzidine	40.000	40.280	-0.7	99	0.00
83	Chrysene	40.000	40.269	-0.7	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.444	1.4	99	0.00
85 c	Di-n-octyl phthalate	40.000	40.429	-1.1	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.495	-1.2	101	-0.01
88	Benzo(b)fluoranthene	40.000	40.380	-1.0	100	0.00
89	Benzo(k)fluoranthene	40.000	40.524	-1.3	101	0.00
90 C	Benzo(a)pyrene	40.000	40.372	-0.9	101	0.00
91	Dibenzo(a,h)anthracene	40.000	40.291	-0.7	101	0.00
92	Benzo(g,h,i)perylene	40.000	40.151	-0.4	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122722\
Data File : BG056134.D
Acq On : 27 Dec 2022 20:24
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG122722

Quant Time: Dec 28 05:22:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 05:20:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0