



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: louie01

Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG No.: 01233

Instrument ID: BNA_F Calibration Date/Time: 01/20/2023 16:40

Lab File ID: BF132164.D Init. Calib. Date(s): 01/19/2023 01/19/2023

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:21 15:11

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.145	1.115		-2.6	
Benzaldehyde	0.927	0.917		-1.1	
Phenol-d6	1.441	1.412		-2.0	
Phenol	1.500	1.494		-0.4	20.0
bis(2-Chloroethyl) ether	1.270	1.244		-2.0	
2-Chlorophenol	1.203	1.195		-0.7	
2-Methylphenol	1.054	1.045		-0.9	
2,2-oxybis(1-Chloropropane)	1.846	1.776		-3.8	
Acetophenone	0.481	0.460		-4.4	
3+4-Methylphenols	1.329	1.328		-0.1	
n-Nitroso-di-n-propylamine	0.883	0.833	0.050	-5.7	
Nitrobenzene-d5	0.376	0.393		4.5	
Hexachloroethane	0.482	0.488		1.2	
Nitrobenzene	0.399	0.408		2.3	
Isophorone	0.737	0.726		-1.5	
2-Nitrophenol	0.127	0.175		37.8	20.0
2,4-Dimethylphenol	0.319	0.319		0.0	
bis(2-Chloroethoxy) methane	0.442	0.434		-1.8	
2,4-Dichlorophenol	0.303	0.311		2.6	20.0
Naphthalene	1.027	1.008		-1.9	
4-Chloroaniline	0.440	0.445		1.1	
Hexachlorobutadiene	0.231	0.224		-3.0	20.0
Caprolactam	0.086	0.098		14.0	
4-Chloro-3-methylphenol	0.305	0.312		2.3	20.0
2-Methylnaphthalene	0.707	0.688		-2.7	
Hexachlorocyclopentadiene	0.380	0.382	0.050	0.5	
2,4,6-Trichlorophenol	0.432	0.453		4.9	20.0
2-Fluorobiphenyl	1.369	1.224		-10.6	
2,4,5-Trichlorophenol	0.505	0.523		3.6	
1,1-Biphenyl	1.578	1.498		-5.1	
2-Chloronaphthalene	1.220	1.186		-2.8	
2-Nitroaniline	0.351	0.417		18.8	
Dimethylphthalate	1.449	1.451		0.1	
Acenaphthylene	1.951	1.915		-1.8	
2,6-Dinitrotoluene	0.281	0.329		17.1	
3-Nitroaniline	0.300	0.354		18.0	
Acenaphthene	1.174	1.145		-2.5	20.0
2,4-Dinitrophenol	0.108	0.144	0.050	33.3	
4-Nitrophenol	0.218	0.262	0.050	20.2	
Dibenzofuran	1.814	1.750		-3.5	
2,4-Dinitrotoluene	0.337	0.414		22.8	



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233

Instrument ID: BNA_F Calibration Date/Time: 01/20/2023 16:40

Lab File ID: BF132164.D Init. Calib. Date(s): 01/19/2023 01/19/2023

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:21 15:11

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Diethylphthalate	1.420	1.458		2.7	
4-Chlorophenyl-phenylether	0.720	0.700		-2.8	
Fluorene	1.344	1.292		-3.9	
4-Nitroaniline	0.279	0.333		19.4	
4,6-Dinitro-2-methylphenol	0.087	0.107		23.0	
n-Nitrosodiphenylamine	0.641	0.597		-6.9	20.0
2,4,6-Tribromophenol	0.214	0.231		7.9	
4-Bromophenyl-phenylether	0.244	0.229		-6.1	
Hexachlorobenzene	0.242	0.229		-5.4	
Atrazine	0.198	0.203		2.5	
Pentachlorophenol	0.156	0.174		11.5	20.0
Phenanthrene	1.050	1.000		-4.8	
Anthracene	1.062	1.012		-4.7	
Carbazole	0.907	0.914		0.8	
Di-n-butylphthalate	1.010	1.056		4.6	
Fluoranthene	1.078	1.118		3.7	20.0
Pyrene	1.843	1.942		5.4	
Terphenyl-d14	1.194	1.203		0.8	
Butylbenzylphthalate	0.534	0.671		25.7	
3,3-Dichlorobenzidine	0.398	0.413		3.8	
Benzo(a)anthracene	1.429	1.449		1.4	
Chrysene	1.360	1.334		-1.9	
Bis(2-ethylhexyl)phthalate	0.705	0.824		16.9	
Di-n-octyl phthalate	1.083	1.018		-6.0	20.0
Benzo(b)fluoranthene	1.218	1.276		4.8	
Benzo(k)fluoranthene	1.242	1.197		-3.6	
Benzo(a)pyrene	1.171	1.185		1.2	20.0
Indeno(1,2,3-cd)pyrene	1.451	1.547		6.6	
Dibenzo(a,h)anthracene	1.184	1.248		5.4	
Benzo(g,h,i)perylene	1.246	1.338		7.4	
1,2,4,5-Tetrachlorobenzene	0.696	0.669		-3.9	
1,4-Dioxane	0.580	0.596		2.8	20.0
2,3,4,6-Tetrachlorophenol	0.384	0.414		7.8	

All other compounds must meet a minimum RRF of 0.010.