



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/23/2023

Lab File ID: BG056380.D Time Analyzed: 11:41

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	35880	8.299	143339	11.14	121624	14.92
UPPER LIMIT	71760	8.799	286678	11.637	243248	15.42
LOWER LIMIT	17940	7.799	71669.5	10.637	60812	14.42
EPA SAMPLE NO.						
01 B-P37A	38461	8.30	155511	11.13	129252	14.91
02 B-P38A	36838	8.30	157784	11.13	127381	14.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.