



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233

EPA Sample No.: SSTDCCC040 Date Analyzed: 01/23/2023

Lab File ID: BG056380.D Time Analyzed: 11:41

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	314903	17.652	266234	21.947	293224	25.384
	UPPER LIMIT	629806	18.152	532468	22.447	586448	25.884
	LOWER LIMIT	157452	17.152	133117	21.447	146612	24.884
	EPA SAMPLE NO.						
01	B-P37A	335669	17.64	299254	21.94	323323	25.37
02	B-P38A	312833	17.65	275682	21.94	296075	25.36

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.