



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01

Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG No.: 01233

Instrument ID: MSVOA_Y Calibration Date/Time: 01/23/2023 12:09

Lab File ID: VY012307.D Init. Calib. Date(s): 01/16/2023 01/16/2023

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:17 14:34

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.437	0.387		-11.44	20
Chloromethane	0.403	0.377	0.1	-6.45	20
Vinyl Chloride	0.471	0.462		-1.91	20
Bromomethane	0.364	0.354		-2.75	20
Chloroethane	0.307	0.305		-0.65	20
Trichlorofluoromethane	0.919	0.884		-3.81	20
1,1,2-Trichlorotrifluoroethane	0.447	0.440		-1.57	20
1,1-Dichloroethene	0.459	0.454		-1.09	20
Acetone	0.139	0.146		5.04	20
Carbon Disulfide	1.420	1.365		-3.87	20
Methyl tert-butyl Ether	1.280	1.242		-2.97	20
Methyl Acetate	0.312	0.292		-6.41	20
Methylene Chloride	0.762	0.518		-32.02	20
trans-1,2-Dichloroethene	0.512	0.505		-1.37	20
1,1-Dichloroethane	0.785	0.778	0.1	-0.89	20
Cyclohexane	0.729	0.682		-6.45	20
2-Butanone	0.154	0.146		-5.2	20
Carbon Tetrachloride	0.581	0.569		-2.07	20
cis-1,2-Dichloroethene	0.563	0.549		-2.66	20
Bromochloromethane	0.171	0.159		-7.02	20
Chloroform	0.927	0.919		-0.86	20
1,1,1-Trichloroethane	0.918	0.900		-1.96	20
Methylcyclohexane	0.550	0.560		1.82	20
Benzene	1.299	1.299		0	20
1,2-Dichloroethane	0.413	0.406		-1.7	20
Trichloroethene	0.388	0.383		-1.29	20
1,2-Dichloropropane	0.271	0.274		1.11	20
Bromodichloromethane	0.457	0.456		-0.22	20
4-Methyl-2-Pentanone	0.197	0.186		-5.58	20
Toluene	0.865	0.875		1.16	20
t-1,3-Dichloropropene	0.478	0.466		-2.51	20
cis-1,3-Dichloropropene	0.514	0.511		-0.58	20
1,1,2-Trichloroethane	0.258	0.247		-4.26	20
2-Hexanone	0.150	0.147		-2	20
Dibromochloromethane	0.342	0.329		-3.8	20
1,2-Dibromoethane	0.249	0.239		-4.02	20
Tetrachloroethene	0.456	0.461		1.1	20
Chlorobenzene	1.023	1.005	0.3	-1.76	20
Ethyl Benzene	1.828	1.852		1.31	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233
Instrument ID: MSVOA_Y Calibration Date/Time: 01/23/2023 12:09
Lab File ID: VY012307.D Init. Calib. Date(s): 01/16/2023 01/16/2023
Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:17 14:34
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
m/p-Xylenes	0.723	0.728		0.69	20
o-Xylene	0.690	0.693		0.44	20
Styrene	1.154	1.171		1.47	20
Bromoform	0.247	0.234	0.1	-5.26	20
Isopropylbenzene	3.605	3.679		2.05	20
1,1,2,2-Tetrachloroethane	0.568	0.540	0.3	-4.93	20
1,3-Dichlorobenzene	1.726	1.693		-1.91	20
1,4-Dichlorobenzene	1.721	1.678		-2.5	20
1,2-Dichlorobenzene	1.540	1.496		-2.86	20
1,2-Dibromo-3-Chloropropane	0.117	0.110		-5.98	20
1,2,4-Trichlorobenzene	0.987	0.948		-3.95	20
1,2,3-Trichlorobenzene	0.849	0.816		-3.89	20
1,2-Dichloroethane-d4	0.465	0.425		-8.6	20
Dibromofluoromethane	0.265	0.258		-2.64	20
Toluene-d8	1.044	0.958		-8.24	20
4-Bromofluorobenzene	0.373	0.357		-4.29	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.