



284 Sheffield Street, Mountainside, New Jersey - 07092

Phone: (908) 789 8900 Fax: (908) 789 8922

## LAB CHRONICLE

|                                                       |                                                           |
|-------------------------------------------------------|-----------------------------------------------------------|
| <b>OrderID:</b> O1233                                 | <b>OrderDate:</b> 1/19/2023 1:53:00 PM                    |
| <b>Client:</b> Louis Berger U.S., Inc., A WSP Company | <b>Project:</b> NYCDDC Phase II SCI Arthur Kill Road CEQR |
| <b>Contact:</b> Jonathan Ganz                         | <b>Location:</b> N11,VOA Ref. #2 Soil,VOA Ref. #3 Water   |

| LabID      | ClientID | Matrix | Test          | Method | Sample Date | Prep Date | Anal Date | Received |
|------------|----------|--------|---------------|--------|-------------|-----------|-----------|----------|
| O1233-02   | B-P34A   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-03   | B-P34B   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-04   | B-P36A   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-05   | B-P36B   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/20/23  | 01/19/23 |
| O1233-05RE | B-P36BRE | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-06   | B-P37A   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-07   | B-P37B   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-08   | B-38A    | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/20/23  | 01/19/23 |
| O1233-08RE | B-38ARE  | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-09   | B-P38B   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/20/23  | 01/19/23 |
| O1233-09RE | B-P38BRE | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |
| O1233-11   | B-P38C   | SOIL   | VOC-TCLVOA-10 | 8260D  | 01/18/23    |           | 01/23/23  | 01/19/23 |



284 Sheffield Street, Mountainside, New Jersey - 07092

Phone: (908) 789 8900 Fax: (908) 789 8922

LAB CHRONICLE

|            |          |       |               |          |          |          |          |
|------------|----------|-------|---------------|----------|----------|----------|----------|
| O1233-11DL | B-P38CDL | SOIL  | VOC-TCLVOA-10 | 8260D    | 01/18/23 | 01/23/23 | 01/19/23 |
| O1233-12   | FB01     | Water | VOC-TCLVOA-10 | 8260-Low | 01/18/23 | 01/20/23 | 01/19/23 |

- A
- B
- C
- D
- E
- F
- G

### Hit Summary Sheet SW-846

SDG No.: O1233

Client: Louis Berger U.S., Inc., A WSP Company

| Sample ID                     | Client ID               | Matrix | Parameter                      | Concentration | C | MDL  | RDL  | Units |
|-------------------------------|-------------------------|--------|--------------------------------|---------------|---|------|------|-------|
| <b>Client ID:</b><br>O1233-02 | <b>B-P34A</b><br>B-P34A | SOIL   | Acetone                        | 22.2          | J | 14.1 | 28.9 | ug/Kg |
|                               |                         |        | <b>Total Voc :</b>             | 22.2          |   |      |      |       |
| O1233-02                      | B-P34A                  | SOIL   | unknown11.034                  | * 6.40        | J | 0    | 0    | ug/Kg |
| O1233-02                      | B-P34A                  | SOIL   | Ethanol, 2-(trimethylsilyl)-   | * 27.6        | J | 0    | 0    | ug/Kg |
|                               |                         |        | <b>Total Tics :</b>            | 34.0          |   |      |      |       |
|                               |                         |        | <b>Total Concentration:</b>    | 56.2          |   |      |      |       |
| <b>Client ID:</b><br>O1233-03 | <b>B-P34B</b><br>B-P34B | SOIL   | Silanol, trimethyl-            | * 24.0        | J | 0    | 0    | ug/Kg |
|                               |                         |        | <b>Total Tics :</b>            | 24.0          |   |      |      |       |
|                               |                         |        | <b>Total Concentration:</b>    | 24.0          |   |      |      |       |
| <b>Client ID:</b><br>O1233-04 | <b>B-P36A</b><br>B-P36A | SOIL   | Acetone                        | 17.6          | J | 13.5 | 27.7 | ug/Kg |
|                               |                         |        | <b>Total Voc :</b>             | 17.6          |   |      |      |       |
| O1233-04                      | B-P36A                  | SOIL   | Silanol, trimethyl-            | * 9.90        | J | 0    | 0    | ug/Kg |
|                               |                         |        | <b>Total Tics :</b>            | 9.90          |   |      |      |       |
|                               |                         |        | <b>Total Concentration:</b>    | 27.5          |   |      |      |       |
| <b>Client ID:</b><br>O1233-06 | <b>B-P37A</b><br>B-P37A | SOIL   | Ethanol, 2-(trimethylsilyl)-   | * 11.8        | J | 0    | 0    | ug/Kg |
|                               |                         |        | <b>Total Tics :</b>            | 11.8          |   |      |      |       |
|                               |                         |        | <b>Total Concentration:</b>    | 11.8          |   |      |      |       |
| <b>Client ID:</b><br>O1233-07 | <b>B-P37B</b><br>B-P37B | SOIL   | Acetone                        | 20.6          | J | 14.6 | 29.8 | ug/Kg |
|                               |                         |        | <b>Total Voc :</b>             | 20.6          |   |      |      |       |
| O1233-07                      | B-P37B                  | SOIL   | unknown10.947                  | * 8.50        | J | 0    | 0    | ug/Kg |
| O1233-07                      | B-P37B                  | SOIL   | Ethylene glycol, TMS derivativ | * 40.0        | J | 0    | 0    | ug/Kg |
|                               |                         |        | <b>Total Tics :</b>            | 48.5          |   |      |      |       |
|                               |                         |        | <b>Total Concentration:</b>    | 69.1          |   |      |      |       |
| <b>Client ID:</b><br>O1233-09 | <b>B-P38B</b><br>B-P38B | SOIL   | Octane                         | * 6.70        | J | 0    | 0    | ug/Kg |
|                               |                         |        | <b>Total Tics :</b>            | 6.70          |   |      |      |       |
|                               |                         |        | <b>Total Concentration:</b>    | 6.70          |   |      |      |       |
| <b>Client ID:</b><br>O1233-11 | <b>B-P38C</b><br>B-P38C | SOIL   | Cyclohexane                    | 4600          |   | 97.2 | 580  | ug/Kg |
| O1233-11                      | B-P38C                  | SOIL   | Methylcyclohexane              | 31500         | E | 92.6 | 580  | ug/Kg |
| O1233-11                      | B-P38C                  | SOIL   | Ethyl Benzene                  | 860           |   | 81.0 | 580  | ug/Kg |
| O1233-11                      | B-P38C                  | SOIL   | m/p-Xylenes                    | 2400          |   | 170  | 1200 | ug/Kg |
| O1233-11                      | B-P38C                  | SOIL   | o-Xylene                       | 320           | J | 91.4 | 580  | ug/Kg |
| O1233-11                      | B-P38C                  | SOIL   | Isopropylbenzene               | 1500          |   | 83.3 | 580  | ug/Kg |
|                               |                         |        | <b>Total Voc :</b>             | 41200         |   |      |      |       |
| O1233-11                      | B-P38C                  | SOIL   | unknown12.600                  | * 47300       | J | 0    | 0    | ug/Kg |

## Hit Summary Sheet

SW-846

**SDG No.:** O1233

**Client:** Louis Berger U.S., Inc., A WSP Company

| Sample ID            | Client ID | Matrix | Parameter              | Concentration | C  | MDL  | RDL   | Units |
|----------------------|-----------|--------|------------------------|---------------|----|------|-------|-------|
| O1233-11             | B-P38C    | SOIL   | Octane                 | * 54700       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | Heptane                | * 22000       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | Heptane, 3-methyl-     | * 22900       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | Heptane, 2-methyl-     | * 38200       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | 1-Phenyl-1-butene      | * 31200       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | Octane, 3-methyl-      | * 30700       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | Octane, 2-methyl-      | * 49000       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | Nonane, 3-methyl-      | * 28400       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | Nonane, 4-methyl-      | * 27300       | J  | 0    | 0     | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | n-propylbenzene        | * 3300        | J  | 82.2 | 580   | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | 1,3,5-Trimethylbenzene | * 4000        | J  | 74.1 | 580   | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | 1,2,4-Trimethylbenzene | * 7700        | J  | 82.2 | 580   | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | sec-Butylbenzene       | * 1300        | J  | 87.9 | 580   | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | p-Isopropyltoluene     | * 2600        | J  | 96.0 | 580   | ug/Kg |
| O1233-11             | B-P38C    | SOIL   | n-Butylbenzene         | * 3700        | J  | 82.2 | 580   | ug/Kg |
| Total Tics :         |           |        |                        | 374000        |    |      |       |       |
| Total Concentration: |           |        |                        | 415000        |    |      |       |       |
| Client ID:           | B-P38CDL  |        |                        |               |    |      |       |       |
| O1233-11DL           | B-P38CDL  | SOIL   | Cyclohexane            | 6300          | D  | 970  | 5800  | ug/Kg |
| O1233-11DL           | B-P38CDL  | SOIL   | Methylcyclohexane      | 33800         | D  | 930  | 5800  | ug/Kg |
| O1233-11DL           | B-P38CDL  | SOIL   | Ethyl Benzene          | 1000          | JD | 810  | 5800  | ug/Kg |
| O1233-11DL           | B-P38CDL  | SOIL   | m/p-Xylenes            | 2800          | JD | 1700 | 11600 | ug/Kg |
| O1233-11DL           | B-P38CDL  | SOIL   | Isopropylbenzene       | 2100          | JD | 830  | 5800  | ug/Kg |
| Total Voc :          |           |        |                        | 46000         |    |      |       |       |
| Total Concentration: |           |        |                        | 46000         |    |      |       |       |
| Client ID:           | FB01      |        |                        |               |    |      |       |       |
| O1233-12             | FB01      | Water  | Methylene Chloride     | 2.80          |    | 0.18 | 1.00  | ug/L  |
| Total Voc :          |           |        |                        | 2.80          |    |      |       |       |
| Total Concentration: |           |        |                        | 2.80          |    |      |       |       |

# SAMPLE DATA

A

B

C

D

E

F

G

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P34A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-02                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.3          |
| Sample Wt/Vol:     | 5.07 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075172.D        | 1         |           | 01/23/23 13:36 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.99  | U         | 0.99 | 5.80       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.20  | U         | 1.20 | 5.80       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.10  | U         | 1.10 | 5.80       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.80       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.00  | U         | 1.00 | 5.80       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.80       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.83  | U         | 0.83 | 5.80       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.99  | U         | 0.99 | 5.80       | ug/Kg             |
| 67-64-1        | Acetone                        | 22.2  | J         | 14.1 | 28.9       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.87  | U         | 0.87 | 5.80       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.10  | U         | 1.10 | 5.80       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.50  | U         | 1.50 | 5.80       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.90  | U         | 6.90 | 11.6       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.79  | U         | 0.79 | 5.80       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.81  | U         | 0.81 | 5.80       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.97  | U         | 0.97 | 5.80       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.40  | U         | 8.40 | 28.9       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.91  | U         | 0.91 | 5.80       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.79  | U         | 0.79 | 5.80       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.94  | U         | 0.94 | 5.80       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.77  | U         | 0.77 | 5.80       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.87  | U         | 0.87 | 5.80       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.92  | U         | 0.92 | 5.80       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.76  | U         | 0.76 | 5.80       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.97  | U         | 0.97 | 5.80       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.84  | U         | 0.84 | 5.80       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.75  | U         | 0.75 | 5.80       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.81  | U         | 0.81 | 5.80       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.30  | U         | 5.30 | 28.9       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.73  | U         | 0.73 | 5.80       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.86  | U         | 0.86 | 5.80       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.82  | U         | 0.82 | 5.80       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P34A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-02                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.3          |
| Sample Wt/Vol:     | 5.07 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075172.D        | 1         |           | 01/23/23 13:36 | VD012323      |

| CAS Number                            | Parameter                    | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|------------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                               | 1,1,2-Trichloroethane        | 0.99   | U         | 0.99     | 5.80       | ug/Kg             |
| 591-78-6                              | 2-Hexanone                   | 5.40   | U         | 5.40     | 28.9       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane         | 0.87   | U         | 0.87     | 5.80       | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane            | 0.87   | U         | 0.87     | 5.80       | ug/Kg             |
| 127-18-4                              | Tetrachloroethene            | 0.88   | U         | 0.88     | 5.80       | ug/Kg             |
| 108-90-7                              | Chlorobenzene                | 0.75   | U         | 0.75     | 5.80       | ug/Kg             |
| 100-41-4                              | Ethyl Benzene                | 0.81   | U         | 0.81     | 5.80       | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                  | 1.70   | U         | 1.70     | 11.6       | ug/Kg             |
| 95-47-6                               | o-Xylene                     | 0.91   | U         | 0.91     | 5.80       | ug/Kg             |
| 100-42-5                              | Styrene                      | 0.91   | U         | 0.91     | 5.80       | ug/Kg             |
| 75-25-2                               | Bromoform                    | 0.94   | U         | 0.94     | 5.80       | ug/Kg             |
| 98-82-8                               | Isopropylbenzene             | 0.83   | U         | 0.83     | 5.80       | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane    | 1.30   | U         | 1.30     | 5.80       | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene          | 0.77   | U         | 0.77     | 5.80       | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene          | 0.73   | U         | 0.73     | 5.80       | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene          | 0.74   | U         | 0.74     | 5.80       | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane  | 1.40   | U         | 1.40     | 5.80       | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene       | 1.10   | U         | 1.10     | 5.80       | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene       | 1.20   | U         | 1.20     | 5.80       | ug/Kg             |
| <b>SURROGATES</b>                     |                              |        |           |          |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4        | 58.5   |           | 50 - 163 | 117%       | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane         | 51.4   |           | 54 - 147 | 103%       | SPK: 50           |
| 2037-26-5                             | Toluene-d8                   | 52.1   |           | 58 - 134 | 104%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene         | 53.8   |           | 30 - 143 | 108%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                              |        |           |          |            |                   |
| 363-72-4                              | Pentafluorobenzene           | 67900  | 7.881     |          |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene          | 130000 | 8.781     |          |            |                   |
| 3114-55-4                             | Chlorobenzene-d5             | 125000 | 11.587    |          |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4       | 54900  | 13.522    |          |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                              |        |           |          |            |                   |
| 002916-68-9                           | Ethanol, 2-(trimethylsilyl)- | 27.6   | J         |          | 7.07       | ug/Kg             |
|                                       | unknown11.034                | 6.40   | J         |          | 11.0       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P34A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-02                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.3          |
| Sample Wt/Vol:     | 5.07 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075172.D        | 1         |           | 01/23/23 13:36 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|------------|-----------|-------|-----------|-----|------------|-------------------|
|------------|-----------|-------|-----------|-----|------------|-------------------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

# Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P34B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-03                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.7          |
| Sample Wt/Vol:     | 5.06 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075173.D        | 1         |           | 01/23/23 14:04 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.94  | U         | 0.94 | 5.40       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.10  | U         | 1.10 | 5.40       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.99  | U         | 0.99 | 5.40       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.40       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.97  | U         | 0.97 | 5.40       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.40       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.78  | U         | 0.78 | 5.40       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.94  | U         | 0.94 | 5.40       | ug/Kg             |
| 67-64-1        | Acetone                        | 13.3  | U         | 13.3 | 27.2       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.82  | U         | 0.82 | 5.40       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.00  | U         | 1.00 | 5.40       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 5.40       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.50  | U         | 6.50 | 10.9       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.74  | U         | 0.74 | 5.40       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.76  | U         | 0.76 | 5.40       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.92  | U         | 0.92 | 5.40       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 7.90  | U         | 7.90 | 27.2       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.86  | U         | 0.86 | 5.40       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.74  | U         | 0.74 | 5.40       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.88  | U         | 0.88 | 5.40       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.73  | U         | 0.73 | 5.40       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.82  | U         | 0.82 | 5.40       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.87  | U         | 0.87 | 5.40       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.72  | U         | 0.72 | 5.40       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.92  | U         | 0.92 | 5.40       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.80  | U         | 0.80 | 5.40       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.71  | U         | 0.71 | 5.40       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.76  | U         | 0.76 | 5.40       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.00  | U         | 5.00 | 27.2       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.69  | U         | 0.69 | 5.40       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.81  | U         | 0.81 | 5.40       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.77  | U         | 0.77 | 5.40       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P34B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-03                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.7          |
| Sample Wt/Vol:     | 5.06 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075173.D        | 1         |           | 01/23/23 14:04 | VD012323      |

| CAS Number                            | Parameter                   | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|-----------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                               | 1,1,2-Trichloroethane       | 0.94  | U         | 0.94     | 5.40       | ug/Kg             |
| 591-78-6                              | 2-Hexanone                  | 5.10  | U         | 5.10     | 27.2       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane        | 0.82  | U         | 0.82     | 5.40       | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane           | 0.82  | U         | 0.82     | 5.40       | ug/Kg             |
| 127-18-4                              | Tetrachloroethene           | 0.83  | U         | 0.83     | 5.40       | ug/Kg             |
| 108-90-7                              | Chlorobenzene               | 0.71  | U         | 0.71     | 5.40       | ug/Kg             |
| 100-41-4                              | Ethyl Benzene               | 0.76  | U         | 0.76     | 5.40       | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                 | 1.60  | U         | 1.60     | 10.9       | ug/Kg             |
| 95-47-6                               | o-Xylene                    | 0.86  | U         | 0.86     | 5.40       | ug/Kg             |
| 100-42-5                              | Styrene                     | 0.86  | U         | 0.86     | 5.40       | ug/Kg             |
| 75-25-2                               | Bromoform                   | 0.88  | U         | 0.88     | 5.40       | ug/Kg             |
| 98-82-8                               | Isopropylbenzene            | 0.78  | U         | 0.78     | 5.40       | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane   | 1.20  | U         | 1.20     | 5.40       | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene         | 0.73  | U         | 0.73     | 5.40       | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene         | 0.69  | U         | 0.69     | 5.40       | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene         | 0.70  | U         | 0.70     | 5.40       | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane | 1.40  | U         | 1.40     | 5.40       | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene      | 1.00  | U         | 1.00     | 5.40       | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene      | 1.10  | U         | 1.10     | 5.40       | ug/Kg             |
| <b>SURROGATES</b>                     |                             |       |           |          |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4       | 66.4  |           | 50 - 163 | 133%       | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane        | 53.5  |           | 54 - 147 | 107%       | SPK: 50           |
| 2037-26-5                             | Toluene-d8                  | 52.7  |           | 58 - 134 | 105%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene        | 56.2  |           | 30 - 143 | 112%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                             |       |           |          |            |                   |
| 363-72-4                              | Pentafluorobenzene          | 49100 | 7.881     |          |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene         | 95000 | 8.781     |          |            |                   |
| 3114-55-4                             | Chlorobenzene-d5            | 93600 | 11.587    |          |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4      | 41800 | 13.522    |          |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                             |       |           |          |            |                   |
| 001066-40-6                           | Silanol, trimethyl-         | 24.0  | J         |          | 7.07       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |                              |
|--------------------|-------------------------------------------|-----------------|------------------------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23                     |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23                     |
| Client Sample ID:  | B-P34B                                    | SDG No.:        | O1233                        |
| Lab Sample ID:     | O1233-03                                  | Matrix:         | SOIL                         |
| Analytical Method: | SW8260                                    | % Solid:        | 90.7                         |
| Sample Wt/Vol:     | 5.06      Units:    g                     | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10                |
| GC Column:         | RTX-VMS      ID :    0.18                 | Level :         | LOW                          |
| Prep Method :      |                                           |                 |                              |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075173.D        | 1         |           | 01/23/23 14:04 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|------------|-----------|-------|-----------|-----|------------|-------------------|
|------------|-----------|-------|-----------|-----|------------|-------------------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-04                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 89.6          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075174.D        | 1         |           | 01/23/23 14:32 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.95  | U         | 0.95 | 5.50       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.20  | U         | 1.20 | 5.50       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.00  | U         | 1.00 | 5.50       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.50       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.99  | U         | 0.99 | 5.50       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.50       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.80  | U         | 0.80 | 5.50       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.95  | U         | 0.95 | 5.50       | ug/Kg             |
| 67-64-1        | Acetone                        | 17.6  | J         | 13.5 | 27.7       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.83  | U         | 0.83 | 5.50       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.00  | U         | 1.00 | 5.50       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 5.50       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.60  | U         | 6.60 | 11.1       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.75  | U         | 0.75 | 5.50       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.78  | U         | 0.78 | 5.50       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.93  | U         | 0.93 | 5.50       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.10  | U         | 8.10 | 27.7       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.88  | U         | 0.88 | 5.50       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.75  | U         | 0.75 | 5.50       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.90  | U         | 0.90 | 5.50       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.74  | U         | 0.74 | 5.50       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.83  | U         | 0.83 | 5.50       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.89  | U         | 0.89 | 5.50       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.73  | U         | 0.73 | 5.50       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.93  | U         | 0.93 | 5.50       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.81  | U         | 0.81 | 5.50       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.72  | U         | 0.72 | 5.50       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.78  | U         | 0.78 | 5.50       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.10  | U         | 5.10 | 27.7       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.70  | U         | 0.70 | 5.50       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.82  | U         | 0.82 | 5.50       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.79  | U         | 0.79 | 5.50       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-04                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 89.6          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075174.D        | 1         |           | 01/23/23 14:32 | VD012323      |

| CAS Number                            | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                               | 1,1,2-Trichloroethane       | 0.95   | U         | 0.95     | 5.50       | ug/Kg             |
| 591-78-6                              | 2-Hexanone                  | 5.20   | U         | 5.20     | 27.7       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane        | 0.83   | U         | 0.83     | 5.50       | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane           | 0.83   | U         | 0.83     | 5.50       | ug/Kg             |
| 127-18-4                              | Tetrachloroethene           | 0.84   | U         | 0.84     | 5.50       | ug/Kg             |
| 108-90-7                              | Chlorobenzene               | 0.72   | U         | 0.72     | 5.50       | ug/Kg             |
| 100-41-4                              | Ethyl Benzene               | 0.78   | U         | 0.78     | 5.50       | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                 | 1.60   | U         | 1.60     | 11.1       | ug/Kg             |
| 95-47-6                               | o-Xylene                    | 0.88   | U         | 0.88     | 5.50       | ug/Kg             |
| 100-42-5                              | Styrene                     | 0.88   | U         | 0.88     | 5.50       | ug/Kg             |
| 75-25-2                               | Bromoform                   | 0.90   | U         | 0.90     | 5.50       | ug/Kg             |
| 98-82-8                               | Isopropylbenzene            | 0.80   | U         | 0.80     | 5.50       | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane   | 1.30   | U         | 1.30     | 5.50       | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene         | 0.74   | U         | 0.74     | 5.50       | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene         | 0.70   | U         | 0.70     | 5.50       | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene         | 0.71   | U         | 0.71     | 5.50       | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane | 1.40   | U         | 1.40     | 5.50       | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene      | 1.00   | U         | 1.00     | 5.50       | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene      | 1.10   | U         | 1.10     | 5.50       | ug/Kg             |
| <b>SURROGATES</b>                     |                             |        |           |          |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4       | 49.8   |           | 50 - 163 | 100%       | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane        | 47.1   |           | 54 - 147 | 94%        | SPK: 50           |
| 2037-26-5                             | Toluene-d8                  | 51.5   |           | 58 - 134 | 103%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene        | 50.5   |           | 30 - 143 | 101%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                             |        |           |          |            |                   |
| 363-72-4                              | Pentafluorobenzene          | 88000  | 7.875     |          |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene         | 171000 | 8.781     |          |            |                   |
| 3114-55-4                             | Chlorobenzene-d5            | 158000 | 11.587    |          |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4      | 64200  | 13.522    |          |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                             |        |           |          |            |                   |
| 001066-40-6                           | Silanol, trimethyl-         | 9.90   | J         |          | 7.07       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |                              |
|--------------------|-------------------------------------------|-----------------|------------------------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23                     |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23                     |
| Client Sample ID:  | B-P36A                                    | SDG No.:        | O1233                        |
| Lab Sample ID:     | O1233-04                                  | Matrix:         | SOIL                         |
| Analytical Method: | SW8260                                    | % Solid:        | 89.6                         |
| Sample Wt/Vol:     | 5.03      Units:    g                     | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10                |
| GC Column:         | RTX-VMS      ID :    0.18                 | Level :         | LOW                          |
| Prep Method :      |                                           |                 |                              |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075174.D        | 1         |           | 01/23/23 14:32 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|------------|-----------|-------|-----------|-----|------------|-------------------|
|------------|-----------|-------|-----------|-----|------------|-------------------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-05                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.4          |
| Sample Wt/Vol:     | 5.08 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075158.D        | 1         |           | 01/20/23 17:26 | VD012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.94  | U         | 0.94 | 5.40       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.10  | U         | 1.10 | 5.40       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.99  | U         | 0.99 | 5.40       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.40       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.97  | U         | 0.97 | 5.40       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.40       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.78  | U         | 0.78 | 5.40       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.94  | U         | 0.94 | 5.40       | ug/Kg             |
| 67-64-1        | Acetone                        | 13.3  | U         | 13.3 | 27.2       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.82  | U         | 0.82 | 5.40       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.00  | U         | 1.00 | 5.40       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 5.40       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.50  | U         | 6.50 | 10.9       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.74  | U         | 0.74 | 5.40       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.76  | U         | 0.76 | 5.40       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.91  | U         | 0.91 | 5.40       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 7.90  | U         | 7.90 | 27.2       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.86  | U         | 0.86 | 5.40       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.74  | U         | 0.74 | 5.40       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.88  | U         | 0.88 | 5.40       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.73  | U         | 0.73 | 5.40       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.82  | U         | 0.82 | 5.40       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.87  | U         | 0.87 | 5.40       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.72  | U         | 0.72 | 5.40       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.91  | U         | 0.91 | 5.40       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.79  | U         | 0.79 | 5.40       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.71  | U         | 0.71 | 5.40       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.76  | U         | 0.76 | 5.40       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.00  | U         | 5.00 | 27.2       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.69  | U         | 0.69 | 5.40       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.81  | U         | 0.81 | 5.40       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.77  | U         | 0.77 | 5.40       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-05                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.4          |
| Sample Wt/Vol:     | 5.08 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075158.D        | 1         |           | 01/20/23 17:26 | VD012023      |

| CAS Number                | Parameter                   | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.94  | U         | 0.94     | 5.40       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 5.10  | U         | 5.10     | 27.2       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.82  | U         | 0.82     | 5.40       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.82  | U         | 0.82     | 5.40       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.83  | U         | 0.83     | 5.40       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.71  | U         | 0.71     | 5.40       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.76  | U         | 0.76     | 5.40       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.60  | U         | 1.60     | 10.9       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.86  | U         | 0.86     | 5.40       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.86  | U         | 0.86     | 5.40       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.88  | U         | 0.88     | 5.40       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.78  | U         | 0.78     | 5.40       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.20  | U         | 1.20     | 5.40       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.73  | U         | 0.73     | 5.40       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.69  | U         | 0.69     | 5.40       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.70  | U         | 0.70     | 5.40       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.40  | U         | 1.40     | 5.40       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1.00  | U         | 1.00     | 5.40       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.10  | U         | 1.10     | 5.40       | ug/Kg             |
| <b>SURROGATES</b>         |                             |       |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 92.5  | *         | 50 - 163 | 185%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 56.1  |           | 54 - 147 | 112%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 51.5  |           | 58 - 134 | 103%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 60.1  |           | 30 - 143 | 120%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |       |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 4930  | 7.875     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 11100 | 8.781     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 10700 | 11.587    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 5310  | 13.522    |          |            |                   |

**Report of Analysis**

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-05                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.4          |
| Sample Wt/Vol:     | 5.08 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075158.D        | 1         |           | 01/20/23 17:26 | VD012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36BRE                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-05RE                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.4          |
| Sample Wt/Vol:     | 5.02 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075175.D        | 1         |           | 01/23/23 15:00 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.95  | U         | 0.95 | 5.50       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.20  | U         | 1.20 | 5.50       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.00  | U         | 1.00 | 5.50       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.50       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.98  | U         | 0.98 | 5.50       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.50       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.79  | U         | 0.79 | 5.50       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.95  | U         | 0.95 | 5.50       | ug/Kg             |
| 67-64-1        | Acetone                        | 13.4  | U         | 13.4 | 27.5       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.83  | U         | 0.83 | 5.50       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.00  | U         | 1.00 | 5.50       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 5.50       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.60  | U         | 6.60 | 11.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.75  | U         | 0.75 | 5.50       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.77  | U         | 0.77 | 5.50       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.93  | U         | 0.93 | 5.50       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.00  | U         | 8.00 | 27.5       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.87  | U         | 0.87 | 5.50       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.75  | U         | 0.75 | 5.50       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.89  | U         | 0.89 | 5.50       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.74  | U         | 0.74 | 5.50       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.83  | U         | 0.83 | 5.50       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.88  | U         | 0.88 | 5.50       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.73  | U         | 0.73 | 5.50       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.93  | U         | 0.93 | 5.50       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.80  | U         | 0.80 | 5.50       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.72  | U         | 0.72 | 5.50       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.77  | U         | 0.77 | 5.50       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.00  | U         | 5.00 | 27.5       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.69  | U         | 0.69 | 5.50       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.82  | U         | 0.82 | 5.50       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.78  | U         | 0.78 | 5.50       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36BRE                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-05RE                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.4          |
| Sample Wt/Vol:     | 5.02 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075175.D        | 1         |           | 01/23/23 15:00 | VD012323      |

| CAS Number                | Parameter                   | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.95  | U         | 0.95     | 5.50       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 5.20  | U         | 5.20     | 27.5       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.83  | U         | 0.83     | 5.50       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.83  | U         | 0.83     | 5.50       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.84  | U         | 0.84     | 5.50       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.72  | U         | 0.72     | 5.50       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.77  | U         | 0.77     | 5.50       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.60  | U         | 1.60     | 11.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.87  | U         | 0.87     | 5.50       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.87  | U         | 0.87     | 5.50       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.89  | U         | 0.89     | 5.50       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.79  | U         | 0.79     | 5.50       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.20  | U         | 1.20     | 5.50       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.74  | U         | 0.74     | 5.50       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.69  | U         | 0.69     | 5.50       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.71  | U         | 0.71     | 5.50       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.40  | U         | 1.40     | 5.50       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1.00  | U         | 1.00     | 5.50       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.10  | U         | 1.10     | 5.50       | ug/Kg             |
| <b>SURROGATES</b>         |                             |       |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 68.1  |           | 50 - 163 | 136%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 52.3  |           | 54 - 147 | 105%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 48.9  |           | 58 - 134 | 98%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 45.8  |           | 30 - 143 | 92%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |       |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 2290  | 7.881     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 4730  | 8.775     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 4220  | 11.581    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 1550  | 13.522    |          |            |                   |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P36BRE                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-05RE                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 90.4          |
| Sample Wt/Vol:     | 5.02 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075175.D        | 1         |           | 01/23/23 15:00 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P37A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-06                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 87.9          |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075176.D        | 1         |           | 01/23/23 15:27 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.98  | U         | 0.98 | 5.70       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.20  | U         | 1.20 | 5.70       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.00  | U         | 1.00 | 5.70       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.70       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.00  | U         | 1.00 | 5.70       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.70       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.82  | U         | 0.82 | 5.70       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.98  | U         | 0.98 | 5.70       | ug/Kg             |
| 67-64-1        | Acetone                        | 13.9  | U         | 13.9 | 28.4       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.85  | U         | 0.85 | 5.70       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.10  | U         | 1.10 | 5.70       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 5.70       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.80  | U         | 6.80 | 11.4       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.77  | U         | 0.77 | 5.70       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.80  | U         | 0.80 | 5.70       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.96  | U         | 0.96 | 5.70       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.30  | U         | 8.30 | 28.4       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.90  | U         | 0.90 | 5.70       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.77  | U         | 0.77 | 5.70       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.92  | U         | 0.92 | 5.70       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.76  | U         | 0.76 | 5.70       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.85  | U         | 0.85 | 5.70       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.91  | U         | 0.91 | 5.70       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.75  | U         | 0.75 | 5.70       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.96  | U         | 0.96 | 5.70       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.83  | U         | 0.83 | 5.70       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.74  | U         | 0.74 | 5.70       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.80  | U         | 0.80 | 5.70       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.20  | U         | 5.20 | 28.4       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.72  | U         | 0.72 | 5.70       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.84  | U         | 0.84 | 5.70       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.81  | U         | 0.81 | 5.70       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P37A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-06                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 87.9          |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075176.D        | 1         |           | 01/23/23 15:27 | VD012323      |

| CAS Number                            | Parameter                    | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|------------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                               | 1,1,2-Trichloroethane        | 0.98   | U         | 0.98     | 5.70       | ug/Kg             |
| 591-78-6                              | 2-Hexanone                   | 5.30   | U         | 5.30     | 28.4       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane         | 0.85   | U         | 0.85     | 5.70       | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane            | 0.85   | U         | 0.85     | 5.70       | ug/Kg             |
| 127-18-4                              | Tetrachloroethene            | 0.86   | U         | 0.86     | 5.70       | ug/Kg             |
| 108-90-7                              | Chlorobenzene                | 0.74   | U         | 0.74     | 5.70       | ug/Kg             |
| 100-41-4                              | Ethyl Benzene                | 0.80   | U         | 0.80     | 5.70       | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                  | 1.70   | U         | 1.70     | 11.4       | ug/Kg             |
| 95-47-6                               | o-Xylene                     | 0.90   | U         | 0.90     | 5.70       | ug/Kg             |
| 100-42-5                              | Styrene                      | 0.90   | U         | 0.90     | 5.70       | ug/Kg             |
| 75-25-2                               | Bromoform                    | 0.92   | U         | 0.92     | 5.70       | ug/Kg             |
| 98-82-8                               | Isopropylbenzene             | 0.82   | U         | 0.82     | 5.70       | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane    | 1.30   | U         | 1.30     | 5.70       | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene          | 0.76   | U         | 0.76     | 5.70       | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene          | 0.72   | U         | 0.72     | 5.70       | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene          | 0.73   | U         | 0.73     | 5.70       | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane  | 1.40   | U         | 1.40     | 5.70       | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene       | 1.10   | U         | 1.10     | 5.70       | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene       | 1.10   | U         | 1.10     | 5.70       | ug/Kg             |
| <b>SURROGATES</b>                     |                              |        |           |          |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4        | 50.8   |           | 50 - 163 | 102%       | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane         | 48.2   |           | 54 - 147 | 96%        | SPK: 50           |
| 2037-26-5                             | Toluene-d8                   | 51.4   |           | 58 - 134 | 103%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene         | 50.0   |           | 30 - 143 | 100%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                              |        |           |          |            |                   |
| 363-72-4                              | Pentafluorobenzene           | 85100  | 7.881     |          |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene          | 168000 | 8.781     |          |            |                   |
| 3114-55-4                             | Chlorobenzene-d5             | 154000 | 11.587    |          |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4       | 63400  | 13.522    |          |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                              |        |           |          |            |                   |
| 002916-68-9                           | Ethanol, 2-(trimethylsilyl)- | 11.8   | J         |          | 7.08       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P37A                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-06                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 87.9          |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075176.D        | 1         |           | 01/23/23 15:27 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|------------|-----------|-------|-----------|-----|------------|-------------------|
|------------|-----------|-------|-----------|-----|------------|-------------------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P37B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-07                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 83.5          |
| Sample Wt/Vol:     | 5.02 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012315.D        | 1         |           | 01/23/23 15:48 | VY012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1.00  | U         | 1.00 | 6.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.30  | U         | 1.30 | 6.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.10  | U         | 1.10 | 6.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.40  | U         | 1.40 | 6.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.10  | U         | 1.10 | 6.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.20  | U         | 1.20 | 6.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.86  | U         | 0.86 | 6.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 1.00  | U         | 1.00 | 6.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 20.6  | J         | 14.6 | 29.8       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.89  | U         | 0.89 | 6.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.10  | U         | 1.10 | 6.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.50  | U         | 1.50 | 6.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 7.10  | U         | 7.10 | 11.9       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.81  | U         | 0.81 | 6.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.83  | U         | 0.83 | 6.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 1.00  | U         | 1.00 | 6.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.70  | U         | 8.70 | 29.8       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.94  | U         | 0.94 | 6.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.81  | U         | 0.81 | 6.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.97  | U         | 0.97 | 6.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.80  | U         | 0.80 | 6.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.89  | U         | 0.89 | 6.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.95  | U         | 0.95 | 6.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.79  | U         | 0.79 | 6.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 1.00  | U         | 1.00 | 6.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.87  | U         | 0.87 | 6.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.78  | U         | 0.78 | 6.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.83  | U         | 0.83 | 6.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.50  | U         | 5.50 | 29.8       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.75  | U         | 0.75 | 6.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.88  | U         | 0.88 | 6.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.85  | U         | 0.85 | 6.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P37B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-07                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 83.5          |
| Sample Wt/Vol:     | 5.02 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012315.D        | 1         |           | 01/23/23 15:48 | VY012323      |

| CAS Number                            | Parameter                       | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|---------------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                               | 1,1,2-Trichloroethane           | 1.00  | U         | 1.00     | 6.00       | ug/Kg             |
| 591-78-6                              | 2-Hexanone                      | 5.60  | U         | 5.60     | 29.8       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane            | 0.89  | U         | 0.89     | 6.00       | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane               | 0.89  | U         | 0.89     | 6.00       | ug/Kg             |
| 127-18-4                              | Tetrachloroethene               | 0.91  | U         | 0.91     | 6.00       | ug/Kg             |
| 108-90-7                              | Chlorobenzene                   | 0.78  | U         | 0.78     | 6.00       | ug/Kg             |
| 100-41-4                              | Ethyl Benzene                   | 0.83  | U         | 0.83     | 6.00       | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                     | 1.80  | U         | 1.80     | 11.9       | ug/Kg             |
| 95-47-6                               | o-Xylene                        | 0.94  | U         | 0.94     | 6.00       | ug/Kg             |
| 100-42-5                              | Styrene                         | 0.94  | U         | 0.94     | 6.00       | ug/Kg             |
| 75-25-2                               | Bromoform                       | 0.97  | U         | 0.97     | 6.00       | ug/Kg             |
| 98-82-8                               | Isopropylbenzene                | 0.86  | U         | 0.86     | 6.00       | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane       | 1.30  | U         | 1.30     | 6.00       | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene             | 0.80  | U         | 0.80     | 6.00       | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene             | 0.75  | U         | 0.75     | 6.00       | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene             | 0.76  | U         | 0.76     | 6.00       | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane     | 1.50  | U         | 1.50     | 6.00       | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene          | 1.10  | U         | 1.10     | 6.00       | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene          | 1.20  | U         | 1.20     | 6.00       | ug/Kg             |
| <b>SURROGATES</b>                     |                                 |       |           |          |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4           | 83.0  | *         | 50 - 163 | 166%       | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane            | 60.1  |           | 54 - 147 | 120%       | SPK: 50           |
| 2037-26-5                             | Toluene-d8                      | 66.9  |           | 58 - 134 | 134%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene            | 48.8  |           | 30 - 143 | 98%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                                 |       |           |          |            |                   |
| 363-72-4                              | Pentafluorobenzene              | 35900 | 7.789     |          |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene             | 57300 | 8.691     |          |            |                   |
| 3114-55-4                             | Chlorobenzene-d5                | 56000 | 11.489    |          |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4          | 24400 | 13.428    |          |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                                 |       |           |          |            |                   |
| 004403-13-8                           | Ethylene glycol, TMS derivative | 40.0  | J         |          | 6.97       | ug/Kg             |
|                                       | unknown10.947                   | 8.50  | J         |          | 10.9       | ug/Kg             |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P37B                                    |           | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-07                                  |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 83.5          |
| Sample Wt/Vol:     | 5.02                                      | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012315.D        | 1         |           | 01/23/23 15:48 | VY012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|------------|-----------|-------|-----------|-----|------------|-------------------|
|------------|-----------|-------|-----------|-----|------------|-------------------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-38A                                     | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-08                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 89.3          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075161.D        | 1         |           | 01/20/23 18:49 | VD012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.96  | U         | 0.96 | 5.60       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.20  | U         | 1.20 | 5.60       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.00  | U         | 1.00 | 5.60       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.60       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.99  | U         | 0.99 | 5.60       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.60       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.80  | U         | 0.80 | 5.60       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.96  | U         | 0.96 | 5.60       | ug/Kg             |
| 67-64-1        | Acetone                        | 13.6  | U         | 13.6 | 27.8       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.83  | U         | 0.83 | 5.60       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.00  | U         | 1.00 | 5.60       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 5.60       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.60  | U         | 6.60 | 11.1       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.76  | U         | 0.76 | 5.60       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.78  | U         | 0.78 | 5.60       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.94  | U         | 0.94 | 5.60       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.10  | U         | 8.10 | 27.8       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.88  | U         | 0.88 | 5.60       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.76  | U         | 0.76 | 5.60       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.90  | U         | 0.90 | 5.60       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.75  | U         | 0.75 | 5.60       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.83  | U         | 0.83 | 5.60       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.89  | U         | 0.89 | 5.60       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.73  | U         | 0.73 | 5.60       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.94  | U         | 0.94 | 5.60       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.81  | U         | 0.81 | 5.60       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.72  | U         | 0.72 | 5.60       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.78  | U         | 0.78 | 5.60       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.10  | U         | 5.10 | 27.8       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.70  | U         | 0.70 | 5.60       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.82  | U         | 0.82 | 5.60       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.79  | U         | 0.79 | 5.60       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-38A                                     | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-08                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 89.3          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075161.D        | 1         |           | 01/20/23 18:49 | VD012023      |

| CAS Number                | Parameter                   | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.96  | U         | 0.96     | 5.60       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 5.20  | U         | 5.20     | 27.8       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.83  | U         | 0.83     | 5.60       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.83  | U         | 0.83     | 5.60       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.85  | U         | 0.85     | 5.60       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.72  | U         | 0.72     | 5.60       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.78  | U         | 0.78     | 5.60       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.60  | U         | 1.60     | 11.1       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.88  | U         | 0.88     | 5.60       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.88  | U         | 0.88     | 5.60       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.90  | U         | 0.90     | 5.60       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.80  | U         | 0.80     | 5.60       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.30  | U         | 1.30     | 5.60       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.75  | U         | 0.75     | 5.60       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.70  | U         | 0.70     | 5.60       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.71  | U         | 0.71     | 5.60       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.40  | U         | 1.40     | 5.60       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1.00  | U         | 1.00     | 5.60       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.10  | U         | 1.10     | 5.60       | ug/Kg             |
| <b>SURROGATES</b>         |                             |       |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 90.7  | *         | 50 - 163 | 181%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 55.9  |           | 54 - 147 | 112%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 55.2  |           | 58 - 134 | 110%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 56.9  |           | 30 - 143 | 114%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |       |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 3460  | 7.875     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 6980  | 8.775     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 6960  | 11.581    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 2980  | 13.516    |          |            |                   |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-38A                                     | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-08                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 89.3          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075161.D        | 1         |           | 01/20/23 18:49 | VD012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |                      |
|--------------------|-------------------------------------------|-----------------|----------------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23             |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23             |
| Client Sample ID:  | B-38ARE                                   | SDG No.:        | O1233                |
| Lab Sample ID:     | O1233-08RE                                | Matrix:         | SOIL                 |
| Analytical Method: | SW8260                                    | % Solid:        | 89.3                 |
| Sample Wt/Vol:     | 5.04      Units:    g                     | Final Vol:      | 5000              uL |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10        |
| GC Column:         | RXI-624              ID :    0.25         | Level :         | LOW                  |
| Prep Method :      |                                           |                 |                      |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012316.D        | 1         |           | 01/23/23 16:11 | VY012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.96  | U         | 0.96 | 5.60       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.20  | U         | 1.20 | 5.60       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.00  | U         | 1.00 | 5.60       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.30  | U         | 1.30 | 5.60       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.99  | U         | 0.99 | 5.60       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.10  | U         | 1.10 | 5.60       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.80  | U         | 0.80 | 5.60       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.96  | U         | 0.96 | 5.60       | ug/Kg             |
| 67-64-1        | Acetone                        | 13.6  | U         | 13.6 | 27.8       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.83  | U         | 0.83 | 5.60       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.00  | U         | 1.00 | 5.60       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.40  | U         | 1.40 | 5.60       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.60  | U         | 6.60 | 11.1       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.76  | U         | 0.76 | 5.60       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.78  | U         | 0.78 | 5.60       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.93  | U         | 0.93 | 5.60       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.10  | U         | 8.10 | 27.8       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.88  | U         | 0.88 | 5.60       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.76  | U         | 0.76 | 5.60       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.90  | U         | 0.90 | 5.60       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.74  | U         | 0.74 | 5.60       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.83  | U         | 0.83 | 5.60       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.89  | U         | 0.89 | 5.60       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.73  | U         | 0.73 | 5.60       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.93  | U         | 0.93 | 5.60       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.81  | U         | 0.81 | 5.60       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.72  | U         | 0.72 | 5.60       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.78  | U         | 0.78 | 5.60       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.10  | U         | 5.10 | 27.8       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.70  | U         | 0.70 | 5.60       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.82  | U         | 0.82 | 5.60       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.79  | U         | 0.79 | 5.60       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-38ARE                                   | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-08RE                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 89.3          |
| Sample Wt/Vol:     | 5.04 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012316.D        | 1         |           | 01/23/23 16:11 | VY012323      |

| CAS Number                | Parameter                   | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.96  | U         | 0.96     | 5.60       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 5.20  | U         | 5.20     | 27.8       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.83  | U         | 0.83     | 5.60       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.83  | U         | 0.83     | 5.60       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.84  | U         | 0.84     | 5.60       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.72  | U         | 0.72     | 5.60       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.78  | U         | 0.78     | 5.60       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.60  | U         | 1.60     | 11.1       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.88  | U         | 0.88     | 5.60       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.88  | U         | 0.88     | 5.60       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.90  | U         | 0.90     | 5.60       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.80  | U         | 0.80     | 5.60       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.30  | U         | 1.30     | 5.60       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.74  | U         | 0.74     | 5.60       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.70  | U         | 0.70     | 5.60       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.71  | U         | 0.71     | 5.60       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.40  | U         | 1.40     | 5.60       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1.00  | U         | 1.00     | 5.60       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.10  | U         | 1.10     | 5.60       | ug/Kg             |
| <b>SURROGATES</b>         |                             |       |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 81.0  |           | 50 - 163 | 162%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 26.2  | *         | 54 - 147 | 52%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 66.0  |           | 58 - 134 | 132%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 43.1  |           | 30 - 143 | 86%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |       |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 4900  | 7.795     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 8010  | 8.691     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 7430  | 11.496    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 2570  | 13.428    |          |            |                   |

## Report of Analysis

|                    |                                           |                 |                              |
|--------------------|-------------------------------------------|-----------------|------------------------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23                     |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23                     |
| Client Sample ID:  | B-38ARE                                   | SDG No.:        | O1233                        |
| Lab Sample ID:     | O1233-08RE                                | Matrix:         | SOIL                         |
| Analytical Method: | SW8260                                    | % Solid:        | 89.3                         |
| Sample Wt/Vol:     | 5.04      Units:    g                     | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10                |
| GC Column:         | RXI-624              ID :    0.25         | Level :         | LOW                          |
| Prep Method :      |                                           |                 |                              |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012316.D        | 1         |           | 01/23/23 16:11 | VY012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-09                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 80.7          |
| Sample Wt/Vol:     | 5.01 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075162.D        | 1         |           | 01/20/23 19:17 | VD012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1.10  | U         | 1.10 | 6.20       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.30  | U         | 1.30 | 6.20       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.10  | U         | 1.10 | 6.20       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.40  | U         | 1.40 | 6.20       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.10  | U         | 1.10 | 6.20       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.20  | U         | 1.20 | 6.20       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.89  | U         | 0.89 | 6.20       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 1.10  | U         | 1.10 | 6.20       | ug/Kg             |
| 67-64-1        | Acetone                        | 15.1  | U         | 15.1 | 30.9       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.93  | U         | 0.93 | 6.20       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.20  | U         | 1.20 | 6.20       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.60  | U         | 1.60 | 6.20       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 7.40  | U         | 7.40 | 12.4       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.84  | U         | 0.84 | 6.20       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.87  | U         | 0.87 | 6.20       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 1.00  | U         | 1.00 | 6.20       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 9.00  | U         | 9.00 | 30.9       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.98  | U         | 0.98 | 6.20       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.84  | U         | 0.84 | 6.20       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 1.00  | U         | 1.00 | 6.20       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.83  | U         | 0.83 | 6.20       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.93  | U         | 0.93 | 6.20       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.99  | U         | 0.99 | 6.20       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.82  | U         | 0.82 | 6.20       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 1.00  | U         | 1.00 | 6.20       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.90  | U         | 0.90 | 6.20       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.80  | U         | 0.80 | 6.20       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.87  | U         | 0.87 | 6.20       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.70  | U         | 5.70 | 30.9       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.78  | U         | 0.78 | 6.20       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.92  | U         | 0.92 | 6.20       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.88  | U         | 0.88 | 6.20       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38B                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-09                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 80.7          |
| Sample Wt/Vol:     | 5.01 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075162.D        | 1         |           | 01/20/23 19:17 | VD012023      |

| CAS Number                            | Parameter                   | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|-----------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                               | 1,1,2-Trichloroethane       | 1.10  | U         | 1.10     | 6.20       | ug/Kg             |
| 591-78-6                              | 2-Hexanone                  | 5.80  | U         | 5.80     | 30.9       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane        | 0.93  | U         | 0.93     | 6.20       | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane           | 0.93  | U         | 0.93     | 6.20       | ug/Kg             |
| 127-18-4                              | Tetrachloroethene           | 0.94  | U         | 0.94     | 6.20       | ug/Kg             |
| 108-90-7                              | Chlorobenzene               | 0.80  | U         | 0.80     | 6.20       | ug/Kg             |
| 100-41-4                              | Ethyl Benzene               | 0.87  | U         | 0.87     | 6.20       | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                 | 1.80  | U         | 1.80     | 12.4       | ug/Kg             |
| 95-47-6                               | o-Xylene                    | 0.98  | U         | 0.98     | 6.20       | ug/Kg             |
| 100-42-5                              | Styrene                     | 0.98  | U         | 0.98     | 6.20       | ug/Kg             |
| 75-25-2                               | Bromoform                   | 1.00  | U         | 1.00     | 6.20       | ug/Kg             |
| 98-82-8                               | Isopropylbenzene            | 0.89  | U         | 0.89     | 6.20       | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane   | 1.40  | U         | 1.40     | 6.20       | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene         | 0.83  | U         | 0.83     | 6.20       | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene         | 0.78  | U         | 0.78     | 6.20       | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene         | 0.79  | U         | 0.79     | 6.20       | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane | 1.50  | U         | 1.50     | 6.20       | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene      | 1.20  | U         | 1.20     | 6.20       | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene      | 1.20  | U         | 1.20     | 6.20       | ug/Kg             |
| <b>SURROGATES</b>                     |                             |       |           |          |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4       | 74.0  |           | 50 - 163 | 148%       | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane        | 54.8  |           | 54 - 147 | 110%       | SPK: 50           |
| 2037-26-5                             | Toluene-d8                  | 54.3  |           | 58 - 134 | 109%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene        | 60.4  |           | 30 - 143 | 121%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                             |       |           |          |            |                   |
| 363-72-4                              | Pentafluorobenzene          | 41900 | 7.881     |          |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene         | 88400 | 8.781     |          |            |                   |
| 3114-55-4                             | Chlorobenzene-d5            | 88300 | 11.586    |          |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4      | 40700 | 13.522    |          |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                             |       |           |          |            |                   |
| 000111-65-9                           | Octane                      | 6.70  | J         |          | 10.5       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |                              |
|--------------------|-------------------------------------------|-----------------|------------------------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23                     |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23                     |
| Client Sample ID:  | B-P38B                                    | SDG No.:        | O1233                        |
| Lab Sample ID:     | O1233-09                                  | Matrix:         | SOIL                         |
| Analytical Method: | SW8260                                    | % Solid:        | 80.7                         |
| Sample Wt/Vol:     | 5.01      Units:    g                     | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10                |
| GC Column:         | RTX-VMS      ID :    0.18                 | Level :         | LOW                          |
| Prep Method :      |                                           |                 |                              |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075162.D        | 1         |           | 01/20/23 19:17 | VD012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|------------|-----------|-------|-----------|-----|------------|-------------------|
|------------|-----------|-------|-----------|-----|------------|-------------------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38BRE                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-09RE                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 80.7          |
| Sample Wt/Vol:     | 5.05 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075178.D        | 1         |           | 01/23/23 16:23 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1.10  | U         | 1.10 | 6.10       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.30  | U         | 1.30 | 6.10       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1.10  | U         | 1.10 | 6.10       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.40  | U         | 1.40 | 6.10       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1.10  | U         | 1.10 | 6.10       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1.20  | U         | 1.20 | 6.10       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.88  | U         | 0.88 | 6.10       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 1.10  | U         | 1.10 | 6.10       | ug/Kg             |
| 67-64-1        | Acetone                        | 15.0  | U         | 15.0 | 30.7       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.92  | U         | 0.92 | 6.10       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1.10  | U         | 1.10 | 6.10       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.50  | U         | 1.50 | 6.10       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 7.30  | U         | 7.30 | 12.3       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.83  | U         | 0.83 | 6.10       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.86  | U         | 0.86 | 6.10       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 1.00  | U         | 1.00 | 6.10       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8.90  | U         | 8.90 | 30.7       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.97  | U         | 0.97 | 6.10       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.83  | U         | 0.83 | 6.10       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.99  | U         | 0.99 | 6.10       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.82  | U         | 0.82 | 6.10       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.92  | U         | 0.92 | 6.10       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.98  | U         | 0.98 | 6.10       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.81  | U         | 0.81 | 6.10       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 1.00  | U         | 1.00 | 6.10       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.90  | U         | 0.90 | 6.10       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.80  | U         | 0.80 | 6.10       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.86  | U         | 0.86 | 6.10       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5.60  | U         | 5.60 | 30.7       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.77  | U         | 0.77 | 6.10       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.91  | U         | 0.91 | 6.10       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.87  | U         | 0.87 | 6.10       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38BRE                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-09RE                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 80.7          |
| Sample Wt/Vol:     | 5.05 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075178.D        | 1         |           | 01/23/23 16:23 | VD012323      |

| CAS Number                | Parameter                   | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|-------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 1.10  | U         | 1.10     | 6.10       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 5.70  | U         | 5.70     | 30.7       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.92  | U         | 0.92     | 6.10       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.92  | U         | 0.92     | 6.10       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.93  | U         | 0.93     | 6.10       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.80  | U         | 0.80     | 6.10       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.86  | U         | 0.86     | 6.10       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.80  | U         | 1.80     | 12.3       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.97  | U         | 0.97     | 6.10       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.97  | U         | 0.97     | 6.10       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.99  | U         | 0.99     | 6.10       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.88  | U         | 0.88     | 6.10       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.40  | U         | 1.40     | 6.10       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.82  | U         | 0.82     | 6.10       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.77  | U         | 0.77     | 6.10       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.79  | U         | 0.79     | 6.10       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.50  | U         | 1.50     | 6.10       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1.20  | U         | 1.20     | 6.10       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.20  | U         | 1.20     | 6.10       | ug/Kg             |
| <b>SURROGATES</b>         |                             |       |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 59.5  |           | 50 - 163 | 119%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 48.3  |           | 54 - 147 | 97%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 56.8  |           | 58 - 134 | 114%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 50.4  |           | 30 - 143 | 101%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |       |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 1300  | 7.875     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 2140  | 8.775     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 1970  | 11.587    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 699   | 13.528    |          |            |                   |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38BRE                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-09RE                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 80.7          |
| Sample Wt/Vol:     | 5.05 Units: g                             | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075178.D        | 1         |           | 01/23/23 16:23 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |        |      |                 |               |    |
|--------------------|-------------------------------------------|--------|------|-----------------|---------------|----|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |        |      | Date Collected: | 01/18/23      |    |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |        |      | Date Received:  | 01/19/23      |    |
| Client Sample ID:  | B-P38C                                    |        |      | SDG No.:        | O1233         |    |
| Lab Sample ID:     | O1233-11                                  |        |      | Matrix:         | SOIL          |    |
| Analytical Method: | SW8260                                    |        |      | % Solid:        | 85.9          |    |
| Sample Wt/Vol:     | 5.03                                      | Units: | g    | Final Vol:      | 10000         | uL |
| Soil Aliquot Vol:  | 100                                       |        | uL   | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624                                   | ID :   | 0.25 | Level :         | MED           |    |
| Prep Method :      |                                           |        |      |                 |               |    |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076364.D        | 1         |           | 01/23/23 14:12 | VN012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 99.5  | U         | 99.5 | 580        | ug/Kg             |
| 74-87-3        | Chloromethane                  | 120   | U         | 120  | 580        | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 110   | U         | 110  | 580        | ug/Kg             |
| 74-83-9        | Bromomethane                   | 130   | U         | 130  | 580        | ug/Kg             |
| 75-00-3        | Chloroethane                   | 100   | U         | 100  | 580        | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 110   | U         | 110  | 580        | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 83.3  | U         | 83.3 | 580        | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 99.5  | U         | 99.5 | 580        | ug/Kg             |
| 67-64-1        | Acetone                        | 1400  | U         | 1400 | 2900       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 86.8  | U         | 86.8 | 580        | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 110   | U         | 110  | 580        | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 150   | U         | 150  | 580        | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 690   | U         | 690  | 1200       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 78.7  | U         | 78.7 | 580        | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 81.0  | U         | 81.0 | 580        | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 4600  |           | 97.2 | 580        | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 840   | U         | 840  | 2900       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 91.4  | U         | 91.4 | 580        | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 78.7  | U         | 78.7 | 580        | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 93.7  | U         | 93.7 | 580        | ug/Kg             |
| 67-66-3        | Chloroform                     | 77.5  | U         | 77.5 | 580        | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 86.8  | U         | 86.8 | 580        | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 31500 | E         | 92.6 | 580        | ug/Kg             |
| 71-43-2        | Benzene                        | 76.4  | U         | 76.4 | 580        | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 97.2  | U         | 97.2 | 580        | ug/Kg             |
| 79-01-6        | Trichloroethene                | 84.5  | U         | 84.5 | 580        | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 75.2  | U         | 75.2 | 580        | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 81.0  | U         | 81.0 | 580        | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 530   | U         | 530  | 2900       | ug/Kg             |
| 108-88-3       | Toluene                        | 72.9  | U         | 72.9 | 580        | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 85.6  | U         | 85.6 | 580        | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 82.2  | U         | 82.2 | 580        | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38C                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-11                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.9          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100 uL                                    | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | MED           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076364.D        | 1         |           | 01/23/23 14:12 | VN012323      |

| CAS Number                            | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                               | 1,1,2-Trichloroethane       | 99.5   | U         | 99.5     | 580        | ug/Kg             |
| 591-78-6                              | 2-Hexanone                  | 540    | U         | 540      | 2900       | ug/Kg             |
| 124-48-1                              | Dibromochloromethane        | 86.8   | U         | 86.8     | 580        | ug/Kg             |
| 106-93-4                              | 1,2-Dibromoethane           | 86.8   | U         | 86.8     | 580        | ug/Kg             |
| 127-18-4                              | Tetrachloroethene           | 87.9   | U         | 87.9     | 580        | ug/Kg             |
| 108-90-7                              | Chlorobenzene               | 75.2   | U         | 75.2     | 580        | ug/Kg             |
| 100-41-4                              | Ethyl Benzene               | 860    |           | 81.0     | 580        | ug/Kg             |
| 179601-23-1                           | m/p-Xylenes                 | 2400   |           | 170      | 1200       | ug/Kg             |
| 95-47-6                               | o-Xylene                    | 320    | J         | 91.4     | 580        | ug/Kg             |
| 100-42-5                              | Styrene                     | 91.4   | U         | 91.4     | 580        | ug/Kg             |
| 75-25-2                               | Bromoform                   | 93.7   | U         | 93.7     | 580        | ug/Kg             |
| 98-82-8                               | Isopropylbenzene            | 1500   |           | 83.3     | 580        | ug/Kg             |
| 79-34-5                               | 1,1,2,2-Tetrachloroethane   | 130    | U         | 130      | 580        | ug/Kg             |
| 541-73-1                              | 1,3-Dichlorobenzene         | 77.5   | U         | 77.5     | 580        | ug/Kg             |
| 106-46-7                              | 1,4-Dichlorobenzene         | 72.9   | U         | 72.9     | 580        | ug/Kg             |
| 95-50-1                               | 1,2-Dichlorobenzene         | 74.1   | U         | 74.1     | 580        | ug/Kg             |
| 96-12-8                               | 1,2-Dibromo-3-Chloropropane | 140    | U         | 140      | 580        | ug/Kg             |
| 120-82-1                              | 1,2,4-Trichlorobenzene      | 110    | U         | 110      | 580        | ug/Kg             |
| 87-61-6                               | 1,2,3-Trichlorobenzene      | 120    | U         | 120      | 580        | ug/Kg             |
| <b>SURROGATES</b>                     |                             |        |           |          |            |                   |
| 17060-07-0                            | 1,2-Dichloroethane-d4       | 47.8   |           | 50 - 163 | 96%        | SPK: 50           |
| 1868-53-7                             | Dibromofluoromethane        | 45.9   |           | 54 - 147 | 92%        | SPK: 50           |
| 2037-26-5                             | Toluene-d8                  | 65.5   |           | 58 - 134 | 131%       | SPK: 50           |
| 460-00-4                              | 4-Bromofluorobenzene        | 91.1   | *         | 30 - 143 | 182%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b>             |                             |        |           |          |            |                   |
| 363-72-4                              | Pentafluorobenzene          | 380000 | 8.224     |          |            |                   |
| 540-36-3                              | 1,4-Difluorobenzene         | 594000 | 9.1       |          |            |                   |
| 3114-55-4                             | Chlorobenzene-d5            | 545000 | 11.871    |          |            |                   |
| 3855-82-1                             | 1,4-Dichlorobenzene-d4      | 319000 | 13.794    |          |            |                   |
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                             |        |           |          |            |                   |
| 000142-82-5                           | Heptane                     | 22000  | J         |          | 8.96       | ug/Kg             |
| 000592-27-8                           | Heptane, 2-methyl-          | 38200  | J         |          | 10.2       | ug/Kg             |
| 000589-81-1                           | Heptane, 3-methyl-          | 22900  | J         |          | 10.3       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38C                                    | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-11                                  | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.9          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100 uL                                    | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | MED           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076364.D        | 1         |           | 01/23/23 14:12 | VN012323      |

| CAS Number  | Parameter              | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|-------------|------------------------|-------|-----------|-----|------------|-------------------|
| 000111-65-9 | Octane                 | 54700 | J         |     | 10.7       | ug/Kg             |
| 003221-61-2 | Octane, 2-methyl-      | 49000 | J         |     | 11.6       | ug/Kg             |
| 002216-33-3 | Octane, 3-methyl-      | 30700 | J         |     | 11.7       | ug/Kg             |
| 005911-04-6 | Nonane, 3-methyl-      | 28400 | J         |     | 12.5       | ug/Kg             |
|             | unknown12.600          | 47300 | J         |     | 12.6       | ug/Kg             |
| 017301-94-9 | Nonane, 4-methyl-      | 27300 | J         |     | 12.8       | ug/Kg             |
| 103-65-1    | n-propylbenzene        | 3300  | J         |     | 13.0       | ug/Kg             |
| 108-67-8    | 1,3,5-Trimethylbenzene | 4000  | J         |     | 13.2       | ug/Kg             |
| 95-63-6     | 1,2,4-Trimethylbenzene | 7700  | J         |     | 13.5       | ug/Kg             |
| 135-98-8    | sec-Butylbenzene       | 1300  | J         |     | 13.6       | ug/Kg             |
| 99-87-6     | p-Isopropyltoluene     | 2600  | J         |     | 13.7       | ug/Kg             |
| 104-51-8    | n-Butylbenzene         | 3700  | J         |     | 14.1       | ug/Kg             |
| 000824-90-8 | 1-Phenyl-1-butene      | 31200 | J         |     | 15.0       | ug/Kg             |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38CDL                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-11DL                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.9          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100 uL                                    | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | MED           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076367.D        | 10        |           | 01/23/23 15:25 | VN012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|-------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |       |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1000  | UD        | 1000  | 5800       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1200  | UD        | 1200  | 5800       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 1100  | UD        | 1100  | 5800       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1300  | UD        | 1300  | 5800       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 1000  | UD        | 1000  | 5800       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1100  | UD        | 1100  | 5800       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 830   | UD        | 830   | 5800       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 1000  | UD        | 1000  | 5800       | ug/Kg             |
| 67-64-1        | Acetone                        | 14100 | UD        | 14100 | 28900      | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 870   | UD        | 870   | 5800       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1100  | UD        | 1100  | 5800       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1500  | UD        | 1500  | 5800       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6900  | UD        | 6900  | 11600      | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 790   | UD        | 790   | 5800       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 810   | UD        | 810   | 5800       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 6300  | D         | 970   | 5800       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 8400  | UD        | 8400  | 28900      | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 910   | UD        | 910   | 5800       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 790   | UD        | 790   | 5800       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 940   | UD        | 940   | 5800       | ug/Kg             |
| 67-66-3        | Chloroform                     | 780   | UD        | 780   | 5800       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 870   | UD        | 870   | 5800       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 33800 | D         | 930   | 5800       | ug/Kg             |
| 71-43-2        | Benzene                        | 760   | UD        | 760   | 5800       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 970   | UD        | 970   | 5800       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 840   | UD        | 840   | 5800       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 750   | UD        | 750   | 5800       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 810   | UD        | 810   | 5800       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 5300  | UD        | 5300  | 28900      | ug/Kg             |
| 108-88-3       | Toluene                        | 730   | UD        | 730   | 5800       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 860   | UD        | 860   | 5800       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 820   | UD        | 820   | 5800       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38CDL                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-11DL                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.9          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100 uL                                    | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | MED           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076367.D        | 10        |           | 01/23/23 15:25 | VN012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 1000   | UD        | 1000     | 5800       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 5400   | UD        | 5400     | 28900      | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 870    | UD        | 870      | 5800       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 870    | UD        | 870      | 5800       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 880    | UD        | 880      | 5800       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 750    | UD        | 750      | 5800       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 1000   | JD        | 810      | 5800       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 2800   | JD        | 1700     | 11600      | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 910    | UD        | 910      | 5800       | ug/Kg             |
| 100-42-5                  | Styrene                     | 910    | UD        | 910      | 5800       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 940    | UD        | 940      | 5800       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 2100   | JD        | 830      | 5800       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1300   | UD        | 1300     | 5800       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 780    | UD        | 780      | 5800       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 730    | UD        | 730      | 5800       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 740    | UD        | 740      | 5800       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1400   | UD        | 1400     | 5800       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1100   | UD        | 1100     | 5800       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1200   | UD        | 1200     | 5800       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 48.6   |           | 50 - 163 | 97%        | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 46.8   |           | 54 - 147 | 94%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 52.4   |           | 58 - 134 | 105%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 56.2   |           | 30 - 143 | 112%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 385000 | 8.23      |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 613000 | 9.106     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 549000 | 11.865    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 255000 | 13.794    |          |            |                   |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | B-P38CDL                                  | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-11DL                                | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 85.9          |
| Sample Wt/Vol:     | 5.03 Units: g                             | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100 uL                                    | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | MED           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076367.D        | 10        |           | 01/23/23 15:25 | VN012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |        |      |                 |               |    |
|--------------------|-------------------------------------------|--------|------|-----------------|---------------|----|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |        |      | Date Collected: | 01/18/23      |    |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |        |      | Date Received:  | 01/19/23      |    |
| Client Sample ID:  | FB01                                      |        |      | SDG No.:        | O1233         |    |
| Lab Sample ID:     | O1233-12                                  |        |      | Matrix:         | Water         |    |
| Analytical Method: | SW8260                                    |        |      | % Solid:        | 0             |    |
| Sample Wt/Vol:     | 5                                         | Units: | mL   | Final Vol:      | 5000          | uL |
| Soil Aliquot Vol:  |                                           |        | uL   | Test:           | VOC-TCLVOA-10 |    |
| GC Column:         | RXI-624                                   | ID :   | 0.25 | Level :         | LOW           |    |
| Prep Method :      |                                           |        |      |                 |               |    |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076347.D        | 1         |           | 01/20/23 14:49 | VN012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|------|------------|-------|
| <b>TARGETS</b> |                                |       |           |      |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 0.22  | U         | 0.22 | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 1.60  | U         | 1.60 | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 0.26  | U         | 0.26 | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 0.23  | U         | 0.23 | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 1.20  | U         | 1.20 | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 0.26  | U         | 0.26 | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 0.53  | U         | 0.53 | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 2.80  |           | 0.18 | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 1.20  | U         | 1.20 | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 0.82  | U         | 0.82 | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 0.19  | U         | 0.19 | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 0.13  | U         | 0.13 | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 0.16  | U         | 0.16 | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 0.27  | U         | 0.27 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 0.87  | U         | 0.87 | 5.00       | ug/L  |
| 108-88-3       | Toluene                        | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.14  | U         | 0.14 | 1.00       | ug/L  |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.16  | U         | 0.16 | 1.00       | ug/L  |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  | 01/19/23      |
| Client Sample ID:  | FB01                                      | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-12                                  | Matrix:         | Water         |
| Analytical Method: | SW8260                                    | % Solid:        | 0             |
| Sample Wt/Vol:     | 5 Units: mL                               | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076347.D        | 1         |           | 01/20/23 14:49 | VN012023      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 0.76   | U         | 0.76     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 0.14   | U         | 0.14     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 0.33   | U         | 0.33     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 0.13   | U         | 0.13     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 0.23   | U         | 0.23     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 0.42   | U         | 0.42     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.23   | U         | 0.23     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 0.33   | U         | 0.33     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.1   |           | 74 - 125 | 98%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 48.1   |           | 75 - 124 | 96%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 49.6   |           | 86 - 113 | 99%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 48.0   |           | 64 - 133 | 96%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 370000 | 8.23      |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 596000 | 9.106     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 524000 | 11.865    |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 212000 | 13.794    |          |            |         |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: | 01/18/23      |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  | 01/19/23      |
| Client Sample ID:  | FB01                                      |           | SDG No.:        | O1233         |
| Lab Sample ID:     | O1233-12                                  |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076347.D        | 1         |           | 01/20/23 14:49 | VN012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

# QC SUMMARY

### Surrogate Summary

SDG No.: O1233

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8260D

| Lab Sample ID | Client ID | Parameter             | Spike | Result | RecoveryQual | Limits |      |
|---------------|-----------|-----------------------|-------|--------|--------------|--------|------|
|               |           |                       |       |        |              | Low    | High |
| O1233-02      | B-P34A    | 1,2-Dichloroethane-d4 | 50    | 58.5   | 117          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 51.4   | 103          | 54     | 147  |
|               |           | Toluene-d8            | 50    | 52.1   | 104          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 53.8   | 108          | 30     | 143  |
| O1233-03      | B-P34B    | 1,2-Dichloroethane-d4 | 50    | 66.4   | 133          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 53.5   | 107          | 54     | 147  |
|               |           | Toluene-d8            | 50    | 52.7   | 105          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 56.2   | 112          | 30     | 143  |
| O1233-04      | B-P36A    | 1,2-Dichloroethane-d4 | 50    | 49.8   | 100          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 47.1   | 94           | 54     | 147  |
|               |           | Toluene-d8            | 50    | 51.5   | 103          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 50.5   | 101          | 30     | 143  |
| O1233-05      | B-P36B    | 1,2-Dichloroethane-d4 | 50    | 92.5   | 185 *        | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 56.1   | 112          | 54     | 147  |
|               |           | Toluene-d8            | 50    | 51.5   | 103          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 60.1   | 120          | 30     | 143  |
| O1233-05RE    | B-P36BRE  | 1,2-Dichloroethane-d4 | 50    | 68.1   | 136          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 52.3   | 105          | 54     | 147  |
|               |           | Toluene-d8            | 50    | 48.9   | 98           | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 45.8   | 92           | 30     | 143  |
| O1233-06      | B-P37A    | 1,2-Dichloroethane-d4 | 50    | 50.8   | 102          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 48.2   | 96           | 54     | 147  |
|               |           | Toluene-d8            | 50    | 51.4   | 103          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 50.0   | 100          | 30     | 143  |
| O1233-07      | B-P37B    | 1,2-Dichloroethane-d4 | 50    | 83.0   | 166 *        | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 60.1   | 120          | 54     | 147  |
|               |           | Toluene-d8            | 50    | 66.9   | 134          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 48.8   | 98           | 30     | 143  |
| O1233-08      | B-38A     | 1,2-Dichloroethane-d4 | 50    | 90.7   | 181 *        | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 55.9   | 112          | 54     | 147  |
|               |           | Toluene-d8            | 50    | 55.2   | 110          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 56.9   | 114          | 30     | 143  |
| O1233-08RE    | B-38ARE   | 1,2-Dichloroethane-d4 | 50    | 81.0   | 162          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 26.2   | 52 *         | 54     | 147  |
|               |           | Toluene-d8            | 50    | 66.0   | 132          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 43.1   | 86           | 30     | 143  |
| O1233-09      | B-P38B    | 1,2-Dichloroethane-d4 | 50    | 74.0   | 148          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 54.8   | 110          | 54     | 147  |
|               |           | Toluene-d8            | 50    | 54.3   | 109          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 60.4   | 121          | 30     | 143  |
| O1233-09RE    | B-P38BRE  | 1,2-Dichloroethane-d4 | 50    | 59.5   | 119          | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 48.3   | 97           | 54     | 147  |
|               |           | Toluene-d8            | 50    | 56.8   | 114          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 50.4   | 101          | 30     | 143  |
| O1233-11      | B-P38C    | 1,2-Dichloroethane-d4 | 50    | 47.8   | 96           | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 45.9   | 92           | 54     | 147  |
|               |           | Toluene-d8            | 50    | 65.5   | 131          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 91.1   | 182 *        | 30     | 143  |
| O1233-11DL    | B-P38CDL  | 1,2-Dichloroethane-d4 | 50    | 48.6   | 97           | 50     | 163  |
|               |           | Dibromofluoromethane  | 50    | 46.9   | 94           | 54     | 147  |
|               |           | Toluene-d8            | 50    | 52.4   | 105          | 58     | 134  |
|               |           | 4-Bromofluorobenzene  | 50    | 56.2   | 112          | 30     | 143  |

### Surrogate Summary

SDG No.: 01233

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8260D

| Lab Sample ID | Client ID    | Parameter             | Spike | Result | RecoveryQual | Limits |      |
|---------------|--------------|-----------------------|-------|--------|--------------|--------|------|
|               |              |                       |       |        |              | Low    | High |
| VD0120SBL04   | VD0120SBL04  | 1,2-Dichloroethane-d4 | 50    | 57.2   | 114          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 50.7   | 101          | 54     | 147  |
|               |              | Toluene-d8            | 50    | 52.1   | 104          | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 53.6   | 107          | 30     | 143  |
| VD0120SBS02   | VD0120SBS02  | 1,2-Dichloroethane-d4 | 50    | 56.1   | 112          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 53.0   | 106          | 54     | 147  |
|               |              | Toluene-d8            | 50    | 53.3   | 107          | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 53.5   | 107          | 30     | 143  |
| VD0123SBL01   | VD0123SBL01  | 1,2-Dichloroethane-d4 | 50    | 51.9   | 104          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 48.5   | 97           | 54     | 147  |
|               |              | Toluene-d8            | 50    | 50.8   | 102          | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 50.5   | 101          | 30     | 143  |
| VD0123SBS01   | VD0123SBS01  | 1,2-Dichloroethane-d4 | 50    | 54.9   | 110          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 53.1   | 106          | 54     | 147  |
|               |              | Toluene-d8            | 50    | 53.7   | 107          | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 51.6   | 103          | 30     | 143  |
| VD0123SBSD01  | VD0123SBSD01 | 1,2-Dichloroethane-d4 | 50    | 53.7   | 107          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 51.7   | 103          | 54     | 147  |
|               |              | Toluene-d8            | 50    | 52.9   | 106          | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 51.2   | 102          | 30     | 143  |
| VN0123MBL01   | VN0123MBL01  | 1,2-Dichloroethane-d4 | 50    | 49.9   | 100          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 49.5   | 99           | 54     | 147  |
|               |              | Toluene-d8            | 50    | 48.9   | 98           | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 44.4   | 89           | 30     | 143  |
| VN0123MBS01   | VN0123MBS01  | 1,2-Dichloroethane-d4 | 50    | 52.1   | 104          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 51.5   | 103          | 54     | 147  |
|               |              | Toluene-d8            | 50    | 50.9   | 102          | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 51.3   | 103          | 30     | 143  |
| VY0123SBL01   | VY0123SBL01  | 1,2-Dichloroethane-d4 | 50    | 52.7   | 105          | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 45.4   | 91           | 54     | 147  |
|               |              | Toluene-d8            | 50    | 47.0   | 94           | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 41.2   | 82           | 30     | 143  |
| VY0123SBS01   | VY0123SBS01  | 1,2-Dichloroethane-d4 | 50    | 49.4   | 99           | 50     | 163  |
|               |              | Dibromofluoromethane  | 50    | 46.3   | 93           | 54     | 147  |
|               |              | Toluene-d8            | 50    | 48.5   | 97           | 58     | 134  |
|               |              | 4-Bromofluorobenzene  | 50    | 44.5   | 89           | 30     | 143  |



Surrogate Summary

SDG No.: O1233  
Client: Louis Berger U.S., Inc., A WSP Company  
Analytical Method: SW8260-Low

| Lab Sample ID | Client ID    | Parameter             | Spike | Result | RecoveryQual | Limits |      |
|---------------|--------------|-----------------------|-------|--------|--------------|--------|------|
|               |              |                       |       |        |              | Low    | High |
| O1233-12      | FB01         | 1,2-Dichloroethane-d4 | 50    | 49.1   | 98           | 74     | 125  |
|               |              | Dibromofluoromethane  | 50    | 48.1   | 96           | 75     | 124  |
|               |              | Toluene-d8            | 50    | 49.6   | 99           | 86     | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 48.0   | 96           | 64     | 133  |
| VN0120WBL01   | VN0120WBL01  | 1,2-Dichloroethane-d4 | 50    | 47.5   | 95           | 74     | 125  |
|               |              | Dibromofluoromethane  | 50    | 48.7   | 97           | 75     | 124  |
|               |              | Toluene-d8            | 50    | 49.7   | 99           | 86     | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 45.5   | 91           | 64     | 133  |
| VN0120WBS01   | VN0120WBS01  | 1,2-Dichloroethane-d4 | 50    | 49.0   | 98           | 74     | 125  |
|               |              | Dibromofluoromethane  | 50    | 50.3   | 101          | 75     | 124  |
|               |              | Toluene-d8            | 50    | 50.7   | 101          | 86     | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 50.9   | 102          | 64     | 133  |
| VN0120WBSD0   | VN0120WBSD01 | 1,2-Dichloroethane-d4 | 50    | 51.3   | 103          | 74     | 125  |
|               |              | Dibromofluoromethane  | 50    | 52.0   | 104          | 75     | 124  |
|               |              | Toluene-d8            | 50    | 51.4   | 103          | 86     | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 53.4   | 107          | 64     | 133  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**  
**SW-846**

**SDG No.:** O1233

**Client:** Louis Berger U.S., Inc., A WSP Company

**Analytical Method:** SW8260D

**Datafile :** VD075150.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                                |       |        |       |     |     |      |     | High   | RPD |
| VD0120SBS02   | Dichlorodifluoromethane        | 20    | 25.8   | ug/Kg | 129 |     |      | 64  | 136    |     |
|               | Chloromethane                  | 20    | 22.6   | ug/Kg | 113 |     |      | 70  | 130    |     |
|               | Vinyl chloride                 | 20    | 20.1   | ug/Kg | 101 |     |      | 72  | 129    |     |
|               | Bromomethane                   | 20    | 21.1   | ug/Kg | 106 |     |      | 58  | 141    |     |
|               | Chloroethane                   | 20    | 20.7   | ug/Kg | 104 |     |      | 69  | 130    |     |
|               | Trichlorofluoromethane         | 20    | 21.2   | ug/Kg | 106 |     |      | 69  | 134    |     |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 21.7   | ug/Kg | 109 |     |      | 81  | 123    |     |
|               | 1,1-Dichloroethene             | 20    | 22.5   | ug/Kg | 113 |     |      | 79  | 121    |     |
|               | Acetone                        | 100   | 99.8   | ug/Kg | 100 |     |      | 60  | 131    |     |
|               | Carbon disulfide               | 20    | 21.8   | ug/Kg | 109 |     |      | 45  | 154    |     |
|               | Methyl tert-butyl Ether        | 20    | 21.8   | ug/Kg | 109 |     |      | 77  | 129    |     |
|               | Methyl Acetate                 | 20    | 22.8   | ug/Kg | 114 |     |      | 69  | 149    |     |
|               | Methylene Chloride             | 20    | 20.1   | ug/Kg | 101 |     |      | 39  | 175    |     |
|               | trans-1,2-Dichloroethene       | 20    | 21.1   | ug/Kg | 106 |     |      | 80  | 123    |     |
|               | 1,1-Dichloroethane             | 20    | 21.6   | ug/Kg | 108 |     |      | 82  | 123    |     |
|               | Cyclohexane                    | 20    | 21.7   | ug/Kg | 109 |     |      | 76  | 122    |     |
|               | 2-Butanone                     | 100   | 110    | ug/Kg | 110 |     |      | 69  | 131    |     |
|               | Carbon Tetrachloride           | 20    | 20.0   | ug/Kg | 100 |     |      | 76  | 129    |     |
|               | cis-1,2-Dichloroethene         | 20    | 21.6   | ug/Kg | 108 |     |      | 82  | 123    |     |
|               | Bromochloromethane             | 20    | 24.1   | ug/Kg | 121 |     |      | 62  | 134    |     |
|               | Chloroform                     | 20    | 21.5   | ug/Kg | 108 |     |      | 82  | 125    |     |
|               | 1,1,1-Trichloroethane          | 20    | 20.8   | ug/Kg | 104 |     |      | 80  | 126    |     |
|               | Methylcyclohexane              | 20    | 21.0   | ug/Kg | 105 |     |      | 77  | 123    |     |
|               | Benzene                        | 20    | 21.4   | ug/Kg | 107 |     |      | 84  | 121    |     |
|               | 1,2-Dichloroethane             | 20    | 20.9   | ug/Kg | 104 |     |      | 81  | 126    |     |
|               | Trichloroethene                | 20    | 20.6   | ug/Kg | 103 |     |      | 83  | 122    |     |
|               | 1,2-Dichloropropane            | 20    | 20.7   | ug/Kg | 104 |     |      | 83  | 122    |     |
|               | Bromodichloromethane           | 20    | 20.4   | ug/Kg | 102 |     |      | 82  | 123    |     |
|               | 4-Methyl-2-Pentanone           | 100   | 110    | ug/Kg | 110 |     |      | 70  | 135    |     |
|               | Toluene                        | 20    | 20.7   | ug/Kg | 104 |     |      | 83  | 122    |     |
|               | t-1,3-Dichloropropene          | 20    | 20.0   | ug/Kg | 100 |     |      | 78  | 124    |     |
|               | cis-1,3-Dichloropropene        | 20    | 20.7   | ug/Kg | 104 |     |      | 81  | 122    |     |
|               | 1,1,2-Trichloroethane          | 20    | 20.9   | ug/Kg | 104 |     |      | 82  | 125    |     |
|               | 2-Hexanone                     | 100   | 110    | ug/Kg | 110 |     |      | 66  | 138    |     |
|               | Dibromochloromethane           | 20    | 20.0   | ug/Kg | 100 |     |      | 79  | 125    |     |
|               | 1,2-Dibromoethane              | 20    | 21.0   | ug/Kg | 105 |     |      | 80  | 125    |     |
|               | Tetrachloroethene              | 20    | 19.2   | ug/Kg | 96  |     |      | 83  | 125    |     |
|               | Chlorobenzene                  | 20    | 20.4   | ug/Kg | 102 |     |      | 84  | 122    |     |
|               | Ethyl Benzene                  | 20    | 20.4   | ug/Kg | 102 |     |      | 82  | 124    |     |
|               | m/p-Xylenes                    | 40    | 41.6   | ug/Kg | 104 |     |      | 83  | 124    |     |
|               | o-Xylene                       | 20    | 20.9   | ug/Kg | 104 |     |      | 83  | 123    |     |
|               | Styrene                        | 20    | 20.7   | ug/Kg | 104 |     |      | 82  | 124    |     |
|               | Bromoform                      | 20    | 19.9   | ug/Kg | 100 |     |      | 75  | 127    |     |
|               | Isopropylbenzene               | 20    | 20.1   | ug/Kg | 101 |     |      | 82  | 124    |     |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 21.7   | ug/Kg | 109 |     |      | 77  | 127    |     |
|               | 1,3-Dichlorobenzene            | 20    | 19.8   | ug/Kg | 99  |     |      | 83  | 122    |     |
|               | 1,4-Dichlorobenzene            | 20    | 20.8   | ug/Kg | 104 |     |      | 84  | 121    |     |
|               | 1,2-Dichlorobenzene            | 20    | 20.4   | ug/Kg | 102 |     |      | 83  | 124    |     |
|               | 1,2-Dibromo-3-Chloropropane    | 20    | 21.7   | ug/Kg | 109 |     |      | 66  | 134    |     |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233  
Client: Louis Berger U.S., Inc., A WSP Company  
Analytical Method: SW8260D

Datafile : VD075150.D

| Lab Sample ID | Parameter              | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                        |       |        |       |     |     |      |     | High   | RPD |
| VD0120SBS02   | 1,2,4-Trichlorobenzene | 20    | 18.2   | ug/Kg | 91  |     |      | 78  | 127    |     |
|               | 1,2,3-Trichlorobenzene | 20    | 18.4   | ug/Kg | 92  |     |      | 70  | 137    |     |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8260D

Datafile : VD075170.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                                |       |        |       |     |     |      |     | High   | RPD |
| VD0123SBS01   | Dichlorodifluoromethane        | 20    | 24.9   | ug/Kg | 125 |     |      | 64  | 136    |     |
|               | Chloromethane                  | 20    | 22.7   | ug/Kg | 114 |     |      | 70  | 130    |     |
|               | Vinyl chloride                 | 20    | 20.4   | ug/Kg | 102 |     |      | 72  | 129    |     |
|               | Bromomethane                   | 20    | 19.9   | ug/Kg | 100 |     |      | 58  | 141    |     |
|               | Chloroethane                   | 20    | 19.7   | ug/Kg | 99  |     |      | 69  | 130    |     |
|               | Trichlorofluoromethane         | 20    | 21.3   | ug/Kg | 106 |     |      | 69  | 134    |     |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 22.1   | ug/Kg | 111 |     |      | 81  | 123    |     |
|               | 1,1-Dichloroethene             | 20    | 21.7   | ug/Kg | 109 |     |      | 79  | 121    |     |
|               | Acetone                        | 100   | 100    | ug/Kg | 100 |     |      | 60  | 131    |     |
|               | Carbon disulfide               | 20    | 21.8   | ug/Kg | 109 |     |      | 45  | 154    |     |
|               | Methyl tert-butyl Ether        | 20    | 21.1   | ug/Kg | 106 |     |      | 77  | 129    |     |
|               | Methyl Acetate                 | 20    | 22.0   | ug/Kg | 110 |     |      | 69  | 149    |     |
|               | Methylene Chloride             | 20    | 16.3   | ug/Kg | 81  |     |      | 39  | 175    |     |
|               | trans-1,2-Dichloroethene       | 20    | 21.2   | ug/Kg | 106 |     |      | 80  | 123    |     |
|               | 1,1-Dichloroethane             | 20    | 21.5   | ug/Kg | 108 |     |      | 82  | 123    |     |
|               | Cyclohexane                    | 20    | 22.0   | ug/Kg | 110 |     |      | 76  | 122    |     |
|               | 2-Butanone                     | 100   | 100    | ug/Kg | 100 |     |      | 69  | 131    |     |
|               | Carbon Tetrachloride           | 20    | 20.3   | ug/Kg | 102 |     |      | 76  | 129    |     |
|               | cis-1,2-Dichloroethene         | 20    | 20.8   | ug/Kg | 104 |     |      | 82  | 123    |     |
|               | Bromochloromethane             | 20    | 23.2   | ug/Kg | 116 |     |      | 62  | 134    |     |
|               | Chloroform                     | 20    | 20.9   | ug/Kg | 104 |     |      | 82  | 125    |     |
|               | 1,1,1-Trichloroethane          | 20    | 21.0   | ug/Kg | 105 |     |      | 80  | 126    |     |
|               | Methylcyclohexane              | 20    | 20.8   | ug/Kg | 104 |     |      | 77  | 123    |     |
|               | Benzene                        | 20    | 21.1   | ug/Kg | 106 |     |      | 84  | 121    |     |
|               | 1,2-Dichloroethane             | 20    | 21.3   | ug/Kg | 106 |     |      | 81  | 126    |     |
|               | Trichloroethene                | 20    | 19.8   | ug/Kg | 99  |     |      | 83  | 122    |     |
|               | 1,2-Dichloropropane            | 20    | 19.9   | ug/Kg | 100 |     |      | 83  | 122    |     |
|               | Bromodichloromethane           | 20    | 20.8   | ug/Kg | 104 |     |      | 82  | 123    |     |
|               | 4-Methyl-2-Pentanone           | 100   | 110    | ug/Kg | 110 |     |      | 70  | 135    |     |
|               | Toluene                        | 20    | 20.8   | ug/Kg | 104 |     |      | 83  | 122    |     |
|               | t-1,3-Dichloropropene          | 20    | 19.8   | ug/Kg | 99  |     |      | 78  | 124    |     |
|               | cis-1,3-Dichloropropene        | 20    | 20.3   | ug/Kg | 102 |     |      | 81  | 122    |     |
|               | 1,1,2-Trichloroethane          | 20    | 21.7   | ug/Kg | 109 |     |      | 82  | 125    |     |
|               | 2-Hexanone                     | 100   | 110    | ug/Kg | 110 |     |      | 66  | 138    |     |
|               | Dibromochloromethane           | 20    | 21.3   | ug/Kg | 106 |     |      | 79  | 125    |     |
|               | 1,2-Dibromoethane              | 20    | 20.6   | ug/Kg | 103 |     |      | 80  | 125    |     |
|               | Tetrachloroethene              | 20    | 19.7   | ug/Kg | 99  |     |      | 83  | 125    |     |
|               | Chlorobenzene                  | 20    | 20.6   | ug/Kg | 103 |     |      | 84  | 122    |     |
|               | Ethyl Benzene                  | 20    | 20.9   | ug/Kg | 104 |     |      | 82  | 124    |     |
|               | m/p-Xylenes                    | 40    | 41.1   | ug/Kg | 103 |     |      | 83  | 124    |     |
|               | o-Xylene                       | 20    | 20.8   | ug/Kg | 104 |     |      | 83  | 123    |     |
|               | Styrene                        | 20    | 20.7   | ug/Kg | 104 |     |      | 82  | 124    |     |
|               | Bromoform                      | 20    | 20.2   | ug/Kg | 101 |     |      | 75  | 127    |     |
|               | Isopropylbenzene               | 20    | 19.9   | ug/Kg | 100 |     |      | 82  | 124    |     |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 20.7   | ug/Kg | 104 |     |      | 77  | 127    |     |
|               | 1,3-Dichlorobenzene            | 20    | 19.9   | ug/Kg | 100 |     |      | 83  | 122    |     |
|               | 1,4-Dichlorobenzene            | 20    | 20.1   | ug/Kg | 101 |     |      | 84  | 121    |     |
|               | 1,2-Dichlorobenzene            | 20    | 19.7   | ug/Kg | 99  |     |      | 83  | 124    |     |
|               | 1,2-Dibromo-3-Chloropropane    | 20    | 20.7   | ug/Kg | 104 |     |      | 66  | 134    |     |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233  
Client: Louis Berger U.S., Inc., A WSP Company  
Analytical Method: SW8260D

Datafile : VD075170.D

| Lab Sample ID | Parameter              | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                        |       |        |       |     |     |      |     | High   | RPD |
| VD0123SBS01   | 1,2,4-Trichlorobenzene | 20    | 18.2   | ug/Kg | 91  |     |      | 78  | 127    |     |
|               | 1,2,3-Trichlorobenzene | 20    | 18.1   | ug/Kg | 91  |     |      | 70  | 137    |     |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**  
**SW-846**

**SDG No.:** O1233

**Client:** Louis Berger U.S., Inc., A WSP Company

**Analytical Method:** SW8260D

**Datafile :** VD075171.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                                |       |        |       |     |     |      |     | High   | RPD |
| VD0123SBSD01  | Dichlorodifluoromethane        | 20    | 26.4   | ug/Kg | 132 | 5   |      | 64  | 136    | 20  |
|               | Chloromethane                  | 20    | 22.1   | ug/Kg | 111 | 3   |      | 70  | 130    | 20  |
|               | Vinyl chloride                 | 20    | 20.9   | ug/Kg | 104 | 2   |      | 72  | 129    | 20  |
|               | Bromomethane                   | 20    | 20.4   | ug/Kg | 102 | 2   |      | 58  | 141    | 20  |
|               | Chloroethane                   | 20    | 21.5   | ug/Kg | 108 | 9   |      | 69  | 130    | 20  |
|               | Trichlorofluoromethane         | 20    | 22.5   | ug/Kg | 113 | 6   |      | 69  | 134    | 20  |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 22.5   | ug/Kg | 113 | 2   |      | 81  | 123    | 20  |
|               | 1,1-Dichloroethene             | 20    | 22.4   | ug/Kg | 112 | 3   |      | 79  | 121    | 20  |
|               | Acetone                        | 100   | 100    | ug/Kg | 100 | 0   |      | 60  | 131    | 20  |
|               | Carbon disulfide               | 20    | 22.3   | ug/Kg | 112 | 3   |      | 45  | 154    | 20  |
|               | Methyl tert-butyl Ether        | 20    | 21.2   | ug/Kg | 106 | 0   |      | 77  | 129    | 20  |
|               | Methyl Acetate                 | 20    | 22.9   | ug/Kg | 115 | 4   |      | 69  | 149    | 20  |
|               | Methylene Chloride             | 20    | 16.9   | ug/Kg | 85  | 5   |      | 39  | 175    | 20  |
|               | trans-1,2-Dichloroethene       | 20    | 21.8   | ug/Kg | 109 | 3   |      | 80  | 123    | 20  |
|               | 1,1-Dichloroethane             | 20    | 21.7   | ug/Kg | 109 | 1   |      | 82  | 123    | 20  |
|               | Cyclohexane                    | 20    | 21.6   | ug/Kg | 108 | 2   |      | 76  | 122    | 20  |
|               | 2-Butanone                     | 100   | 110    | ug/Kg | 110 | 10  |      | 69  | 131    | 20  |
|               | Carbon Tetrachloride           | 20    | 20.2   | ug/Kg | 101 | 1   |      | 76  | 129    | 20  |
|               | cis-1,2-Dichloroethene         | 20    | 21.5   | ug/Kg | 108 | 4   |      | 82  | 123    | 20  |
|               | Bromochloromethane             | 20    | 24.2   | ug/Kg | 121 | 4   |      | 62  | 134    | 20  |
|               | Chloroform                     | 20    | 21.3   | ug/Kg | 106 | 2   |      | 82  | 125    | 20  |
|               | 1,1,1-Trichloroethane          | 20    | 21.3   | ug/Kg | 106 | 1   |      | 80  | 126    | 20  |
|               | Methylcyclohexane              | 20    | 21.1   | ug/Kg | 106 | 2   |      | 77  | 123    | 20  |
|               | Benzene                        | 20    | 21.6   | ug/Kg | 108 | 2   |      | 84  | 121    | 20  |
|               | 1,2-Dichloroethane             | 20    | 21.1   | ug/Kg | 106 | 0   |      | 81  | 126    | 20  |
|               | Trichloroethene                | 20    | 20.0   | ug/Kg | 100 | 1   |      | 83  | 122    | 20  |
|               | 1,2-Dichloropropane            | 20    | 21.3   | ug/Kg | 106 | 6   |      | 83  | 122    | 20  |
|               | Bromodichloromethane           | 20    | 21.5   | ug/Kg | 108 | 4   |      | 82  | 123    | 20  |
|               | 4-Methyl-2-Pentanone           | 100   | 110    | ug/Kg | 110 | 0   |      | 70  | 135    | 20  |
|               | Toluene                        | 20    | 21.3   | ug/Kg | 106 | 2   |      | 83  | 122    | 20  |
|               | t-1,3-Dichloropropene          | 20    | 19.7   | ug/Kg | 99  | 0   |      | 78  | 124    | 20  |
|               | cis-1,3-Dichloropropene        | 20    | 20.4   | ug/Kg | 102 | 0   |      | 81  | 122    | 20  |
|               | 1,1,2-Trichloroethane          | 20    | 21.3   | ug/Kg | 106 | 3   |      | 82  | 125    | 20  |
|               | 2-Hexanone                     | 100   | 110    | ug/Kg | 110 | 0   |      | 66  | 138    | 20  |
|               | Dibromochloromethane           | 20    | 20.9   | ug/Kg | 104 | 2   |      | 79  | 125    | 20  |
|               | 1,2-Dibromoethane              | 20    | 21.2   | ug/Kg | 106 | 3   |      | 80  | 125    | 20  |
|               | Tetrachloroethene              | 20    | 20.1   | ug/Kg | 101 | 2   |      | 83  | 125    | 20  |
|               | Chlorobenzene                  | 20    | 20.5   | ug/Kg | 103 | 0   |      | 84  | 122    | 20  |
|               | Ethyl Benzene                  | 20    | 21.0   | ug/Kg | 105 | 1   |      | 82  | 124    | 20  |
|               | m/p-Xylenes                    | 40    | 42.9   | ug/Kg | 107 | 4   |      | 83  | 124    | 20  |
|               | o-Xylene                       | 20    | 20.4   | ug/Kg | 102 | 2   |      | 83  | 123    | 20  |
|               | Styrene                        | 20    | 21.5   | ug/Kg | 108 | 4   |      | 82  | 124    | 20  |
|               | Bromoform                      | 20    | 20.2   | ug/Kg | 101 | 0   |      | 75  | 127    | 20  |
|               | Isopropylbenzene               | 20    | 20.1   | ug/Kg | 101 | 1   |      | 82  | 124    | 20  |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 21.6   | ug/Kg | 108 | 4   |      | 77  | 127    | 20  |
|               | 1,3-Dichlorobenzene            | 20    | 20.2   | ug/Kg | 101 | 1   |      | 83  | 122    | 20  |
|               | 1,4-Dichlorobenzene            | 20    | 20.6   | ug/Kg | 103 | 2   |      | 84  | 121    | 20  |
|               | 1,2-Dichlorobenzene            | 20    | 20.6   | ug/Kg | 103 | 4   |      | 83  | 124    | 20  |
|               | 1,2-Dibromo-3-Chloropropane    | 20    | 22.7   | ug/Kg | 114 | 9   |      | 66  | 134    | 20  |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233  
Client: Louis Berger U.S., Inc., A WSP Company  
Analytical Method: SW8260D

Datafile : VD075171.D

| Lab Sample ID | Parameter              | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                        |       |        |       |     |     |      |     | High   | RPD |
| VD0123SBSD01  | 1,2,4-Trichlorobenzene | 20    | 19.0   | ug/Kg | 95  | 4   |      | 78  | 127    | 20  |
|               | 1,2,3-Trichlorobenzene | 20    | 18.8   | ug/Kg | 94  | 3   |      | 70  | 137    | 20  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**  
**SW-846**

**SDG No.:** O1233

**Client:** Louis Berger U.S., Inc., A WSP Company

**Analytical Method:** SW8260-Low

**Datafile :** VN076340.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits |     |
|---------------|--------------------------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |                                |       |        |      |     |     |      |     | High   | RPD |
| VN0120WBS01   | Dichlorodifluoromethane        | 20    | 17.4   | ug/L | 87  |     |      | 69  | 116    |     |
|               | Chloromethane                  | 20    | 16.6   | ug/L | 83  |     |      | 65  | 116    |     |
|               | Vinyl chloride                 | 20    | 17.5   | ug/L | 88  |     |      | 65  | 117    |     |
|               | Bromomethane                   | 20    | 17.6   | ug/L | 88  |     |      | 42  | 173    |     |
|               | Chloroethane                   | 20    | 18.8   | ug/L | 94  |     |      | 44  | 165    |     |
|               | Trichlorofluoromethane         | 20    | 16.9   | ug/L | 85  |     |      | 73  | 115    |     |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 17.2   | ug/L | 86  |     |      | 80  | 112    |     |
|               | 1,1-Dichloroethene             | 20    | 16.8   | ug/L | 84  |     |      | 74  | 110    |     |
|               | Acetone                        | 100   | 85.8   | ug/L | 86  |     |      | 60  | 125    |     |
|               | Carbon disulfide               | 20    | 15.2   | ug/L | 76  |     |      | 64  | 112    |     |
|               | Methyl tert-butyl Ether        | 20    | 17.6   | ug/L | 88  |     |      | 78  | 114    |     |
|               | Methyl Acetate                 | 20    | 16.4   | ug/L | 82  |     |      | 67  | 125    |     |
|               | Methylene Chloride             | 20    | 18.5   | ug/L | 93  |     |      | 72  | 114    |     |
|               | trans-1,2-Dichloroethene       | 20    | 16.2   | ug/L | 81  |     |      | 75  | 108    |     |
|               | 1,1-Dichloroethane             | 20    | 16.8   | ug/L | 84  |     |      | 78  | 112    |     |
|               | Cyclohexane                    | 20    | 16.2   | ug/L | 81  |     |      | 75  | 110    |     |
|               | 2-Butanone                     | 100   | 83.8   | ug/L | 84  |     |      | 65  | 122    |     |
|               | Carbon Tetrachloride           | 20    | 17.2   | ug/L | 86  |     |      | 77  | 113    |     |
|               | cis-1,2-Dichloroethene         | 20    | 17.3   | ug/L | 86  |     |      | 77  | 110    |     |
|               | Bromochloromethane             | 20    | 18.8   | ug/L | 94  |     |      | 70  | 124    |     |
|               | Chloroform                     | 20    | 17.0   | ug/L | 85  |     |      | 79  | 113    |     |
|               | 1,1,1-Trichloroethane          | 20    | 17.5   | ug/L | 88  |     |      | 80  | 108    |     |
|               | Methylcyclohexane              | 20    | 17.6   | ug/L | 88  |     |      | 72  | 115    |     |
|               | Benzene                        | 20    | 17.2   | ug/L | 86  |     |      | 82  | 109    |     |
|               | 1,2-Dichloroethane             | 20    | 17.3   | ug/L | 86  |     |      | 80  | 115    |     |
|               | Trichloroethene                | 20    | 16.7   | ug/L | 84  |     |      | 77  | 113    |     |
|               | 1,2-Dichloropropane            | 20    | 17.0   | ug/L | 85  |     |      | 83  | 111    |     |
|               | Bromodichloromethane           | 20    | 17.6   | ug/L | 88  |     |      | 83  | 110    |     |
|               | 4-Methyl-2-Pentanone           | 100   | 88.6   | ug/L | 89  |     |      | 74  | 118    |     |
|               | Toluene                        | 20    | 16.7   | ug/L | 84  |     |      | 82  | 110    |     |
|               | t-1,3-Dichloropropene          | 20    | 17.9   | ug/L | 90  |     |      | 79  | 110    |     |
|               | cis-1,3-Dichloropropene        | 20    | 17.3   | ug/L | 86  |     |      | 82  | 110    |     |
|               | 1,1,2-Trichloroethane          | 20    | 17.5   | ug/L | 88  |     |      | 83  | 112    |     |
|               | 2-Hexanone                     | 100   | 88.3   | ug/L | 88  |     |      | 73  | 117    |     |
|               | Dibromochloromethane           | 20    | 18.0   | ug/L | 90  |     |      | 82  | 110    |     |
|               | 1,2-Dibromoethane              | 20    | 17.8   | ug/L | 89  |     |      | 81  | 110    |     |
|               | Tetrachloroethene              | 20    | 17.6   | ug/L | 88  |     |      | 67  | 123    |     |
|               | Chlorobenzene                  | 20    | 17.8   | ug/L | 89  |     |      | 82  | 109    |     |
|               | Ethyl Benzene                  | 20    | 18.3   | ug/L | 92  |     |      | 83  | 109    |     |
|               | m/p-Xylenes                    | 40    | 36.8   | ug/L | 92  |     |      | 82  | 110    |     |
|               | o-Xylene                       | 20    | 18.5   | ug/L | 93  |     |      | 83  | 109    |     |
|               | Styrene                        | 20    | 18.3   | ug/L | 92  |     |      | 80  | 111    |     |
|               | Bromoform                      | 20    | 18.1   | ug/L | 91  |     |      | 79  | 109    |     |
|               | Isopropylbenzene               | 20    | 19.4   | ug/L | 97  |     |      | 83  | 112    |     |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 18.1   | ug/L | 91  |     |      | 76  | 118    |     |
|               | 1,3-Dichlorobenzene            | 20    | 17.6   | ug/L | 88  |     |      | 82  | 108    |     |
|               | 1,4-Dichlorobenzene            | 20    | 18.0   | ug/L | 90  |     |      | 82  | 107    |     |
|               | 1,2-Dichlorobenzene            | 20    | 17.4   | ug/L | 87  |     |      | 82  | 109    |     |
|               | 1,2-Dibromo-3-Chloropropane    | 20    | 17.9   | ug/L | 90  |     |      | 68  | 112    |     |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233  
Client: Louis Berger U.S., Inc., A WSP Company  
Analytical Method: SW8260-Low

Datafile : VN076340.D

| Lab Sample ID | Parameter              | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits |     |
|---------------|------------------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |                        |       |        |      |     |     |      |     | High   | RPD |
| VN0120WBS01   | 1,2,4-Trichlorobenzene | 20    | 16.7   | ug/L | 84  |     |      | 74  | 114    |     |
|               | 1,2,3-Trichlorobenzene | 20    | 17.1   | ug/L | 86  |     |      | 77  | 113    |     |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**  
**SW-846**

**SDG No.:** O1233

**Client:** Louis Berger U.S., Inc., A WSP Company

**Analytical Method:** SW8260-Low

**Datafile :** VN076341.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits |     |
|---------------|--------------------------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |                                |       |        |      |     |     |      |     | High   | RPD |
| VN0120WBSD01  | Dichlorodifluoromethane        | 20    | 18.1   | ug/L | 91  | 4   |      | 69  | 116    | 20  |
|               | Chloromethane                  | 20    | 17.3   | ug/L | 86  | 4   |      | 65  | 116    | 20  |
|               | Vinyl chloride                 | 20    | 18.3   | ug/L | 92  | 4   |      | 65  | 117    | 20  |
|               | Bromomethane                   | 20    | 18.7   | ug/L | 94  | 7   |      | 42  | 173    | 20  |
|               | Chloroethane                   | 20    | 18.9   | ug/L | 95  | 1   |      | 44  | 165    | 20  |
|               | Trichlorofluoromethane         | 20    | 18.0   | ug/L | 90  | 6   |      | 73  | 115    | 20  |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 18.7   | ug/L | 94  | 9   |      | 80  | 112    | 20  |
|               | 1,1-Dichloroethene             | 20    | 17.4   | ug/L | 87  | 4   |      | 74  | 110    | 20  |
|               | Acetone                        | 100   | 91.0   | ug/L | 91  | 6   |      | 60  | 125    | 20  |
|               | Carbon disulfide               | 20    | 16.3   | ug/L | 81  | 6   |      | 64  | 112    | 20  |
|               | Methyl tert-butyl Ether        | 20    | 19.9   | ug/L | 100 | 13  |      | 78  | 114    | 20  |
|               | Methyl Acetate                 | 20    | 18.5   | ug/L | 93  | 13  |      | 67  | 125    | 20  |
|               | Methylene Chloride             | 20    | 20.5   | ug/L | 103 | 10  |      | 72  | 114    | 20  |
|               | trans-1,2-Dichloroethene       | 20    | 18.0   | ug/L | 90  | 11  |      | 75  | 108    | 20  |
|               | 1,1-Dichloroethane             | 20    | 18.4   | ug/L | 92  | 9   |      | 78  | 112    | 20  |
|               | Cyclohexane                    | 20    | 17.4   | ug/L | 87  | 7   |      | 75  | 110    | 20  |
|               | 2-Butanone                     | 100   | 92.2   | ug/L | 92  | 9   |      | 65  | 122    | 20  |
|               | Carbon Tetrachloride           | 20    | 18.7   | ug/L | 94  | 9   |      | 77  | 113    | 20  |
|               | cis-1,2-Dichloroethene         | 20    | 18.6   | ug/L | 93  | 8   |      | 77  | 110    | 20  |
|               | Bromochloromethane             | 20    | 20.4   | ug/L | 102 | 8   |      | 70  | 124    | 20  |
|               | Chloroform                     | 20    | 18.8   | ug/L | 94  | 10  |      | 79  | 113    | 20  |
|               | 1,1,1-Trichloroethane          | 20    | 18.5   | ug/L | 93  | 6   |      | 80  | 108    | 20  |
|               | Methylcyclohexane              | 20    | 19.1   | ug/L | 96  | 9   |      | 72  | 115    | 20  |
|               | Benzene                        | 20    | 18.8   | ug/L | 94  | 9   |      | 82  | 109    | 20  |
|               | 1,2-Dichloroethane             | 20    | 19.5   | ug/L | 98  | 13  |      | 80  | 115    | 20  |
|               | Trichloroethene                | 20    | 18.0   | ug/L | 90  | 7   |      | 77  | 113    | 20  |
|               | 1,2-Dichloropropane            | 20    | 19.2   | ug/L | 96  | 12  |      | 83  | 111    | 20  |
|               | Bromodichloromethane           | 20    | 19.5   | ug/L | 98  | 11  |      | 83  | 110    | 20  |
|               | 4-Methyl-2-Pentanone           | 100   | 98.9   | ug/L | 99  | 11  |      | 74  | 118    | 20  |
|               | Toluene                        | 20    | 18.5   | ug/L | 93  | 10  |      | 82  | 110    | 20  |
|               | t-1,3-Dichloropropene          | 20    | 19.8   | ug/L | 99  | 10  |      | 79  | 110    | 20  |
|               | cis-1,3-Dichloropropene        | 20    | 19.4   | ug/L | 97  | 12  |      | 82  | 110    | 20  |
|               | 1,1,2-Trichloroethane          | 20    | 19.5   | ug/L | 98  | 11  |      | 83  | 112    | 20  |
|               | 2-Hexanone                     | 100   | 98.1   | ug/L | 98  | 11  |      | 73  | 117    | 20  |
|               | Dibromochloromethane           | 20    | 20.2   | ug/L | 101 | 12  |      | 82  | 110    | 20  |
|               | 1,2-Dibromoethane              | 20    | 19.4   | ug/L | 97  | 9   |      | 81  | 110    | 20  |
|               | Tetrachloroethene              | 20    | 18.4   | ug/L | 92  | 4   |      | 67  | 123    | 20  |
|               | Chlorobenzene                  | 20    | 18.8   | ug/L | 94  | 5   |      | 82  | 109    | 20  |
|               | Ethyl Benzene                  | 20    | 19.1   | ug/L | 96  | 4   |      | 83  | 109    | 20  |
|               | m/p-Xylenes                    | 40    | 38.2   | ug/L | 96  | 4   |      | 82  | 110    | 20  |
|               | o-Xylene                       | 20    | 19.2   | ug/L | 96  | 3   |      | 83  | 109    | 20  |
|               | Styrene                        | 20    | 19.4   | ug/L | 97  | 5   |      | 80  | 111    | 20  |
|               | Bromoform                      | 20    | 19.8   | ug/L | 99  | 8   |      | 79  | 109    | 20  |
|               | Isopropylbenzene               | 20    | 19.9   | ug/L | 100 | 3   |      | 83  | 112    | 20  |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 19.3   | ug/L | 97  | 6   |      | 76  | 118    | 20  |
|               | 1,3-Dichlorobenzene            | 20    | 18.1   | ug/L | 91  | 3   |      | 82  | 108    | 20  |
|               | 1,4-Dichlorobenzene            | 20    | 18.8   | ug/L | 94  | 4   |      | 82  | 107    | 20  |
|               | 1,2-Dichlorobenzene            | 20    | 18.1   | ug/L | 91  | 4   |      | 82  | 109    | 20  |
|               | 1,2-Dibromo-3-Chloropropane    | 20    | 19.9   | ug/L | 100 | 11  |      | 68  | 112    | 20  |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233  
Client: Louis Berger U.S., Inc., A WSP Company  
Analytical Method: SW8260-Low

Datafile : VN076341.D

| Lab Sample ID | Parameter              | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits |     |
|---------------|------------------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |                        |       |        |      |     |     |      |     | High   | RPD |
| VN0120WBSD01  | 1,2,4-Trichlorobenzene | 20    | 18.4   | ug/L | 92  | 9   |      | 74  | 114    | 20  |
|               | 1,2,3-Trichlorobenzene | 20    | 18.5   | ug/L | 93  | 8   |      | 77  | 113    | 20  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**  
**SW-846**

**SDG No.:** O1233

**Client:** Louis Berger U.S., Inc., A WSP Company

**Analytical Method:** SW8260D

**Datafile :** VN076358.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                                |       |        |       |     |     |      |     | High   | RPD |
| VN0123MBS01   | Dichlorodifluoromethane        | 2000  | 1800   | ug/Kg | 90  |     |      | 64  | 136    |     |
|               | Chloromethane                  | 2000  | 1800   | ug/Kg | 90  |     |      | 70  | 130    |     |
|               | Vinyl chloride                 | 2000  | 2000   | ug/Kg | 100 |     |      | 72  | 129    |     |
|               | Bromomethane                   | 2000  | 1900   | ug/Kg | 95  |     |      | 58  | 141    |     |
|               | Chloroethane                   | 2000  | 2000   | ug/Kg | 100 |     |      | 69  | 130    |     |
|               | Trichlorofluoromethane         | 2000  | 1700   | ug/Kg | 85  |     |      | 69  | 134    |     |
|               | 1,1,2-Trichlorotrifluoroethane | 2000  | 1800   | ug/Kg | 90  |     |      | 81  | 123    |     |
|               | 1,1-Dichloroethene             | 2000  | 1700   | ug/Kg | 85  |     |      | 79  | 121    |     |
|               | Acetone                        | 10000 | 10700  | ug/Kg | 107 |     |      | 60  | 131    |     |
|               | Carbon disulfide               | 2000  | 1700   | ug/Kg | 85  |     |      | 45  | 154    |     |
|               | Methyl tert-butyl Ether        | 2000  | 1800   | ug/Kg | 90  |     |      | 77  | 129    |     |
|               | Methyl Acetate                 | 2000  | 1800   | ug/Kg | 90  |     |      | 69  | 149    |     |
|               | Methylene Chloride             | 2000  | 1700   | ug/Kg | 85  |     |      | 39  | 175    |     |
|               | trans-1,2-Dichloroethene       | 2000  | 1700   | ug/Kg | 85  |     |      | 80  | 123    |     |
|               | 1,1-Dichloroethane             | 2000  | 1800   | ug/Kg | 90  |     |      | 82  | 123    |     |
|               | Cyclohexane                    | 2000  | 1700   | ug/Kg | 85  |     |      | 76  | 122    |     |
|               | 2-Butanone                     | 10000 | 9300   | ug/Kg | 93  |     |      | 69  | 131    |     |
|               | Carbon Tetrachloride           | 2000  | 1900   | ug/Kg | 95  |     |      | 76  | 129    |     |
|               | cis-1,2-Dichloroethene         | 2000  | 1700   | ug/Kg | 85  |     |      | 82  | 123    |     |
|               | Bromochloromethane             | 2000  | 2100   | ug/Kg | 105 |     |      | 62  | 134    |     |
|               | Chloroform                     | 2000  | 1800   | ug/Kg | 90  |     |      | 82  | 125    |     |
|               | 1,1,1-Trichloroethane          | 2000  | 1800   | ug/Kg | 90  |     |      | 80  | 126    |     |
|               | Methylcyclohexane              | 2000  | 1700   | ug/Kg | 85  |     |      | 77  | 123    |     |
|               | Benzene                        | 2000  | 1800   | ug/Kg | 90  |     |      | 84  | 121    |     |
|               | 1,2-Dichloroethane             | 2000  | 1800   | ug/Kg | 90  |     |      | 81  | 126    |     |
|               | Trichloroethene                | 2000  | 1700   | ug/Kg | 85  |     |      | 83  | 122    |     |
|               | 1,2-Dichloropropane            | 2000  | 1800   | ug/Kg | 90  |     |      | 83  | 122    |     |
|               | Bromodichloromethane           | 2000  | 1900   | ug/Kg | 95  |     |      | 82  | 123    |     |
|               | 4-Methyl-2-Pentanone           | 10000 | 9500   | ug/Kg | 95  |     |      | 70  | 135    |     |
|               | Toluene                        | 2000  | 1700   | ug/Kg | 85  |     |      | 83  | 122    |     |
|               | t-1,3-Dichloropropene          | 2000  | 1900   | ug/Kg | 95  |     |      | 78  | 124    |     |
|               | cis-1,3-Dichloropropene        | 2000  | 1900   | ug/Kg | 95  |     |      | 81  | 122    |     |
|               | 1,1,2-Trichloroethane          | 2000  | 1800   | ug/Kg | 90  |     |      | 82  | 125    |     |
|               | 2-Hexanone                     | 10000 | 9400   | ug/Kg | 94  |     |      | 66  | 138    |     |
|               | Dibromochloromethane           | 2000  | 2000   | ug/Kg | 100 |     |      | 79  | 125    |     |
|               | 1,2-Dibromoethane              | 2000  | 1800   | ug/Kg | 90  |     |      | 80  | 125    |     |
|               | Tetrachloroethene              | 2000  | 1700   | ug/Kg | 85  |     |      | 83  | 125    |     |
|               | Chlorobenzene                  | 2000  | 1800   | ug/Kg | 90  |     |      | 84  | 122    |     |
|               | Ethyl Benzene                  | 2000  | 1900   | ug/Kg | 95  |     |      | 82  | 124    |     |
|               | m/p-Xylenes                    | 4000  | 3700   | ug/Kg | 93  |     |      | 83  | 124    |     |
|               | o-Xylene                       | 2000  | 1800   | ug/Kg | 90  |     |      | 83  | 123    |     |
|               | Styrene                        | 2000  | 1900   | ug/Kg | 95  |     |      | 82  | 124    |     |
|               | Bromoform                      | 2000  | 2100   | ug/Kg | 105 |     |      | 75  | 127    |     |
|               | Isopropylbenzene               | 2000  | 2000   | ug/Kg | 100 |     |      | 82  | 124    |     |
|               | 1,1,2,2-Tetrachloroethane      | 2000  | 1900   | ug/Kg | 95  |     |      | 77  | 127    |     |
|               | 1,3-Dichlorobenzene            | 2000  | 1800   | ug/Kg | 90  |     |      | 83  | 122    |     |
|               | 1,4-Dichlorobenzene            | 2000  | 1800   | ug/Kg | 90  |     |      | 84  | 121    |     |
|               | 1,2-Dichlorobenzene            | 2000  | 1700   | ug/Kg | 85  |     |      | 83  | 124    |     |
|               | 1,2-Dibromo-3-Chloropropane    | 2000  | 1800   | ug/Kg | 90  |     |      | 66  | 134    |     |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233

Client: Louis Berger U.S., Inc., A WSP Company

Analytical Method: SW8260D

Datafile : VN076358.D

| Lab Sample ID | Parameter              | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                        |       |        |       |     |     |      |     | High   | RPD |
| VN0123MBS01   | 1,2,4-Trichlorobenzene | 2000  | 1700   | ug/Kg | 85  |     |      | 78  | 127    |     |
|               | 1,2,3-Trichlorobenzene | 2000  | 1700   | ug/Kg | 85  |     |      | 70  | 137    |     |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**  
**SW-846**

**SDG No.:** O1233

**Client:** Louis Berger U.S., Inc., A WSP Company

**Analytical Method:** SW8260D

**Datafile :** VY012309.D

| Lab Sample ID | Parameter                      | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|--------------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                                |       |        |       |     |     |      |     | High   | RPD |
| VY0123SBS01   | Dichlorodifluoromethane        | 20    | 21.2   | ug/Kg | 106 |     |      | 64  | 136    |     |
|               | Chloromethane                  | 20    | 19.7   | ug/Kg | 99  |     |      | 70  | 130    |     |
|               | Vinyl chloride                 | 20    | 20.0   | ug/Kg | 100 |     |      | 72  | 129    |     |
|               | Bromomethane                   | 20    | 21.0   | ug/Kg | 105 |     |      | 58  | 141    |     |
|               | Chloroethane                   | 20    | 20.8   | ug/Kg | 104 |     |      | 69  | 130    |     |
|               | Trichlorofluoromethane         | 20    | 21.2   | ug/Kg | 106 |     |      | 69  | 134    |     |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 20.8   | ug/Kg | 104 |     |      | 81  | 123    |     |
|               | 1,1-Dichloroethene             | 20    | 20.6   | ug/Kg | 103 |     |      | 79  | 121    |     |
|               | Acetone                        | 100   | 87.9   | ug/Kg | 88  |     |      | 60  | 131    |     |
|               | Carbon disulfide               | 20    | 20.2   | ug/Kg | 101 |     |      | 45  | 154    |     |
|               | Methyl tert-butyl Ether        | 20    | 19.4   | ug/Kg | 97  |     |      | 77  | 129    |     |
|               | Methyl Acetate                 | 20    | 18.3   | ug/Kg | 92  |     |      | 69  | 149    |     |
|               | Methylene Chloride             | 20    | 15.2   | ug/Kg | 76  |     |      | 39  | 175    |     |
|               | trans-1,2-Dichloroethene       | 20    | 20.1   | ug/Kg | 101 |     |      | 80  | 123    |     |
|               | 1,1-Dichloroethane             | 20    | 20.5   | ug/Kg | 103 |     |      | 82  | 123    |     |
|               | Cyclohexane                    | 20    | 20.2   | ug/Kg | 101 |     |      | 76  | 122    |     |
|               | 2-Butanone                     | 100   | 92.3   | ug/Kg | 92  |     |      | 69  | 131    |     |
|               | Carbon Tetrachloride           | 20    | 20.7   | ug/Kg | 104 |     |      | 76  | 129    |     |
|               | cis-1,2-Dichloroethene         | 20    | 20.5   | ug/Kg | 103 |     |      | 82  | 123    |     |
|               | Bromochloromethane             | 20    | 19.2   | ug/Kg | 96  |     |      | 62  | 134    |     |
|               | Chloroform                     | 20    | 20.4   | ug/Kg | 102 |     |      | 82  | 125    |     |
|               | 1,1,1-Trichloroethane          | 20    | 20.7   | ug/Kg | 104 |     |      | 80  | 126    |     |
|               | Methylcyclohexane              | 20    | 20.8   | ug/Kg | 104 |     |      | 77  | 123    |     |
|               | Benzene                        | 20    | 20.5   | ug/Kg | 103 |     |      | 84  | 121    |     |
|               | 1,2-Dichloroethane             | 20    | 20.2   | ug/Kg | 101 |     |      | 81  | 126    |     |
|               | Trichloroethene                | 20    | 20.1   | ug/Kg | 101 |     |      | 83  | 122    |     |
|               | 1,2-Dichloropropane            | 20    | 20.3   | ug/Kg | 102 |     |      | 83  | 122    |     |
|               | Bromodichloromethane           | 20    | 20.3   | ug/Kg | 102 |     |      | 82  | 123    |     |
|               | 4-Methyl-2-Pentanone           | 100   | 93.5   | ug/Kg | 94  |     |      | 70  | 135    |     |
|               | Toluene                        | 20    | 20.4   | ug/Kg | 102 |     |      | 83  | 122    |     |
|               | t-1,3-Dichloropropene          | 20    | 19.6   | ug/Kg | 98  |     |      | 78  | 124    |     |
|               | cis-1,3-Dichloropropene        | 20    | 20.1   | ug/Kg | 101 |     |      | 81  | 122    |     |
|               | 1,1,2-Trichloroethane          | 20    | 19.6   | ug/Kg | 98  |     |      | 82  | 125    |     |
|               | 2-Hexanone                     | 100   | 92.0   | ug/Kg | 92  |     |      | 66  | 138    |     |
|               | Dibromochloromethane           | 20    | 19.7   | ug/Kg | 99  |     |      | 79  | 125    |     |
|               | 1,2-Dibromoethane              | 20    | 19.3   | ug/Kg | 97  |     |      | 80  | 125    |     |
|               | Tetrachloroethene              | 20    | 21.0   | ug/Kg | 105 |     |      | 83  | 125    |     |
|               | Chlorobenzene                  | 20    | 20.2   | ug/Kg | 101 |     |      | 84  | 122    |     |
|               | Ethyl Benzene                  | 20    | 20.3   | ug/Kg | 102 |     |      | 82  | 124    |     |
|               | m/p-Xylenes                    | 40    | 41.2   | ug/Kg | 103 |     |      | 83  | 124    |     |
|               | o-Xylene                       | 20    | 20.0   | ug/Kg | 100 |     |      | 83  | 123    |     |
|               | Styrene                        | 20    | 20.4   | ug/Kg | 102 |     |      | 82  | 124    |     |
|               | Bromoform                      | 20    | 19.1   | ug/Kg | 96  |     |      | 75  | 127    |     |
|               | Isopropylbenzene               | 20    | 20.6   | ug/Kg | 103 |     |      | 82  | 124    |     |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 19.3   | ug/Kg | 97  |     |      | 77  | 127    |     |
|               | 1,3-Dichlorobenzene            | 20    | 19.8   | ug/Kg | 99  |     |      | 83  | 122    |     |
|               | 1,4-Dichlorobenzene            | 20    | 19.5   | ug/Kg | 98  |     |      | 84  | 121    |     |
|               | 1,2-Dichlorobenzene            | 20    | 19.4   | ug/Kg | 97  |     |      | 83  | 124    |     |
|               | 1,2-Dibromo-3-Chloropropane    | 20    | 18.7   | ug/Kg | 94  |     |      | 66  | 134    |     |



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary  
SW-846

SDG No.: O1233  
Client: Louis Berger U.S., Inc., A WSP Company  
Analytical Method: SW8260D

Datafile : VY012309.D

| Lab Sample ID | Parameter              | Spike | Result | Unit  | Rec | RPD | Qual | Low | Limits |     |
|---------------|------------------------|-------|--------|-------|-----|-----|------|-----|--------|-----|
|               |                        |       |        |       |     |     |      |     | High   | RPD |
| VY0123SBS01   | 1,2,4-Trichlorobenzene | 20    | 17.9   | ug/Kg | 90  |     |      | 78  | 127    |     |
|               | 1,2,3-Trichlorobenzene | 20    | 17.2   | ug/Kg | 86  |     |      | 70  | 137    |     |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VD0120SBL04

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM Case No.: O1233

SAS No.: O1233 SDG NO.: O1233

Lab File ID: VD075147.D

Lab Sample ID: VD0120SBL04

Date Analyzed: 01/20/2023

Time Analyzed: 12:23

GC Column: RTX-VMS ID: 0.18 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA\_D

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VD0120SBS02       | VD0120SBS02      | VD075150.D     | 01/20/2023       |
| B-P36B            | O1233-05         | VD075158.D     | 01/20/2023       |
| B-38A             | O1233-08         | VD075161.D     | 01/20/2023       |
| B-P38B            | O1233-09         | VD075162.D     | 01/20/2023       |

COMMENTS: \_\_\_\_\_  
\_\_\_\_\_



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VD0123SBL01

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM Case No.: 01233

SAS No.: 01233 SDG NO.: 01233

Lab File ID: VD075169.D

Lab Sample ID: VD0123SBL01

Date Analyzed: 01/23/2023

Time Analyzed: 12:14

GC Column: RTX-VMS ID: 0.18 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA\_D

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VD0123SBS01       | VD0123SBS01      | VD075170.D     | 01/23/2023       |
| VD0123SBSD01      | VD0123SBSD01     | VD075171.D     | 01/23/2023       |
| B-P34A            | 01233-02         | VD075172.D     | 01/23/2023       |
| B-P34B            | 01233-03         | VD075173.D     | 01/23/2023       |
| B-P36A            | 01233-04         | VD075174.D     | 01/23/2023       |
| B-P36BRE          | 01233-05RE       | VD075175.D     | 01/23/2023       |
| B-P37A            | 01233-06         | VD075176.D     | 01/23/2023       |
| B-P38BRE          | 01233-09RE       | VD075178.D     | 01/23/2023       |

COMMENTS:



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0123MBL01

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM Case No.: O1233

SAS No.: O1233 SDG NO.: O1233

Lab File ID: VN076356.D

Lab Sample ID: VN0123MBL01

Date Analyzed: 01/23/2023

Time Analyzed: 10:51

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VN0123MBS01       | VN0123MBS01      | VN076358.D     | 01/23/2023       |
| B-P38C            | O1233-11         | VN076364.D     | 01/23/2023       |
| B-P38CDL          | O1233-11DL       | VN076367.D     | 01/23/2023       |

COMMENTS: \_\_\_\_\_  
\_\_\_\_\_



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0123SBL01

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM Case No.: O1233

SAS No.: O1233 SDG NO.: O1233

Lab File ID: VY012308.D

Lab Sample ID: VY0123SBL01

Date Analyzed: 01/23/2023

Time Analyzed: 13:04

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA\_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VY0123SBS01       | VY0123SBS01      | VY012309.D     | 01/23/2023       |
| B-P37B            | O1233-07         | VY012315.D     | 01/23/2023       |
| B-38ARE           | O1233-08RE       | VY012316.D     | 01/23/2023       |

COMMENTS: \_\_\_\_\_  
\_\_\_\_\_



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0120WBL01

Lab Name: CHEMTECH

Contract: loui01

Lab Code: CHEM Case No.: O1233

SAS No.: O1233 SDG NO.: O1233

Lab File ID: VN076339.D

Lab Sample ID: VN0120WBL01

Date Analyzed: 01/20/2023

Time Analyzed: 11:25

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VN0120WBS01       | VN0120WBS01      | VN076340.D     | 01/20/2023       |
| VN0120WBSD01      | VN0120WBSD01     | VN076341.D     | 01/20/2023       |
| FB01              | O1233-12         | VN076347.D     | 01/20/2023       |

COMMENTS: \_\_\_\_\_  
\_\_\_\_\_



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VD075123.D BFB Injection Date: 01/18/2023  
Instrument ID: MSVOA\_D BFB Injection Time: 09:22  
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 20.3                 |
| 75  | 30.0 - 60.0% of mass 95            | 52.1                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.1                  |
| 173 | Less than 2.0% of mass 174         | 1 ( 1.3 ) 1          |
| 174 | 50.0 - 100.0% of mass 95           | 80.6                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.5 ( 8.1 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 77.5 ( 96.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.4 ( 7 ) 2          |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDIC005         | VSTDIC005        | VD075124.D     | 01/18/2023       | 10:25            |
| VSTDIC010         | VSTDIC010        | VD075125.D     | 01/18/2023       | 11:20            |
| VSTDIC020         | VSTDIC020        | VD075126.D     | 01/18/2023       | 11:48            |
| VSTDIC050         | VSTDIC050        | VD075127.D     | 01/18/2023       | 12:15            |
| VSTDIC100         | VSTDIC100        | VD075128.D     | 01/18/2023       | 12:42            |
| VSTDIC150         | VSTDIC150        | VD075129.D     | 01/18/2023       | 13:10            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VD075145.D BFB Injection Date: 01/20/2023  
Instrument ID: MSVOA\_D BFB Injection Time: 10:40  
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 20.1                 |
| 75  | 30.0 - 60.0% of mass 95            | 52.6                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.8                  |
| 173 | Less than 2.0% of mass 174         | 1.2 ( 1.6 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 74                   |
| 175 | 5.0 - 9.0% of mass 174             | 5.8 ( 7.9 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 72.5 ( 98 ) 1        |
| 177 | 5.0 - 9.0% of mass 176             | 4.6 ( 6.4 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDCCC050        | VSTDCCC050       | VD075146.D     | 01/20/2023       | 11:13            |
| VD0120SBL04       | VD0120SBL04      | VD075147.D     | 01/20/2023       | 12:23            |
| VD0120SBS02       | VD0120SBS02      | VD075150.D     | 01/20/2023       | 13:45            |
| B-P36B            | O1233-05         | VD075158.D     | 01/20/2023       | 17:26            |
| B-38A             | O1233-08         | VD075161.D     | 01/20/2023       | 18:49            |
| B-P38B            | O1233-09         | VD075162.D     | 01/20/2023       | 19:17            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VD075167.D BFB Injection Date: 01/23/2023  
Instrument ID: MSVOA\_D BFB Injection Time: 09:38  
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 20.9                 |
| 75  | 30.0 - 60.0% of mass 95            | 53.1                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.9                  |
| 173 | Less than 2.0% of mass 174         | 1.2 ( 1.5 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 77.3                 |
| 175 | 5.0 - 9.0% of mass 174             | 5.8 ( 7.5 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 74.3 ( 96.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.5 ( 6.1 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDCCC050        | VSTDCCC050       | VD075168.D     | 01/23/2023       | 10:15            |
| VD0123SBL01       | VD0123SBL01      | VD075169.D     | 01/23/2023       | 12:14            |
| VD0123SBS01       | VD0123SBS01      | VD075170.D     | 01/23/2023       | 12:41            |
| VD0123SBSD01      | VD0123SBSD01     | VD075171.D     | 01/23/2023       | 13:08            |
| B-P34A            | O1233-02         | VD075172.D     | 01/23/2023       | 13:36            |
| B-P34B            | O1233-03         | VD075173.D     | 01/23/2023       | 14:04            |
| B-P36A            | O1233-04         | VD075174.D     | 01/23/2023       | 14:32            |
| B-P36BRE          | O1233-05RE       | VD075175.D     | 01/23/2023       | 15:00            |
| B-P37A            | O1233-06         | VD075176.D     | 01/23/2023       | 15:27            |
| B-P38BRE          | O1233-09RE       | VD075178.D     | 01/23/2023       | 16:23            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VN076259.D BFB Injection Date: 01/16/2023  
Instrument ID: MSVOA\_N BFB Injection Time: 09:57  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 18.7                 |
| 75  | 30.0 - 60.0% of mass 95            | 53.4                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7                    |
| 173 | Less than 2.0% of mass 174         | 1.2 ( 1.4 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 88.5                 |
| 175 | 5.0 - 9.0% of mass 174             | 6 ( 6.7 ) 1          |
| 176 | 95.0 - 101.0% of mass 174          | 86.5 ( 97.7 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.1 ( 5.9 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDIC005         | VSTDIC005        | VN076261.D     | 01/16/2023       | 10:54            |
| VSTDIC010         | VSTDIC010        | VN076262.D     | 01/16/2023       | 11:18            |
| VSTDIC020         | VSTDIC020        | VN076263.D     | 01/16/2023       | 11:42            |
| VSTDIC050         | VSTDIC050        | VN076264.D     | 01/16/2023       | 12:06            |
| VSTDIC075         | VSTDIC075        | VN076265.D     | 01/16/2023       | 12:30            |
| VSTDIC001         | VSTDIC001        | VN076267.D     | 01/16/2023       | 13:17            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VN076336.D BFB Injection Date: 01/20/2023  
Instrument ID: MSVOA\_N BFB Injection Time: 09:54  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 18.9                 |
| 75  | 30.0 - 60.0% of mass 95            | 51.9                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7                    |
| 173 | Less than 2.0% of mass 174         | 1.1 ( 1.3 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 87.5                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.6 ( 7.5 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 84.4 ( 96.5 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.8 ( 6.8 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDCCC050        | VSTDCCC050       | VN076337.D     | 01/20/2023       | 10:27            |
| VN0120WBL01       | VN0120WBL01      | VN076339.D     | 01/20/2023       | 11:25            |
| VN0120WBS01       | VN0120WBS01      | VN076340.D     | 01/20/2023       | 11:49            |
| VN0120WBSD01      | VN0120WBSD01     | VN076341.D     | 01/20/2023       | 12:25            |
| FB01              | O1233-12         | VN076347.D     | 01/20/2023       | 14:49            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VN076354.D BFB Injection Date: 01/23/2023  
Instrument ID: MSVOA\_N BFB Injection Time: 09:44  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 18.8                 |
| 75  | 30.0 - 60.0% of mass 95            | 52.2                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.9                  |
| 173 | Less than 2.0% of mass 174         | 1 ( 1.1 ) 1          |
| 174 | 50.0 - 100.0% of mass 95           | 86.3                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.5 ( 7.6 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 84.7 ( 98.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.5 ( 6.5 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDCCC050        | VSTDCCC050       | VN076355.D     | 01/23/2023       | 10:17            |
| VN0123MBL01       | VN0123MBL01      | VN076356.D     | 01/23/2023       | 10:51            |
| VN0123MBS01       | VN0123MBS01      | VN076358.D     | 01/23/2023       | 11:39            |
| B-P38C            | O1233-11         | VN076364.D     | 01/23/2023       | 14:12            |
| B-P38CDL          | O1233-11DL       | VN076367.D     | 01/23/2023       | 15:25            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VY012193.D BFB Injection Date: 01/16/2023  
Instrument ID: MSVOA\_Y BFB Injection Time: 09:00  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 17.8                 |
| 75  | 30.0 - 60.0% of mass 95            | 54.6                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.6                  |
| 173 | Less than 2.0% of mass 174         | 1 ( 1.1 ) 1          |
| 174 | 50.0 - 100.0% of mass 95           | 88.4                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.7 ( 7.6 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 85.2 ( 96.4 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.9 ( 6.9 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDIC005         | VSTDIC005        | VY012194.D     | 01/16/2023       | 12:17            |
| VSTDIC010         | VSTDIC010        | VY012195.D     | 01/16/2023       | 13:03            |
| VSTDIC020         | VSTDIC020        | VY012196.D     | 01/16/2023       | 13:26            |
| VSTDIC050         | VSTDIC050        | VY012197.D     | 01/16/2023       | 13:48            |
| VSTDIC100         | VSTDIC100        | VY012198.D     | 01/16/2023       | 14:11            |
| VSTDIC150         | VSTDIC150        | VY012199.D     | 01/16/2023       | 14:34            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233  
Lab File ID: VY012306.D BFB Injection Date: 01/23/2023  
Instrument ID: MSVOA\_Y BFB Injection Time: 10:03  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 17.7                 |
| 75  | 30.0 - 60.0% of mass 95            | 54.6                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.6                  |
| 173 | Less than 2.0% of mass 174         | 1 ( 1.1 ) 1          |
| 174 | 50.0 - 100.0% of mass 95           | 90.6                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.3 ( 7 ) 1          |
| 176 | 95.0 - 101.0% of mass 174          | 87.1 ( 96.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.8 ( 6.6 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDCCC050        | VSTDCCC050       | VY012307.D     | 01/23/2023       | 12:09            |
| VY0123SBL01       | VY0123SBL01      | VY012308.D     | 01/23/2023       | 13:04            |
| VY0123SBS01       | VY0123SBS01      | VY012309.D     | 01/23/2023       | 13:27            |
| B-P37B            | O1233-07         | VY012315.D     | 01/23/2023       | 15:48            |
| B-38ARE           | O1233-08RE       | VY012316.D     | 01/23/2023       | 16:11            |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: louie01  
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
Lab File ID: VD075146.D Date Analyzed: 01/20/2023  
Instrument ID: MSVOA\_D Time Analyzed: 11:13  
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 87549         | 7.88  | 150535        | 8.78  | 130836        | 11.59  |
| UPPER LIMIT    | 175098        | 8.375 | 301070        | 9.281 | 261672        | 12.087 |
| LOWER LIMIT    | 43774.5       | 7.375 | 75267.5       | 8.281 | 65418         | 11.087 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| B-P36B         | 4925 *        | 7.88  | 11060 *       | 8.78  | 10688 *       | 11.59  |
| B-38A          | 3464 *        | 7.88  | 6979 *        | 8.78  | 6961 *        | 11.58  |
| B-P38B         | 41902 *       | 7.88  | 88407         | 8.78  | 88309         | 11.59  |
| VD0120SBL04    | 86545         | 7.88  | 180550        | 8.78  | 168873        | 11.58  |
| VD0120SBS02    | 78332         | 7.88  | 142964        | 8.78  | 128592        | 11.59  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
Lab File ID: VD075146.D Date Analyzed: 01/20/2023  
Instrument ID: MSVOA\_D Time Analyzed: 11:13  
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 59366         | 13.522 |  |  |  |  |
| UPPER LIMIT    | 118732        | 14.022 |  |  |  |  |
| LOWER LIMIT    | 29683         | 13.022 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| B-P36B         | 5310 *        | 13.52  |  |  |  |  |
| B-38A          | 2980 *        | 13.52  |  |  |  |  |
| B-P38B         | 40736         | 13.52  |  |  |  |  |
| VD0120SBL04    | 72499         | 13.52  |  |  |  |  |
| VD0120SBS02    | 58018         | 13.52  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.

## VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
 Lab File ID: VD075168.D Date Analyzed: 01/23/2023  
 Instrument ID: MSVOA\_D Time Analyzed: 10:15  
 GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 83058         | 7.88  | 142829        | 8.78  | 124569        | 11.59  |
| UPPER LIMIT    | 166116        | 8.375 | 285658        | 9.281 | 249138        | 12.087 |
| LOWER LIMIT    | 41529         | 7.375 | 71414.5       | 8.281 | 62284.5       | 11.087 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| B-P34A         | 67921         | 7.88  | 129748        | 8.78  | 124838        | 11.59  |
| B-P34B         | 49085         | 7.88  | 95032         | 8.78  | 93578         | 11.59  |
| B-P36A         | 87972         | 7.88  | 171041        | 8.78  | 157598        | 11.59  |
| B-P36BRE       | 2289 *        | 7.88  | 4728 *        | 8.78  | 4222 *        | 11.58  |
| B-P37A         | 85119         | 7.88  | 168462        | 8.78  | 154125        | 11.59  |
| B-P38BRE       | 1300 *        | 7.88  | 2143 *        | 8.78  | 1966 *        | 11.59  |
| VD0123SBL01    | 95373         | 7.88  | 184136        | 8.78  | 168096        | 11.59  |
| VD0123SBS01    | 75965         | 7.88  | 134615        | 8.78  | 120882        | 11.59  |
| VD0123SBS01    | 72362         | 7.88  | 130215        | 8.78  | 115554        | 11.59  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
Lab File ID: VD075168.D Date Analyzed: 01/23/2023  
Instrument ID: MSVOA\_D Time Analyzed: 10:15  
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 56972         | 13.522 |  |  |  |  |
| UPPER LIMIT    | 113944        | 14.022 |  |  |  |  |
| LOWER LIMIT    | 28486         | 13.022 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| B-P34A         | 54879         | 13.52  |  |  |  |  |
| B-P34B         | 41757         | 13.52  |  |  |  |  |
| B-P36A         | 64243         | 13.52  |  |  |  |  |
| B-P36BRE       | 1552 *        | 13.52  |  |  |  |  |
| B-P37A         | 63432         | 13.52  |  |  |  |  |
| B-P38BRE       | 699 *         | 13.53  |  |  |  |  |
| VD0123SBL01    | 71751         | 13.52  |  |  |  |  |
| VD0123SBS01    | 55819         | 13.52  |  |  |  |  |
| VD0123SBSD01   | 53241         | 13.52  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.

## VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
 Lab File ID: VN076337.D Date Analyzed: 01/20/2023  
 Instrument ID: MSVOA\_N Time Analyzed: 10:27  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT # | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 383049        | 8.23 | 610968        | 9.11  | 549444        | 11.87  |
| UPPER LIMIT    | 766098        | 8.73 | 1221940       | 9.606 | 1098890       | 12.365 |
| LOWER LIMIT    | 191525        | 7.73 | 305484        | 8.606 | 274722        | 11.365 |
| EPA SAMPLE NO. |               |      |               |       |               |        |
| FB01           | 370105        | 8.23 | 595933        | 9.11  | 523917        | 11.87  |
| VN0120WBL01    | 490775        | 8.23 | 773684        | 9.11  | 657515        | 11.87  |
| VN0120WBS01    | 434458        | 8.23 | 686884        | 9.11  | 590812        | 11.87  |
| VN0120WBSD01   | 359856        | 8.23 | 567895        | 9.11  | 510094        | 11.87  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
Lab File ID: VN076337.D Date Analyzed: 01/20/2023  
Instrument ID: MSVOA\_N Time Analyzed: 10:27  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 272155        | 13.794 |  |  |  |  |
| UPPER LIMIT    | 544310        | 14.294 |  |  |  |  |
| LOWER LIMIT    | 136078        | 13.294 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| FB01           | 212240        | 13.79  |  |  |  |  |
| VN0120WBL01    | 251353        | 13.79  |  |  |  |  |
| VN0120WBS01    | 268276        | 13.79  |  |  |  |  |
| VN0120WBSD01   | 238179        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.

## VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
 Lab File ID: VN076355.D Date Analyzed: 01/23/2023  
 Instrument ID: MSVOA\_N Time Analyzed: 10:17  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT # | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 360494        | 8.23 | 576114        | 9.11  | 518968        | 11.87 |
| UPPER LIMIT    | 720988        | 8.73 | 1152230       | 9.606 | 1037940       | 12.37 |
| LOWER LIMIT    | 180247        | 7.73 | 288057        | 8.606 | 259484        | 11.37 |
| EPA SAMPLE NO. |               |      |               |       |               |       |
| B-P38C         | 380260        | 8.22 | 594043        | 9.10  | 544762        | 11.87 |
| B-P38CDL       | 384506        | 8.23 | 612820        | 9.11  | 549267        | 11.87 |
| VN0123MBL01    | 303559        | 8.23 | 491398        | 9.11  | 403787        | 11.87 |
| VN0123MBS01    | 381993        | 8.23 | 609908        | 9.11  | 527333        | 11.87 |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: louie01  
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
Lab File ID: VN076355.D Date Analyzed: 01/23/2023  
Instrument ID: MSVOA\_N Time Analyzed: 10:17  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 248688        | 13.794 |  |  |  |  |
| UPPER LIMIT    | 497376        | 14.294 |  |  |  |  |
| LOWER LIMIT    | 124344        | 13.294 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| B-P38C         | 318779        | 13.79  |  |  |  |  |
| B-P38CDL       | 255428        | 13.79  |  |  |  |  |
| VN0123MBL01    | 158993        | 13.79  |  |  |  |  |
| VN0123MBS01    | 238215        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.

## VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
 Lab File ID: VY012307.D Date Analyzed: 01/23/2023  
 Instrument ID: MSVOA\_Y Time Analyzed: 12:09  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|-------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 179464        | 7.79  | 268117        | 8.69  | 243165        | 11.49 |
| UPPER LIMIT    | 358928        | 8.289 | 536234        | 9.191 | 486330        | 11.99 |
| LOWER LIMIT    | 89732         | 7.289 | 134059        | 8.191 | 121583        | 10.99 |
| EPA SAMPLE NO. |               |       |               |       |               |       |
| B-P37B         | 35937 *       | 7.79  | 57297 *       | 8.69  | 55952 *       | 11.49 |
| B-38ARE        | 4898 *        | 7.80  | 8005 *        | 8.69  | 7433 *        | 11.50 |
| VY0123SBL01    | 162197        | 7.79  | 255207        | 8.69  | 227538        | 11.49 |
| VY0123SBS01    | 173816        | 7.79  | 263834        | 8.69  | 239765        | 11.49 |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.

## VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233  
 Lab File ID: VY012307.D Date Analyzed: 01/23/2023  
 Instrument ID: MSVOA\_Y Time Analyzed: 12:09  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 123436        | 13.428 |  |  |  |  |
| UPPER LIMIT    | 246872        | 13.928 |  |  |  |  |
| LOWER LIMIT    | 61718         | 12.928 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| B-P37B         | 24351 *       | 13.43  |  |  |  |  |
| B-38ARE        | 2570 *        | 13.43  |  |  |  |  |
| VY0123SBL01    | 110055        | 13.43  |  |  |  |  |
| VY0123SBS01    | 122385        | 13.43  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.

# QC SAMPLE DATA

A

B

C

D

E

F

G

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0120SBL04                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0120SBL04                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075147.D        | 1         |           | 01/20/23 12:23 | VD012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.86  | U         | 0.86 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.91  | U         | 0.91 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.20  | U         | 1.20 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.89  | U         | 0.89 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 0.97  | U         | 0.97 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.72  | U         | 0.72 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.86  | U         | 0.86 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 12.2  | U         | 12.2 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.75  | U         | 0.75 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.93  | U         | 0.93 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.30  | U         | 1.30 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.00  | U         | 6.00 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.68  | U         | 0.68 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.70  | U         | 0.70 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.84  | U         | 0.84 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 7.30  | U         | 7.30 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.79  | U         | 0.79 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.68  | U         | 0.68 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.81  | U         | 0.81 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.67  | U         | 0.67 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.75  | U         | 0.75 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.80  | U         | 0.80 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.66  | U         | 0.66 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.84  | U         | 0.84 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.73  | U         | 0.73 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.65  | U         | 0.65 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.70  | U         | 0.70 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 4.60  | U         | 4.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.63  | U         | 0.63 | 5.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.74  | U         | 0.74 | 5.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.71  | U         | 0.71 | 5.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0120SBL04                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0120SBL04                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075147.D        | 1         |           | 01/20/23 12:23 | VD012023      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.86   | U         | 0.86     | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 4.70   | U         | 4.70     | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.75   | U         | 0.75     | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.75   | U         | 0.75     | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.76   | U         | 0.76     | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.65   | U         | 0.65     | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.70   | U         | 0.70     | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.50   | U         | 1.50     | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.79   | U         | 0.79     | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.79   | U         | 0.79     | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.81   | U         | 0.81     | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.72   | U         | 0.72     | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.10   | U         | 1.10     | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.67   | U         | 0.67     | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.63   | U         | 0.63     | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.64   | U         | 0.64     | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.20   | U         | 1.20     | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.94   | U         | 0.94     | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.00   | U         | 1.00     | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 57.2   |           | 50 - 163 | 114%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 50.7   |           | 54 - 147 | 101%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 52.1   |           | 58 - 134 | 104%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 53.6   |           | 30 - 143 | 107%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 86500  | 7.875     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 181000 | 8.781     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 169000 | 11.581    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 72500  | 13.522    |          |            |                   |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0120SBL04                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0120SBL04                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075147.D        | 1         |           | 01/20/23 12:23 | VD012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0123SBL01                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBL01                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075169.D        | 1         |           | 01/23/23 12:14 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.86  | U         | 0.86 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.91  | U         | 0.91 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.20  | U         | 1.20 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.89  | U         | 0.89 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 0.97  | U         | 0.97 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.72  | U         | 0.72 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.86  | U         | 0.86 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 12.2  | U         | 12.2 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.75  | U         | 0.75 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.93  | U         | 0.93 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.30  | U         | 1.30 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.00  | U         | 6.00 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.68  | U         | 0.68 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.70  | U         | 0.70 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.84  | U         | 0.84 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 7.30  | U         | 7.30 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.79  | U         | 0.79 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.68  | U         | 0.68 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.81  | U         | 0.81 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.67  | U         | 0.67 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.75  | U         | 0.75 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.80  | U         | 0.80 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.66  | U         | 0.66 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.84  | U         | 0.84 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.73  | U         | 0.73 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.65  | U         | 0.65 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.70  | U         | 0.70 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 4.60  | U         | 4.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.63  | U         | 0.63 | 5.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.74  | U         | 0.74 | 5.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.71  | U         | 0.71 | 5.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0123SBL01                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBL01                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075169.D        | 1         |           | 01/23/23 12:14 | VD012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.86   | U         | 0.86     | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 4.70   | U         | 4.70     | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.75   | U         | 0.75     | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.75   | U         | 0.75     | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.76   | U         | 0.76     | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.65   | U         | 0.65     | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.70   | U         | 0.70     | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.50   | U         | 1.50     | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.79   | U         | 0.79     | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.79   | U         | 0.79     | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.81   | U         | 0.81     | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.72   | U         | 0.72     | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.10   | U         | 1.10     | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.67   | U         | 0.67     | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.63   | U         | 0.63     | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.64   | U         | 0.64     | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.20   | U         | 1.20     | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.94   | U         | 0.94     | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.00   | U         | 1.00     | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 51.9   |           | 50 - 163 | 104%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 48.5   |           | 54 - 147 | 97%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 50.8   |           | 58 - 134 | 102%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 50.4   |           | 30 - 143 | 101%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 95400  | 7.875     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 184000 | 8.781     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 168000 | 11.587    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 71800  | 13.522    |          |            |                   |

**Report of Analysis**

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0123SBL01                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBL01                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075169.D        | 1         |           | 01/23/23 12:14 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0120WBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBL01                               |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076339.D        | 1         |           | 01/20/23 11:25 | VN012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|------|------------|-------|
| <b>TARGETS</b> |                                |       |           |      |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 0.22  | U         | 0.22 | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 1.60  | U         | 1.60 | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 0.26  | U         | 0.26 | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 0.23  | U         | 0.23 | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 1.20  | U         | 1.20 | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 0.26  | U         | 0.26 | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 0.53  | U         | 0.53 | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 0.20  | U         | 0.20 | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 1.20  | U         | 1.20 | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 0.82  | U         | 0.82 | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 0.19  | U         | 0.19 | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 0.13  | U         | 0.13 | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 0.16  | U         | 0.16 | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 0.27  | U         | 0.27 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 0.18  | U         | 0.18 | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 0.87  | U         | 0.87 | 5.00       | ug/L  |
| 108-88-3       | Toluene                        | 0.17  | U         | 0.17 | 1.00       | ug/L  |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.14  | U         | 0.14 | 1.00       | ug/L  |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.16  | U         | 0.16 | 1.00       | ug/L  |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VN0120WBL01                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBL01                               | Matrix:         | Water         |
| Analytical Method: | SW8260                                    | % Solid:        | 0             |
| Sample Wt/Vol:     | 5 Units: mL                               | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076339.D        | 1         |           | 01/20/23 11:25 | VN012023      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 0.76   | U         | 0.76     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 0.14   | U         | 0.14     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 0.33   | U         | 0.33     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 0.13   | U         | 0.13     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 0.23   | U         | 0.23     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 0.42   | U         | 0.42     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.23   | U         | 0.23     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 0.33   | U         | 0.33     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 47.5   |           | 74 - 125 | 95%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 48.7   |           | 75 - 124 | 97%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 49.7   |           | 86 - 113 | 99%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 45.5   |           | 64 - 133 | 91%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 491000 | 8.23      |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 774000 | 9.106     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 658000 | 11.871    |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 251000 | 13.794    |          |            |         |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0120WBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBL01                               |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076339.D        | 1         |           | 01/20/23 11:25 | VN012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0123MBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0123MBL01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100                                       | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076356.D        | 1         |           | 01/23/23 10:51 | VN012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 86.0  | U         | 86.0 | 500        | ug/Kg             |
| 74-87-3        | Chloromethane                  | 110   | U         | 110  | 500        | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 91.0  | U         | 91.0 | 500        | ug/Kg             |
| 74-83-9        | Bromomethane                   | 120   | U         | 120  | 500        | ug/Kg             |
| 75-00-3        | Chloroethane                   | 89.0  | U         | 89.0 | 500        | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 97.0  | U         | 97.0 | 500        | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 72.0  | U         | 72.0 | 500        | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 86.0  | U         | 86.0 | 500        | ug/Kg             |
| 67-64-1        | Acetone                        | 1200  | U         | 1200 | 2500       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 75.0  | U         | 75.0 | 500        | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 93.0  | U         | 93.0 | 500        | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 130   | U         | 130  | 500        | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 600   | U         | 600  | 1000       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 68.0  | U         | 68.0 | 500        | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 70.0  | U         | 70.0 | 500        | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 84.0  | U         | 84.0 | 500        | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 730   | U         | 730  | 2500       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 79.0  | U         | 79.0 | 500        | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 68.0  | U         | 68.0 | 500        | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 81.0  | U         | 81.0 | 500        | ug/Kg             |
| 67-66-3        | Chloroform                     | 67.0  | U         | 67.0 | 500        | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 75.0  | U         | 75.0 | 500        | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 80.0  | U         | 80.0 | 500        | ug/Kg             |
| 71-43-2        | Benzene                        | 66.0  | U         | 66.0 | 500        | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 84.0  | U         | 84.0 | 500        | ug/Kg             |
| 79-01-6        | Trichloroethene                | 73.0  | U         | 73.0 | 500        | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 65.0  | U         | 65.0 | 500        | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 70.0  | U         | 70.0 | 500        | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 460   | U         | 460  | 2500       | ug/Kg             |
| 108-88-3       | Toluene                        | 63.0  | U         | 63.0 | 500        | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 74.0  | U         | 74.0 | 500        | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 71.0  | U         | 71.0 | 500        | ug/Kg             |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0123MBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0123MBL01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100                                       | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076356.D        | 1         |           | 01/23/23 10:51 | VN012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 86.0   | U         | 86.0     | 500        | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 470    | U         | 470      | 2500       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 75.0   | U         | 75.0     | 500        | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 75.0   | U         | 75.0     | 500        | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 76.0   | U         | 76.0     | 500        | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 65.0   | U         | 65.0     | 500        | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 70.0   | U         | 70.0     | 500        | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 150    | U         | 150      | 1000       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 79.0   | U         | 79.0     | 500        | ug/Kg             |
| 100-42-5                  | Styrene                     | 79.0   | U         | 79.0     | 500        | ug/Kg             |
| 75-25-2                   | Bromoform                   | 81.0   | U         | 81.0     | 500        | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 72.0   | U         | 72.0     | 500        | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 110    | U         | 110      | 500        | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 67.0   | U         | 67.0     | 500        | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 63.0   | U         | 63.0     | 500        | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 64.0   | U         | 64.0     | 500        | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 120    | U         | 120      | 500        | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 94.0   | U         | 94.0     | 500        | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 100    | U         | 100      | 500        | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.9   |           | 50 - 163 | 100%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 49.5   |           | 54 - 147 | 99%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 48.9   |           | 58 - 134 | 98%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 44.4   |           | 30 - 143 | 89%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 304000 | 8.23      |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 491000 | 9.106     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 404000 | 11.87     |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 159000 | 13.794    |          |            |                   |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0123MBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0123MBL01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100                                       | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076356.D        | 1         |           | 01/23/23 10:51 | VN012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VY0123SBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VY0123SBL01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012308.D        | 1         |           | 01/23/23 13:04 | VY012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 0.86  | U         | 0.86 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1.10  | U         | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 0.91  | U         | 0.91 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1.20  | U         | 1.20 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 0.89  | U         | 0.89 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 0.97  | U         | 0.97 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.72  | U         | 0.72 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 0.86  | U         | 0.86 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 12.2  | U         | 12.2 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 0.75  | U         | 0.75 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.93  | U         | 0.93 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1.30  | U         | 1.30 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 6.00  | U         | 6.00 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.68  | U         | 0.68 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 0.70  | U         | 0.70 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 0.84  | U         | 0.84 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 7.30  | U         | 7.30 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 0.79  | U         | 0.79 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.68  | U         | 0.68 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 0.81  | U         | 0.81 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 0.67  | U         | 0.67 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.75  | U         | 0.75 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 0.80  | U         | 0.80 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 0.66  | U         | 0.66 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 0.84  | U         | 0.84 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 0.73  | U         | 0.73 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 0.65  | U         | 0.65 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 0.70  | U         | 0.70 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 4.60  | U         | 4.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 0.63  | U         | 0.63 | 5.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 0.74  | U         | 0.74 | 5.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 0.71  | U         | 0.71 | 5.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VY0123SBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VY0123SBL01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012308.D        | 1         |           | 01/23/23 13:04 | VY012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.86   | U         | 0.86     | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 4.70   | U         | 4.70     | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 0.75   | U         | 0.75     | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 0.75   | U         | 0.75     | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 0.76   | U         | 0.76     | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 0.65   | U         | 0.65     | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 0.70   | U         | 0.70     | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 1.50   | U         | 1.50     | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 0.79   | U         | 0.79     | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 0.79   | U         | 0.79     | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 0.81   | U         | 0.81     | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 0.72   | U         | 0.72     | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1.10   | U         | 1.10     | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.67   | U         | 0.67     | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.63   | U         | 0.63     | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.64   | U         | 0.64     | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1.20   | U         | 1.20     | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.94   | U         | 0.94     | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1.00   | U         | 1.00     | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 52.7   |           | 50 - 163 | 105%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 45.4   |           | 54 - 147 | 91%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 47.0   |           | 58 - 134 | 94%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 41.2   |           | 30 - 143 | 82%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 162000 | 7.789     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 255000 | 8.691     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 228000 | 11.489    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 110000 | 13.428    |          |            |                   |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VY0123SBL01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VY0123SBL01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012308.D        | 1         |           | 01/23/23 13:04 | VY012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0120SBS02                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0120SBS02                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075150.D        | 1         |           | 01/20/23 13:45 | VD012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 25.8  |           | 0.86 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 22.6  |           | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 20.1  |           | 0.91 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 21.1  |           | 1.20 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 20.7  |           | 0.89 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 21.2  |           | 0.97 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 21.7  |           | 0.72 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 22.5  |           | 0.86 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 99.8  |           | 12.2 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 21.8  |           | 0.75 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 21.8  |           | 0.93 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 22.8  |           | 1.30 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 20.1  |           | 6.00 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 21.1  |           | 0.68 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 21.6  |           | 0.70 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 21.7  |           | 0.84 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 110   |           | 7.30 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 20.0  |           | 0.79 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 21.6  |           | 0.68 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 24.1  |           | 0.81 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 21.5  |           | 0.67 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 20.8  |           | 0.75 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 21.0  |           | 0.80 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 21.4  |           | 0.66 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 20.9  |           | 0.84 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 20.6  |           | 0.73 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 20.7  |           | 0.65 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 20.4  |           | 0.70 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 110   |           | 4.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 20.7  |           | 0.63 | 5.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 20.0  |           | 0.74 | 5.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 20.7  |           | 0.71 | 5.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0120SBS02                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0120SBS02                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075150.D        | 1         |           | 01/20/23 13:45 | VD012023      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 20.9   |           | 0.86     | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 110    |           | 4.70     | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 20.0   |           | 0.75     | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 21.0   |           | 0.75     | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 19.2   |           | 0.76     | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 20.4   |           | 0.65     | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 20.4   |           | 0.70     | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 41.6   |           | 1.50     | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 20.9   |           | 0.79     | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 20.7   |           | 0.79     | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 19.9   |           | 0.81     | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 20.1   |           | 0.72     | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 21.7   |           | 1.10     | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 19.8   |           | 0.67     | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 20.8   |           | 0.63     | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 20.4   |           | 0.64     | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 21.7   |           | 1.20     | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 18.2   |           | 0.94     | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 18.4   |           | 1.00     | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 56.1   |           | 50 - 163 | 112%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 53.0   |           | 54 - 147 | 106%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 53.3   |           | 58 - 134 | 107%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 53.5   |           | 30 - 143 | 107%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 78300  | 7.875     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 143000 | 8.781     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 129000 | 11.586    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 58000  | 13.522    |          |            |                   |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0120SBS02                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0120SBS02                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075150.D        | 1         |           | 01/20/23 13:45 | VD012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0123SBS01                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBS01                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075170.D        | 1         |           | 01/23/23 12:41 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 24.9  |           | 0.86 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 22.7  |           | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 20.4  |           | 0.91 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 19.9  |           | 1.20 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 19.7  |           | 0.89 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 21.3  |           | 0.97 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 22.1  |           | 0.72 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 21.7  |           | 0.86 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 100   |           | 12.2 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 21.8  |           | 0.75 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 21.1  |           | 0.93 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 22.0  |           | 1.30 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 16.3  |           | 6.00 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 21.2  |           | 0.68 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 21.5  |           | 0.70 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 22.0  |           | 0.84 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 100   |           | 7.30 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 20.3  |           | 0.79 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 20.8  |           | 0.68 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 23.2  |           | 0.81 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 20.9  |           | 0.67 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 21.0  |           | 0.75 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 20.8  |           | 0.80 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 21.1  |           | 0.66 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 21.3  |           | 0.84 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 19.8  |           | 0.73 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 19.9  |           | 0.65 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 20.8  |           | 0.70 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 110   |           | 4.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 20.8  |           | 0.63 | 5.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 19.8  |           | 0.74 | 5.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 20.3  |           | 0.71 | 5.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0123SBS01                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBS01                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075170.D        | 1         |           | 01/23/23 12:41 | VD012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 21.7   |           | 0.86     | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 110    |           | 4.70     | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 21.3   |           | 0.75     | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 20.6   |           | 0.75     | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 19.7   |           | 0.76     | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 20.6   |           | 0.65     | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 20.9   |           | 0.70     | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 41.1   |           | 1.50     | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 20.8   |           | 0.79     | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 20.7   |           | 0.79     | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 20.2   |           | 0.81     | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 19.9   |           | 0.72     | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 20.7   |           | 1.10     | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 19.9   |           | 0.67     | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 20.1   |           | 0.63     | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 19.7   |           | 0.64     | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 20.7   |           | 1.20     | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 18.2   |           | 0.94     | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 18.1   |           | 1.00     | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 54.9   |           | 50 - 163 | 110%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 53.1   |           | 54 - 147 | 106%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 53.7   |           | 58 - 134 | 107%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 51.6   |           | 30 - 143 | 103%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 76000  | 7.881     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 135000 | 8.781     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 121000 | 11.586    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 55800  | 13.522    |          |            |                   |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VD0123SBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBS01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS                                   | ID : 0.18 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075170.D        | 1         |           | 01/23/23 12:41 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0120WBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBS01                               |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076340.D        | 1         |           | 01/20/23 11:49 | VN012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|------|------------|-------|
| <b>TARGETS</b> |                                |       |           |      |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 17.4  |           | 0.17 | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 16.6  |           | 0.20 | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 17.5  |           | 0.22 | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 17.6  |           | 1.60 | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 18.8  |           | 0.26 | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 16.9  |           | 0.20 | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 17.2  |           | 0.17 | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 16.8  |           | 0.23 | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 85.8  |           | 1.20 | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 15.2  |           | 0.26 | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 17.6  |           | 0.18 | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 16.4  |           | 0.53 | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 18.5  |           | 0.18 | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 16.2  |           | 0.20 | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 16.8  |           | 0.20 | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 16.2  |           | 1.20 | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 83.8  |           | 0.82 | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 17.2  |           | 0.18 | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 17.3  |           | 0.17 | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 18.8  |           | 0.19 | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 17.0  |           | 0.18 | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 17.5  |           | 0.18 | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 17.6  |           | 0.13 | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 17.2  |           | 0.16 | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 17.3  |           | 0.18 | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 16.7  |           | 0.27 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 17.0  |           | 0.17 | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 17.6  |           | 0.18 | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 88.6  |           | 0.87 | 5.00       | ug/L  |
| 108-88-3       | Toluene                        | 16.7  |           | 0.17 | 1.00       | ug/L  |
| 10061-02-6     | t-1,3-Dichloropropene          | 17.9  |           | 0.14 | 1.00       | ug/L  |
| 10061-01-5     | cis-1,3-Dichloropropene        | 17.3  |           | 0.16 | 1.00       | ug/L  |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0120WBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBS01                               |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076340.D        | 1         |           | 01/20/23 11:49 | VN012023      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 17.5   |           | 0.19     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 88.3   |           | 0.76     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 18.0   |           | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 17.8   |           | 0.14     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 17.6   |           | 0.18     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 17.8   |           | 0.17     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 18.3   |           | 0.17     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 36.8   |           | 0.33     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 18.5   |           | 0.18     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 18.3   |           | 0.13     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 18.1   |           | 0.16     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 19.4   |           | 0.19     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 18.1   |           | 0.23     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 17.6   |           | 0.20     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 18.0   |           | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 17.4   |           | 0.17     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 17.9   |           | 0.42     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 16.7   |           | 0.23     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 17.1   |           | 0.33     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.0   |           | 74 - 125 | 98%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 50.3   |           | 75 - 124 | 101%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 50.7   |           | 86 - 113 | 101%       | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 50.9   |           | 64 - 133 | 102%       | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 434000 | 8.23      |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 687000 | 9.106     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 591000 | 11.865    |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 268000 | 13.794    |          |            |         |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0120WBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBS01                               |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076340.D        | 1         |           | 01/20/23 11:49 | VN012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0123MBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0123MBS01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100                                       | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076358.D        | 1         |           | 01/23/23 11:39 | VN012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 1800  |           | 86.0 | 500        | ug/Kg             |
| 74-87-3        | Chloromethane                  | 1800  |           | 110  | 500        | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 2000  |           | 91.0 | 500        | ug/Kg             |
| 74-83-9        | Bromomethane                   | 1900  |           | 120  | 500        | ug/Kg             |
| 75-00-3        | Chloroethane                   | 2000  |           | 89.0 | 500        | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 1700  |           | 97.0 | 500        | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 1800  |           | 72.0 | 500        | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 1700  |           | 86.0 | 500        | ug/Kg             |
| 67-64-1        | Acetone                        | 10700 |           | 1200 | 2500       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 1700  |           | 75.0 | 500        | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 1800  |           | 93.0 | 500        | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 1800  |           | 130  | 500        | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 1700  |           | 600  | 1000       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 1700  |           | 68.0 | 500        | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 1800  |           | 70.0 | 500        | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 1700  |           | 84.0 | 500        | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 9300  |           | 730  | 2500       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 1900  |           | 79.0 | 500        | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 1700  |           | 68.0 | 500        | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 2100  |           | 81.0 | 500        | ug/Kg             |
| 67-66-3        | Chloroform                     | 1800  |           | 67.0 | 500        | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 1800  |           | 75.0 | 500        | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 1700  |           | 80.0 | 500        | ug/Kg             |
| 71-43-2        | Benzene                        | 1800  |           | 66.0 | 500        | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 1800  |           | 84.0 | 500        | ug/Kg             |
| 79-01-6        | Trichloroethene                | 1700  |           | 73.0 | 500        | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 1800  |           | 65.0 | 500        | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 1900  |           | 70.0 | 500        | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 9500  |           | 460  | 2500       | ug/Kg             |
| 108-88-3       | Toluene                        | 1700  |           | 63.0 | 500        | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 1900  |           | 74.0 | 500        | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 1900  |           | 71.0 | 500        | ug/Kg             |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0123MBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0123MBS01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100                                       | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076358.D        | 1         |           | 01/23/23 11:39 | VN012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 1800   |           | 86.0     | 500        | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 9400   |           | 470      | 2500       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 2000   |           | 75.0     | 500        | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 1800   |           | 75.0     | 500        | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 1700   |           | 76.0     | 500        | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 1800   |           | 65.0     | 500        | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 1900   |           | 70.0     | 500        | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 3700   |           | 150      | 1000       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 1800   |           | 79.0     | 500        | ug/Kg             |
| 100-42-5                  | Styrene                     | 1900   |           | 79.0     | 500        | ug/Kg             |
| 75-25-2                   | Bromoform                   | 2100   |           | 81.0     | 500        | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 2000   |           | 72.0     | 500        | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 1900   |           | 110      | 500        | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 1800   |           | 67.0     | 500        | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 1800   |           | 63.0     | 500        | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 1700   |           | 64.0     | 500        | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 1800   |           | 120      | 500        | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 1700   |           | 94.0     | 500        | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 1700   |           | 100      | 500        | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 52.1   |           | 50 - 163 | 104%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 51.4   |           | 54 - 147 | 103%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 50.9   |           | 58 - 134 | 102%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 51.3   |           | 30 - 143 | 103%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 382000 | 8.23      |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 610000 | 9.106     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 527000 | 11.865    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 238000 | 13.794    |          |            |                   |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0123MBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0123MBS01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 10000 uL      |
| Soil Aliquot Vol:  | 100                                       | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | MED           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076358.D        | 1         |           | 01/23/23 11:39 | VN012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VY0123SBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VY0123SBS01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012309.D        | 1         |           | 01/23/23 13:27 | VY012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 21.2  |           | 0.86 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 19.7  |           | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 20.0  |           | 0.91 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 21.0  |           | 1.20 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 20.8  |           | 0.89 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 21.2  |           | 0.97 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 20.8  |           | 0.72 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 20.6  |           | 0.86 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 87.9  |           | 12.2 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 20.2  |           | 0.75 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 19.4  |           | 0.93 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 18.3  |           | 1.30 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 15.2  |           | 6.00 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 20.1  |           | 0.68 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 20.5  |           | 0.70 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 20.2  |           | 0.84 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 92.3  |           | 7.30 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 20.7  |           | 0.79 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 20.5  |           | 0.68 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 19.2  |           | 0.81 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 20.4  |           | 0.67 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 20.7  |           | 0.75 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 20.8  |           | 0.80 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 20.5  |           | 0.66 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 20.2  |           | 0.84 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 20.1  |           | 0.73 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 20.3  |           | 0.65 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 20.3  |           | 0.70 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 93.5  |           | 4.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 20.4  |           | 0.63 | 5.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 19.6  |           | 0.74 | 5.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 20.1  |           | 0.71 | 5.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VY0123SBS01                               | SDG No.:        | O1233         |
| Lab Sample ID:     | VY0123SBS01                               | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624 ID : 0.25                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012309.D        | 1         |           | 01/23/23 13:27 | VY012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 19.6   |           | 0.86     | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 92.0   |           | 4.70     | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 19.7   |           | 0.75     | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 19.3   |           | 0.75     | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 21.0   |           | 0.76     | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 20.2   |           | 0.65     | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 20.3   |           | 0.70     | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 41.2   |           | 1.50     | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 20.0   |           | 0.79     | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 20.4   |           | 0.79     | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 19.1   |           | 0.81     | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 20.6   |           | 0.72     | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 19.3   |           | 1.10     | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 19.8   |           | 0.67     | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 19.5   |           | 0.63     | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 19.4   |           | 0.64     | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 18.7   |           | 1.20     | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 17.9   |           | 0.94     | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 17.2   |           | 1.00     | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.4   |           | 50 - 163 | 99%        | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 46.3   |           | 54 - 147 | 93%        | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 48.5   |           | 58 - 134 | 97%        | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 44.5   |           | 30 - 143 | 89%        | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 174000 | 7.789     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 264000 | 8.691     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 240000 | 11.489    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 122000 | 13.428    |          |            |                   |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VY0123SBS01                               |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VY0123SBS01                               |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VY012309.D        | 1         |           | 01/23/23 13:27 | VY012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VD0123SBSD01                              |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBSD01                              |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS                                   | ID : 0.18 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075171.D        | 1         |           | 01/23/23 13:08 | VD012323      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units(Dry Weight) |
|----------------|--------------------------------|-------|-----------|------|------------|-------------------|
| <b>TARGETS</b> |                                |       |           |      |            |                   |
| 75-71-8        | Dichlorodifluoromethane        | 26.4  |           | 0.86 | 5.00       | ug/Kg             |
| 74-87-3        | Chloromethane                  | 22.1  |           | 1.10 | 5.00       | ug/Kg             |
| 75-01-4        | Vinyl Chloride                 | 20.9  |           | 0.91 | 5.00       | ug/Kg             |
| 74-83-9        | Bromomethane                   | 20.4  |           | 1.20 | 5.00       | ug/Kg             |
| 75-00-3        | Chloroethane                   | 21.5  |           | 0.89 | 5.00       | ug/Kg             |
| 75-69-4        | Trichlorofluoromethane         | 22.5  |           | 0.97 | 5.00       | ug/Kg             |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 22.5  |           | 0.72 | 5.00       | ug/Kg             |
| 75-35-4        | 1,1-Dichloroethene             | 22.4  |           | 0.86 | 5.00       | ug/Kg             |
| 67-64-1        | Acetone                        | 100   |           | 12.2 | 25.0       | ug/Kg             |
| 75-15-0        | Carbon Disulfide               | 22.3  |           | 0.75 | 5.00       | ug/Kg             |
| 1634-04-4      | Methyl tert-butyl Ether        | 21.2  |           | 0.93 | 5.00       | ug/Kg             |
| 79-20-9        | Methyl Acetate                 | 22.9  |           | 1.30 | 5.00       | ug/Kg             |
| 75-09-2        | Methylene Chloride             | 16.9  |           | 6.00 | 10.0       | ug/Kg             |
| 156-60-5       | trans-1,2-Dichloroethene       | 21.8  |           | 0.68 | 5.00       | ug/Kg             |
| 75-34-3        | 1,1-Dichloroethane             | 21.7  |           | 0.70 | 5.00       | ug/Kg             |
| 110-82-7       | Cyclohexane                    | 21.6  |           | 0.84 | 5.00       | ug/Kg             |
| 78-93-3        | 2-Butanone                     | 110   |           | 7.30 | 25.0       | ug/Kg             |
| 56-23-5        | Carbon Tetrachloride           | 20.2  |           | 0.79 | 5.00       | ug/Kg             |
| 156-59-2       | cis-1,2-Dichloroethene         | 21.5  |           | 0.68 | 5.00       | ug/Kg             |
| 74-97-5        | Bromochloromethane             | 24.2  |           | 0.81 | 5.00       | ug/Kg             |
| 67-66-3        | Chloroform                     | 21.3  |           | 0.67 | 5.00       | ug/Kg             |
| 71-55-6        | 1,1,1-Trichloroethane          | 21.3  |           | 0.75 | 5.00       | ug/Kg             |
| 108-87-2       | Methylcyclohexane              | 21.1  |           | 0.80 | 5.00       | ug/Kg             |
| 71-43-2        | Benzene                        | 21.6  |           | 0.66 | 5.00       | ug/Kg             |
| 107-06-2       | 1,2-Dichloroethane             | 21.1  |           | 0.84 | 5.00       | ug/Kg             |
| 79-01-6        | Trichloroethene                | 20.0  |           | 0.73 | 5.00       | ug/Kg             |
| 78-87-5        | 1,2-Dichloropropane            | 21.3  |           | 0.65 | 5.00       | ug/Kg             |
| 75-27-4        | Bromodichloromethane           | 21.5  |           | 0.70 | 5.00       | ug/Kg             |
| 108-10-1       | 4-Methyl-2-Pentanone           | 110   |           | 4.60 | 25.0       | ug/Kg             |
| 108-88-3       | Toluene                        | 21.3  |           | 0.63 | 5.00       | ug/Kg             |
| 10061-02-6     | t-1,3-Dichloropropene          | 19.7  |           | 0.74 | 5.00       | ug/Kg             |
| 10061-01-5     | cis-1,3-Dichloropropene        | 20.4  |           | 0.71 | 5.00       | ug/Kg             |

## Report of Analysis

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VD0123SBSD01                              | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBSD01                              | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    | % Solid:        | 100           |
| Sample Wt/Vol:     | 5 Units: g                                | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  | uL                                        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS ID : 0.18                         | Level :         | LOW           |
| Prep Method :      |                                           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075171.D        | 1         |           | 01/23/23 13:08 | VD012323      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|---------------------------|-----------------------------|--------|-----------|----------|------------|-------------------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 21.3   |           | 0.86     | 5.00       | ug/Kg             |
| 591-78-6                  | 2-Hexanone                  | 110    |           | 4.70     | 25.0       | ug/Kg             |
| 124-48-1                  | Dibromochloromethane        | 20.9   |           | 0.75     | 5.00       | ug/Kg             |
| 106-93-4                  | 1,2-Dibromoethane           | 21.2   |           | 0.75     | 5.00       | ug/Kg             |
| 127-18-4                  | Tetrachloroethene           | 20.1   |           | 0.76     | 5.00       | ug/Kg             |
| 108-90-7                  | Chlorobenzene               | 20.5   |           | 0.65     | 5.00       | ug/Kg             |
| 100-41-4                  | Ethyl Benzene               | 21.0   |           | 0.70     | 5.00       | ug/Kg             |
| 179601-23-1               | m/p-Xylenes                 | 42.9   |           | 1.50     | 10.0       | ug/Kg             |
| 95-47-6                   | o-Xylene                    | 20.4   |           | 0.79     | 5.00       | ug/Kg             |
| 100-42-5                  | Styrene                     | 21.5   |           | 0.79     | 5.00       | ug/Kg             |
| 75-25-2                   | Bromoform                   | 20.2   |           | 0.81     | 5.00       | ug/Kg             |
| 98-82-8                   | Isopropylbenzene            | 20.1   |           | 0.72     | 5.00       | ug/Kg             |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 21.6   |           | 1.10     | 5.00       | ug/Kg             |
| 541-73-1                  | 1,3-Dichlorobenzene         | 20.2   |           | 0.67     | 5.00       | ug/Kg             |
| 106-46-7                  | 1,4-Dichlorobenzene         | 20.6   |           | 0.63     | 5.00       | ug/Kg             |
| 95-50-1                   | 1,2-Dichlorobenzene         | 20.6   |           | 0.64     | 5.00       | ug/Kg             |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 22.7   |           | 1.20     | 5.00       | ug/Kg             |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 19.0   |           | 0.94     | 5.00       | ug/Kg             |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 18.8   |           | 1.00     | 5.00       | ug/Kg             |
| <b>SURROGATES</b>         |                             |        |           |          |            |                   |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 53.7   |           | 50 - 163 | 107%       | SPK: 50           |
| 1868-53-7                 | Dibromofluoromethane        | 51.7   |           | 54 - 147 | 103%       | SPK: 50           |
| 2037-26-5                 | Toluene-d8                  | 52.9   |           | 58 - 134 | 106%       | SPK: 50           |
| 460-00-4                  | 4-Bromofluorobenzene        | 51.2   |           | 30 - 143 | 102%       | SPK: 50           |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |                   |
| 363-72-4                  | Pentafluorobenzene          | 72400  | 7.875     |          |            |                   |
| 540-36-3                  | 1,4-Difluorobenzene         | 130000 | 8.781     |          |            |                   |
| 3114-55-4                 | Chlorobenzene-d5            | 116000 | 11.587    |          |            |                   |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 53200  | 13.522    |          |            |                   |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VD0123SBSD01                              |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VD0123SBSD01                              |           | Matrix:         | SOIL          |
| Analytical Method: | SW8260                                    |           | % Solid:        | 100           |
| Sample Wt/Vol:     | 5                                         | Units: g  | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RTX-VMS                                   | ID : 0.18 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VD075171.D        | 1         |           | 01/23/23 13:08 | VD012323      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0120WBSD01                              |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBSD01                              |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076341.D        | 1         |           | 01/20/23 12:25 | VN012023      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|------|------------|-------|
| <b>TARGETS</b> |                                |       |           |      |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 18.1  |           | 0.17 | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 17.3  |           | 0.20 | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 18.3  |           | 0.22 | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 18.7  |           | 1.60 | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 18.9  |           | 0.26 | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 18.0  |           | 0.20 | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 18.7  |           | 0.17 | 1.00       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 17.4  |           | 0.23 | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 91.0  |           | 1.20 | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 16.3  |           | 0.26 | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 19.9  |           | 0.18 | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 18.5  |           | 0.53 | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 20.5  |           | 0.18 | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 18.0  |           | 0.20 | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 18.4  |           | 0.20 | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 17.4  |           | 1.20 | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 92.2  |           | 0.82 | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 18.7  |           | 0.18 | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 18.6  |           | 0.17 | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 20.4  |           | 0.19 | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 18.8  |           | 0.18 | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 18.5  |           | 0.18 | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 19.1  |           | 0.13 | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 18.8  |           | 0.16 | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 19.5  |           | 0.18 | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 18.0  |           | 0.27 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 19.2  |           | 0.17 | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 19.5  |           | 0.18 | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 98.9  |           | 0.87 | 5.00       | ug/L  |
| 108-88-3       | Toluene                        | 18.5  |           | 0.17 | 1.00       | ug/L  |
| 10061-02-6     | t-1,3-Dichloropropene          | 19.8  |           | 0.14 | 1.00       | ug/L  |
| 10061-01-5     | cis-1,3-Dichloropropene        | 19.4  |           | 0.16 | 1.00       | ug/L  |

## Report of Analysis

|                    |                                           |           |                 |               |
|--------------------|-------------------------------------------|-----------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    |           | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR |           | Date Received:  |               |
| Client Sample ID:  | VN0120WBSD01                              |           | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBSD01                              |           | Matrix:         | Water         |
| Analytical Method: | SW8260                                    |           | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units: mL | Final Vol:      | 5000 uL       |
| Soil Aliquot Vol:  |                                           | uL        | Test:           | VOC-TCLVOA-10 |
| GC Column:         | RXI-624                                   | ID : 0.25 | Level :         | LOW           |
| Prep Method :      |                                           |           |                 |               |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076341.D        | 1         |           | 01/20/23 12:25 | VN012023      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 79-00-5                   | 1,1,2-Trichloroethane       | 19.5   |           | 0.19     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 98.1   |           | 0.76     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 20.2   |           | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 19.4   |           | 0.14     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 18.4   |           | 0.18     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 18.8   |           | 0.17     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 19.1   |           | 0.17     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 38.2   |           | 0.33     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 19.2   |           | 0.18     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 19.4   |           | 0.13     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 19.8   |           | 0.16     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 19.9   |           | 0.19     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 19.3   |           | 0.23     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 18.1   |           | 0.20     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 18.8   |           | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 18.1   |           | 0.17     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 19.9   |           | 0.42     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 18.4   |           | 0.23     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 18.5   |           | 0.33     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 51.3   |           | 74 - 125 | 103%       | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 51.9   |           | 75 - 124 | 104%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 51.4   |           | 86 - 113 | 103%       | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 53.4   |           | 64 - 133 | 107%       | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 360000 | 8.23      |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 568000 | 9.106     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 510000 | 11.871    |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 238000 | 13.794    |          |            |         |

**Report of Analysis**

|                    |                                           |                 |               |
|--------------------|-------------------------------------------|-----------------|---------------|
| Client:            | Louis Berger U.S., Inc., A WSP Company    | Date Collected: |               |
| Project:           | NYCDDC Phase II SCI Arthur Kill Road CEQR | Date Received:  |               |
| Client Sample ID:  | VN0120WBSD01                              | SDG No.:        | O1233         |
| Lab Sample ID:     | VN0120WBSD01                              | Matrix:         | Water         |
| Analytical Method: | SW8260                                    | % Solid:        | 0             |
| Sample Wt/Vol:     | 5                                         | Units:          | mL            |
| Soil Aliquot Vol:  |                                           |                 | uL            |
| GC Column:         | RXI-624                                   | ID :            | 0.25          |
| Prep Method :      |                                           | Level :         | LOW           |
|                    |                                           | Final Vol:      | 5000 uL       |
|                    |                                           | Test:           | VOC-TCLVOA-10 |

|                   |           |           |                |               |
|-------------------|-----------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed  | Prep Batch ID |
| VN076341.D        | 1         |           | 01/20/23 12:25 | VN012023      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

# CALIBRATION SUMMARY

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233  
 Instrument ID: MSVOA\_D Calibration Date(s): 01/18/2023 01/18/2023  
 Heated Purge: (Y/N) Y Calibration Time(s): 10:25 13:10  
 GC Column: RTX-VMS ID: 0.18 (mm)

| LAB FILE ID:                   | RRF005 = VD075124.D | RRF010 = VD075125.D | RRF020 = VD075126.D |        |        |        |       |       |
|--------------------------------|---------------------|---------------------|---------------------|--------|--------|--------|-------|-------|
|                                | RRF050 = VD075127.D | RRF100 = VD075128.D | RRF150 = VD075129.D |        |        |        |       |       |
| COMPOUND                       | RRF005              | RRF010              | RRF020              | RRF050 | RRF100 | RRF150 | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.476               | 0.507               | 0.518               | 0.365  | 0.373  | 0.374  | 0.435 | 16.6  |
| Chloromethane                  | 0.725               | 0.734               | 0.686               | 0.608  | 0.556  | 0.566  | 0.646 | 12.3  |
| Vinyl Chloride                 | 0.741               | 0.724               | 0.728               | 0.656  | 0.641  | 0.668  | 0.693 | 6.2   |
| Bromomethane                   | 0.534               | 0.516               | 0.461               | 0.451  | 0.415  | 0.451  | 0.471 | 9.5   |
| Chloroethane                   | 0.496               | 0.511               | 0.487               | 0.473  | 0.445  | 0.471  | 0.481 | 4.7   |
| Trichlorofluoromethane         | 0.959               | 0.897               | 0.901               | 0.850  | 0.832  | 0.847  | 0.881 | 5.4   |
| 1,1,2-Trichlorotrifluoroethane | 0.580               | 0.547               | 0.579               | 0.517  | 0.519  | 0.521  | 0.544 | 5.5   |
| 1,1-Dichloroethene             | 0.596               | 0.603               | 0.595               | 0.563  | 0.537  | 0.541  | 0.573 | 5.2   |
| Acetone                        | 0.109               | 0.139               | 0.126               | 0.160  | 0.136  | 0.138  | 0.135 | 12.6  |
| Carbon Disulfide               | 1.982               | 1.978               | 1.942               | 1.832  | 1.710  | 1.698  | 1.857 | 7     |
| Methyl tert-butyl Ether        | 1.168               | 1.255               | 1.164               | 1.290  | 1.128  | 1.170  | 1.196 | 5.2   |
| Methyl Acetate                 | 0.390               | 0.401               | 0.353               | 0.387  | 0.325  | 0.338  | 0.366 | 8.5   |
| Methylene Chloride             | 0.972               | 1.161               | 0.911               | 0.794  | 0.629  | 0.625  | 0.849 | 24.6  |
| trans-1,2-Dichloroethene       | 0.695               | 0.676               | 0.631               | 0.657  | 0.583  | 0.583  | 0.637 | 7.4   |
| 1,1-Dichloroethane             | 1.130               | 1.181               | 1.083               | 1.112  | 1.002  | 1.012  | 1.087 | 6.4   |
| Cyclohexane                    | 1.295               | 1.059               | 1.049               | 0.935  | 0.895  | 0.907  | 1.023 | 14.7  |
| 2-Butanone                     | 0.159               | 0.176               | 0.150               | 0.182  | 0.151  | 0.158  | 0.163 | 8.2   |
| Carbon Tetrachloride           | 0.398               | 0.375               | 0.404               | 0.406  | 0.434  | 0.435  | 0.409 | 5.6   |
| cis-1,2-Dichloroethene         | 0.720               | 0.728               | 0.707               | 0.726  | 0.648  | 0.640  | 0.695 | 5.8   |
| Bromochloromethane             | 0.312               | 0.338               | 0.320               | 0.297  | 0.258  | 0.255  | 0.297 | 11.3  |
| Chloroform                     | 1.169               | 1.159               | 1.067               | 1.122  | 0.986  | 0.999  | 1.084 | 7.3   |
| 1,1,1-Trichloroethane          | 0.981               | 0.987               | 0.926               | 0.937  | 0.878  | 0.894  | 0.934 | 4.7   |
| Methylcyclohexane              | 0.646               | 0.590               | 0.599               | 0.570  | 0.597  | 0.588  | 0.598 | 4.2   |
| Benzene                        | 1.458               | 1.425               | 1.368               | 1.438  | 1.328  | 1.290  | 1.384 | 4.8   |
| 1,2-Dichloroethane             | 0.359               | 0.376               | 0.350               | 0.384  | 0.355  | 0.358  | 0.364 | 3.6   |
| Trichloroethene                | 0.410               | 0.401               | 0.373               | 0.400  | 0.376  | 0.374  | 0.389 | 4.2   |
| 1,2-Dichloropropane            | 0.365               | 0.355               | 0.323               | 0.355  | 0.327  | 0.321  | 0.341 | 5.7   |
| Bromodichloromethane           | 0.449               | 0.455               | 0.417               | 0.473  | 0.443  | 0.440  | 0.446 | 4.2   |
| 4-Methyl-2-Pentanone           | 0.171               | 0.192               | 0.172               | 0.192  | 0.180  | 0.190  | 0.183 | 5.4   |
| Toluene                        | 0.912               | 0.906               | 0.864               | 0.916  | 0.829  | 0.822  | 0.875 | 4.9   |
| t-1,3-Dichloropropene          | 0.402               | 0.426               | 0.399               | 0.447  | 0.426  | 0.433  | 0.422 | 4.4   |
| cis-1,3-Dichloropropene        | 0.486               | 0.519               | 0.501               | 0.556  | 0.511  | 0.521  | 0.516 | 4.6   |
| 1,1,2-Trichloroethane          | 0.243               | 0.245               | 0.229               | 0.255  | 0.225  | 0.232  | 0.238 | 4.7   |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                     |                 |                      |                   |
|---------------------|-----------------|----------------------|-------------------|
| Lab Name:           | <u>CHEMTECH</u> | Contract:            | <u>loui01</u>     |
| Lab Code:           | <u>CHEM</u>     | Case No.:            | <u>O1233</u>      |
| Instrument ID:      | <u>MSVOA_D</u>  | SAS No.:             | <u>O1233</u>      |
|                     |                 | SDG No.:             | <u>O1233</u>      |
| Heated Purge: (Y/N) | <u>Y</u>        | Calibration Date(s): | <u>01/18/2023</u> |
|                     |                 |                      | <u>01/18/2023</u> |
|                     |                 | Calibration Time(s): | <u>10:25</u>      |
|                     |                 |                      | <u>13:10</u>      |
| GC Column:          | <u>RTX-VMS</u>  | ID:                  | <u>0.18</u> (mm)  |

|                             |        |                     |        |                     |        |                     |       |       |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                |        | RRF005 = VD075124.D |        | RRF010 = VD075125.D |        | RRF020 = VD075126.D |       |       |
|                             |        | RRF050 = VD075127.D |        | RRF100 = VD075128.D |        | RRF150 = VD075129.D |       |       |
| COMPOUND                    | RRF005 | RRF010              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |
| 2-Hexanone                  | 0.115  | 0.140               | 0.123  | 0.144               | 0.131  | 0.140               | 0.132 | 8.7   |
| Dibromochloromethane        | 0.279  | 0.299               | 0.262  | 0.310               | 0.283  | 0.288               | 0.287 | 5.7   |
| 1,2-Dibromoethane           | 0.234  | 0.228               | 0.221  | 0.242               | 0.217  | 0.222               | 0.227 | 4     |
| Tetrachloroethene           | 0.380  | 0.363               | 0.358  | 0.350               | 0.346  | 0.344               | 0.357 | 3.8   |
| Chlorobenzene               | 1.129  | 1.025               | 1.005  | 1.054               | 0.972  | 0.962               | 1.025 | 6     |
| Ethyl Benzene               | 1.914  | 1.826               | 1.855  | 1.865               | 1.805  | 1.800               | 1.844 | 2.3   |
| m/p-Xylenes                 | 0.755  | 0.701               | 0.722  | 0.718               | 0.689  | 0.690               | 0.712 | 3.5   |
| o-Xylene                    | 0.687  | 0.668               | 0.661  | 0.696               | 0.647  | 0.651               | 0.668 | 2.9   |
| Styrene                     | 1.119  | 1.139               | 1.102  | 1.153               | 1.073  | 1.079               | 1.111 | 2.9   |
| Bromoform                   | 0.166  | 0.176               | 0.166  | 0.188               | 0.184  | 0.187               | 0.178 | 5.7   |
| Isopropylbenzene            | 4.215  | 3.842               | 3.974  | 3.847               | 3.954  | 3.864               | 3.949 | 3.6   |
| 1,1,2,2-Tetrachloroethane   | 0.601  | 0.674               | 0.626  | 0.645               | 0.614  | 0.615               | 0.629 | 4.2   |
| 1,3-Dichlorobenzene         | 1.817  | 1.801               | 1.783  | 1.756               | 1.684  | 1.651               | 1.749 | 3.8   |
| 1,4-Dichlorobenzene         | 1.818  | 1.764               | 1.700  | 1.721               | 1.676  | 1.644               | 1.720 | 3.6   |
| 1,2-Dichlorobenzene         | 1.532  | 1.491               | 1.515  | 1.532               | 1.454  | 1.453               | 1.496 | 2.4   |
| 1,2-Dibromo-3-Chloropropane | 0.084  | 0.095               | 0.095  | 0.102               | 0.097  | 0.104               | 0.096 | 7.4   |
| 1,2,4-Trichlorobenzene      | 0.983  | 0.990               | 0.933  | 0.997               | 0.987  | 1.005               | 0.982 | 2.6   |
| 1,2,3-Trichlorobenzene      | 0.813  | 0.833               | 0.820  | 0.854               | 0.847  | 0.863               | 0.838 | 2.3   |
| 1,2-Dichloroethane-d4       | 0.524  | 0.543               | 0.501  | 0.537               | 0.472  | 0.485               | 0.510 | 5.7   |
| Dibromofluoromethane        | 0.288  | 0.312               | 0.301  | 0.322               | 0.297  | 0.293               | 0.302 | 4.2   |
| Toluene-d8                  | 1.228  | 1.239               | 1.152  | 1.232               | 1.153  | 1.133               | 1.190 | 4.1   |
| 4-Bromofluorobenzene        | 0.409  | 0.376               | 0.363  | 0.397               | 0.357  | 0.363               | 0.378 | 5.6   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                                                |                                                          |
|------------------------------------------------|----------------------------------------------------------|
| Lab Name: <u>CHEMTECH</u>                      | Contract: <u>loui01</u>                                  |
| Lab Code: <u>CHEM</u> Case No.: <u>O1233</u>   | SAS No.: <u>O1233</u> SDG No.: <u>O1233</u>              |
| Instrument ID: <u>MSVOA_N</u>                  | Calibration Date(s): <u>01/16/2023</u> <u>01/16/2023</u> |
| Heated Purge: (Y/N) <u>N</u>                   | Calibration Time(s): <u>10:54</u> <u>13:17</u>           |
| GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm) |                                                          |

| LAB FILE ID:                   | RRF005 = VN076261.D | RRF010 = VN076262.D | RRF020 = VN076263.D |        |        |        |       |       |
|--------------------------------|---------------------|---------------------|---------------------|--------|--------|--------|-------|-------|
|                                | RRF050 = VN076264.D | RRF075 = VN076265.D | RRF001 = VN076267.D |        |        |        |       |       |
| COMPOUND                       | RRF005              | RRF010              | RRF020              | RRF050 | RRF075 | RRF001 | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.452               | 0.537               | 0.523               | 0.526  | 0.541  | 0.496  | 0.512 | 6.6   |
| Chloromethane                  | 0.503               | 0.544               | 0.523               | 0.522  | 0.522  | 0.569  | 0.531 | 4.3   |
| Vinyl Chloride                 | 0.658               | 0.653               | 0.644               | 0.644  | 0.657  | 0.647  | 0.650 | 1     |
| Bromomethane                   | 0.558               | 0.534               | 0.525               | 0.515  | 0.532  |        | 0.533 | 3     |
| Chloroethane                   | 0.475               | 0.484               | 0.471               | 0.457  | 0.487  | 0.553  | 0.488 | 6.9   |
| Trichlorofluoromethane         | 0.887               | 0.887               | 0.880               | 0.875  | 0.898  | 0.942  | 0.895 | 2.7   |
| 1,1,2-Trichlorotrifluoroethane | 0.511               | 0.531               | 0.502               | 0.498  | 0.511  | 0.487  | 0.507 | 3     |
| 1,1-Dichloroethene             | 0.470               | 0.479               | 0.459               | 0.477  | 0.486  | 0.503  | 0.479 | 3.1   |
| Acetone                        | 0.187               | 0.185               | 0.173               | 0.175  | 0.182  | 0.269  | 0.195 | 18.8  |
| Carbon Disulfide               | 1.073               | 1.179               | 1.141               | 1.184  | 1.228  | 1.350  | 1.192 | 7.8   |
| Methyl tert-butyl Ether        | 1.517               | 1.535               | 1.537               | 1.597  | 1.633  | 1.535  | 1.559 | 2.9   |
| Methyl Acetate                 | 0.618               | 0.606               | 0.574               | 0.588  | 0.599  | 0.664  | 0.608 | 5.1   |
| Methylene Chloride             | 0.579               | 0.550               | 0.543               | 0.517  | 0.535  | 1.144  | 0.645 | 38.1  |
| trans-1,2-Dichloroethene       | 0.520               | 0.501               | 0.500               | 0.506  | 0.513  | 0.573  | 0.519 | 5.4   |
| 1,1-Dichloroethane             | 0.910               | 0.897               | 0.854               | 0.868  | 0.893  | 0.842  | 0.877 | 3.1   |
| Cyclohexane                    | 1.007               | 0.875               | 0.833               | 0.787  | 0.798  |        | 0.860 | 10.4  |
| 2-Butanone                     | 0.265               | 0.275               | 0.263               | 0.273  | 0.278  | 0.290  | 0.274 | 3.5   |
| Carbon Tetrachloride           | 0.489               | 0.490               | 0.486               | 0.499  | 0.514  | 0.496  | 0.496 | 2     |
| cis-1,2-Dichloroethene         | 0.615               | 0.591               | 0.597               | 0.599  | 0.614  | 0.614  | 0.605 | 1.7   |
| Bromochloromethane             | 0.364               | 0.333               | 0.353               | 0.362  | 0.380  | 0.342  | 0.356 | 4.7   |
| Chloroform                     | 0.966               | 0.972               | 0.931               | 0.965  | 0.978  | 1.013  | 0.971 | 2.7   |
| 1,1,1-Trichloroethane          | 0.885               | 0.878               | 0.870               | 0.884  | 0.924  | 0.876  | 0.886 | 2.2   |
| Methylcyclohexane              | 0.532               | 0.525               | 0.557               | 0.572  | 0.597  | 0.474  | 0.543 | 7.9   |
| Benzene                        | 1.325               | 1.322               | 1.316               | 1.333  | 1.358  | 1.320  | 1.329 | 1.1   |
| 1,2-Dichloroethane             | 0.458               | 0.448               | 0.441               | 0.450  | 0.460  | 0.471  | 0.455 | 2.4   |
| Trichloroethene                | 0.398               | 0.372               | 0.354               | 0.361  | 0.370  | 0.399  | 0.376 | 5     |
| 1,2-Dichloropropane            | 0.326               | 0.315               | 0.314               | 0.315  | 0.325  | 0.315  | 0.318 | 1.8   |
| Bromodichloromethane           | 0.450               | 0.459               | 0.456               | 0.474  | 0.486  | 0.471  | 0.466 | 2.8   |
| 4-Methyl-2-Pentanone           | 0.327               | 0.341               | 0.341               | 0.356  | 0.367  | 0.310  | 0.340 | 5.9   |
| Toluene                        | 0.891               | 0.894               | 0.870               | 0.885  | 0.909  | 0.883  | 0.889 | 1.5   |
| t-1,3-Dichloropropene          | 0.409               | 0.426               | 0.442               | 0.489  | 0.514  | 0.386  | 0.444 | 10.9  |
| cis-1,3-Dichloropropene        | 0.468               | 0.506               | 0.505               | 0.531  | 0.565  | 0.490  | 0.511 | 6.6   |
| 1,1,2-Trichloroethane          | 0.340               | 0.336               | 0.330               | 0.331  | 0.350  | 0.350  | 0.340 | 2.6   |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                     |                 |                      |                   |
|---------------------|-----------------|----------------------|-------------------|
| Lab Name:           | <u>CHEMTECH</u> | Contract:            | <u>loui01</u>     |
| Lab Code:           | <u>CHEM</u>     | Case No.:            | <u>O1233</u>      |
| Instrument ID:      | <u>MSVOA_N</u>  | SAS No.:             | <u>O1233</u>      |
|                     |                 | SDG No.:             | <u>O1233</u>      |
| Heated Purge: (Y/N) | <u>N</u>        | Calibration Date(s): | <u>01/16/2023</u> |
|                     |                 |                      | <u>01/16/2023</u> |
|                     |                 | Calibration Time(s): | <u>10:54</u>      |
|                     |                 |                      | <u>13:17</u>      |
| GC Column:          | <u>RXI-624</u>  | ID:                  | <u>0.25</u> (mm)  |

|                             |        |                     |        |                     |        |                     |       |       |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                |        | RRF005 = VN076261.D |        | RRF010 = VN076262.D |        | RRF020 = VN076263.D |       |       |
|                             |        | RRF050 = VN076264.D |        | RRF075 = VN076265.D |        | RRF001 = VN076267.D |       |       |
| COMPOUND                    | RRF005 | RRF010              | RRF020 | RRF050              | RRF075 | RRF001              | RRF   | % RSD |
| 2-Hexanone                  | 0.232  | 0.242               | 0.249  | 0.261               | 0.269  | 0.228               | 0.247 | 6.5   |
| Dibromochloromethane        | 0.342  | 0.354               | 0.358  | 0.382               | 0.400  | 0.310               | 0.358 | 8.8   |
| 1,2-Dibromoethane           | 0.334  | 0.340               | 0.335  | 0.350               | 0.361  | 0.314               | 0.339 | 4.7   |
| Tetrachloroethene           | 0.501  | 0.460               | 0.422  | 0.393               | 0.388  | 0.541               | 0.451 | 13.6  |
| Chlorobenzene               | 1.040  | 1.054               | 1.017  | 1.034               | 1.051  | 1.146               | 1.057 | 4.3   |
| Ethyl Benzene               | 1.707  | 1.820               | 1.814  | 1.880               | 1.921  | 1.711               | 1.809 | 4.8   |
| m/p-Xylenes                 | 0.678  | 0.727               | 0.721  | 0.745               | 0.771  | 0.710               | 0.725 | 4.4   |
| o-Xylene                    | 0.670  | 0.706               | 0.710  | 0.739               | 0.751  | 0.699               | 0.712 | 4.1   |
| Styrene                     | 1.037  | 1.097               | 1.137  | 1.206               | 1.250  | 1.096               | 1.137 | 6.9   |
| Bromoform                   | 0.252  | 0.270               | 0.282  | 0.304               | 0.323  | 0.273               | 0.284 | 9     |
| Isopropylbenzene            | 3.802  | 4.040               | 3.943  | 3.806               | 3.824  | 3.793               | 3.868 | 2.6   |
| 1,1,2,2-Tetrachloroethane   | 1.108  | 1.079               | 1.071  | 1.057               | 1.047  | 1.309               | 1.112 | 8.9   |
| 1,3-Dichlorobenzene         | 1.692  | 1.760               | 1.675  | 1.724               | 1.759  | 2.328               | 1.823 | 13.7  |
| 1,4-Dichlorobenzene         | 1.740  | 1.728               | 1.692  | 1.701               | 1.752  | 2.439               | 1.842 | 15.9  |
| 1,2-Dichlorobenzene         | 1.766  | 1.760               | 1.703  | 1.716               | 1.725  | 2.224               | 1.816 | 11.1  |
| 1,2-Dibromo-3-Chloropropane | 0.182  | 0.180               | 0.187  | 0.207               | 0.201  | 0.204               | 0.193 | 6.2   |
| 1,2,4-Trichlorobenzene      | 0.652  | 0.757               | 0.760  | 0.875               | 0.892  | 1.282               | 0.870 | 25.3  |
| 1,2,3-Trichlorobenzene      | 0.648  | 0.758               | 0.730  | 0.832               | 0.850  | 1.121               | 0.823 | 19.8  |
| 1,2-Dichloroethane-d4       | 0.584  | 0.526               | 0.549  | 0.554               | 0.586  |                     | 0.560 | 4.5   |
| Dibromofluoromethane        | 0.331  | 0.295               | 0.312  | 0.306               | 0.321  |                     | 0.313 | 4.4   |
| Toluene-d8                  | 1.176  | 1.062               | 1.151  | 1.165               | 1.222  |                     | 1.155 | 5.1   |
| 4-Bromofluorobenzene        | 0.356  | 0.334               | 0.371  | 0.393               | 0.431  |                     | 0.377 | 9.8   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                                                |                                                          |
|------------------------------------------------|----------------------------------------------------------|
| Lab Name: <u>CHEMTECH</u>                      | Contract: <u>loui01</u>                                  |
| Lab Code: <u>CHEM</u> Case No.: <u>O1233</u>   | SAS No.: <u>O1233</u> SDG No.: <u>O1233</u>              |
| Instrument ID: <u>MSVOA_Y</u>                  | Calibration Date(s): <u>01/16/2023</u> <u>01/16/2023</u> |
| Heated Purge: (Y/N) <u>Y</u>                   | Calibration Time(s): <u>12:17</u> <u>14:34</u>           |
| GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm) |                                                          |

| LAB FILE ID:                   | RRF005 = VY012194.D | RRF010 = VY012195.D | RRF020 = VY012196.D |        |        |        |       |       |
|--------------------------------|---------------------|---------------------|---------------------|--------|--------|--------|-------|-------|
|                                | RRF050 = VY012197.D | RRF100 = VY012198.D | RRF150 = VY012199.D |        |        |        |       |       |
| COMPOUND                       | RRF005              | RRF010              | RRF020              | RRF050 | RRF100 | RRF150 | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.456               | 0.520               | 0.486               | 0.372  | 0.410  | 0.378  | 0.437 | 13.8  |
| Chloromethane                  | 0.452               | 0.440               | 0.400               | 0.368  | 0.389  | 0.372  | 0.403 | 8.7   |
| Vinyl Chloride                 | 0.503               | 0.508               | 0.479               | 0.435  | 0.464  | 0.436  | 0.471 | 6.8   |
| Bromomethane                   | 0.428               | 0.398               | 0.355               | 0.334  | 0.352  | 0.317  | 0.364 | 11.4  |
| Chloroethane                   | 0.324               | 0.324               | 0.310               | 0.287  | 0.310  | 0.288  | 0.307 | 5.3   |
| Trichlorofluoromethane         | 0.951               | 0.994               | 0.945               | 0.842  | 0.929  | 0.853  | 0.919 | 6.5   |
| 1,1,2-Trichlorotrifluoroethane | 0.454               | 0.481               | 0.455               | 0.407  | 0.457  | 0.430  | 0.447 | 5.7   |
| 1,1-Dichloroethene             | 0.470               | 0.477               | 0.459               | 0.438  | 0.469  | 0.443  | 0.459 | 3.4   |
| Acetone                        | 0.111               | 0.149               | 0.123               | 0.154  | 0.149  | 0.150  | 0.139 | 12.9  |
| Carbon Disulfide               | 1.466               | 1.490               | 1.413               | 1.343  | 1.443  | 1.367  | 1.420 | 4     |
| Methyl tert-butyl Ether        | 1.159               | 1.395               | 1.250               | 1.280  | 1.325  | 1.273  | 1.280 | 6.1   |
| Methyl Acetate                 | 0.308               | 0.340               | 0.287               | 0.301  | 0.308  | 0.326  | 0.312 | 6     |
| Methylene Chloride             | 1.174               | 1.020               | 0.736               | 0.570  | 0.559  | 0.513  | 0.762 | 36.1  |
| trans-1,2-Dichloroethene       | 0.519               | 0.543               | 0.512               | 0.495  | 0.519  | 0.487  | 0.512 | 3.9   |
| 1,1-Dichloroethane             | 0.804               | 0.812               | 0.772               | 0.760  | 0.805  | 0.759  | 0.785 | 3.1   |
| Cyclohexane                    | 0.845               | 0.795               | 0.707               | 0.633  | 0.713  | 0.681  | 0.729 | 10.6  |
| 2-Butanone                     | 0.125               | 0.175               | 0.143               | 0.157  | 0.156  | 0.166  | 0.154 | 11.5  |
| Carbon Tetrachloride           | 0.600               | 0.615               | 0.594               | 0.552  | 0.583  | 0.543  | 0.581 | 4.9   |
| cis-1,2-Dichloroethene         | 0.571               | 0.583               | 0.555               | 0.548  | 0.582  | 0.538  | 0.563 | 3.3   |
| Bromochloromethane             | 0.214               | 0.177               | 0.160               | 0.159  | 0.160  | 0.158  | 0.171 | 12.8  |
| Chloroform                     | 0.940               | 0.997               | 0.912               | 0.900  | 0.952  | 0.860  | 0.927 | 5.1   |
| 1,1,1-Trichloroethane          | 0.932               | 0.969               | 0.901               | 0.885  | 0.950  | 0.868  | 0.918 | 4.3   |
| Methylcyclohexane              | 0.549               | 0.566               | 0.556               | 0.514  | 0.564  | 0.550  | 0.550 | 3.4   |
| Benzene                        | 1.354               | 1.375               | 1.297               | 1.237  | 1.289  | 1.239  | 1.299 | 4.4   |
| 1,2-Dichloroethane             | 0.404               | 0.453               | 0.413               | 0.408  | 0.414  | 0.389  | 0.413 | 5.2   |
| Trichloroethene                | 0.404               | 0.405               | 0.395               | 0.372  | 0.387  | 0.366  | 0.388 | 4.2   |
| 1,2-Dichloropropane            | 0.267               | 0.288               | 0.268               | 0.261  | 0.277  | 0.266  | 0.271 | 3.6   |
| Bromodichloromethane           | 0.437               | 0.488               | 0.460               | 0.446  | 0.468  | 0.440  | 0.457 | 4.2   |
| 4-Methyl-2-Pentanone           | 0.165               | 0.218               | 0.193               | 0.199  | 0.195  | 0.211  | 0.197 | 9.3   |
| Toluene                        | 0.875               | 0.896               | 0.872               | 0.840  | 0.877  | 0.828  | 0.865 | 3     |
| t-1,3-Dichloropropene          | 0.454               | 0.501               | 0.478               | 0.475  | 0.488  | 0.470  | 0.478 | 3.3   |
| cis-1,3-Dichloropropene        | 0.492               | 0.547               | 0.514               | 0.508  | 0.525  | 0.500  | 0.514 | 3.8   |
| 1,1,2-Trichloroethane          | 0.252               | 0.286               | 0.255               | 0.254  | 0.253  | 0.246  | 0.258 | 5.5   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                                                |                                                          |
|------------------------------------------------|----------------------------------------------------------|
| Lab Name: <u>CHEMTECH</u>                      | Contract: <u>loui01</u>                                  |
| Lab Code: <u>CHEM</u> Case No.: <u>O1233</u>   | SAS No.: <u>O1233</u> SDG No.: <u>O1233</u>              |
| Instrument ID: <u>MSVOA_Y</u>                  | Calibration Date(s): <u>01/16/2023</u> <u>01/16/2023</u> |
| Heated Purge: (Y/N) <u>Y</u>                   | Calibration Time(s): <u>12:17</u> <u>14:34</u>           |
| GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm) |                                                          |

|                             |        |                     |        |                     |        |                     |       |       |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                |        | RRF005 = VY012194.D |        | RRF010 = VY012195.D |        | RRF020 = VY012196.D |       |       |
|                             |        | RRF050 = VY012197.D |        | RRF100 = VY012198.D |        | RRF150 = VY012199.D |       |       |
| COMPOUND                    | RRF005 | RRF010              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |
| 2-Hexanone                  | 0.115  | 0.166               | 0.145  | 0.158               | 0.153  | 0.163               | 0.150 | 12.5  |
| Dibromochloromethane        | 0.321  | 0.369               | 0.345  | 0.343               | 0.344  | 0.328               | 0.342 | 4.8   |
| 1,2-Dibromoethane           | 0.233  | 0.273               | 0.247  | 0.247               | 0.250  | 0.243               | 0.249 | 5.4   |
| Tetrachloroethene           | 0.461  | 0.494               | 0.468  | 0.451               | 0.447  | 0.417               | 0.456 | 5.6   |
| Chlorobenzene               | 1.097  | 1.064               | 1.013  | 0.990               | 1.013  | 0.961               | 1.023 | 4.8   |
| Ethyl Benzene               | 1.877  | 1.892               | 1.821  | 1.791               | 1.838  | 1.752               | 1.828 | 2.9   |
| m/p-Xylenes                 | 0.732  | 0.752               | 0.730  | 0.711               | 0.726  | 0.687               | 0.723 | 3.1   |
| o-Xylene                    | 0.706  | 0.715               | 0.695  | 0.680               | 0.692  | 0.654               | 0.690 | 3.1   |
| Styrene                     | 1.140  | 1.203               | 1.157  | 1.149               | 1.167  | 1.106               | 1.154 | 2.8   |
| Bromoform                   | 0.227  | 0.272               | 0.250  | 0.249               | 0.242  | 0.240               | 0.247 | 6     |
| Isopropylbenzene            | 3.672  | 3.672               | 3.604  | 3.501               | 3.715  | 3.469               | 3.605 | 2.8   |
| 1,1,2,2-Tetrachloroethane   | 0.544  | 0.622               | 0.546  | 0.563               | 0.561  | 0.572               | 0.568 | 5     |
| 1,3-Dichlorobenzene         | 1.837  | 1.799               | 1.736  | 1.665               | 1.725  | 1.591               | 1.726 | 5.2   |
| 1,4-Dichlorobenzene         | 1.888  | 1.818               | 1.695  | 1.642               | 1.705  | 1.577               | 1.721 | 6.6   |
| 1,2-Dichlorobenzene         | 1.660  | 1.639               | 1.521  | 1.485               | 1.511  | 1.422               | 1.540 | 6     |
| 1,2-Dibromo-3-Chloropropane | 0.103  | 0.133               | 0.116  | 0.116               | 0.115  | 0.121               | 0.117 | 8.3   |
| 1,2,4-Trichlorobenzene      | 1.049  | 0.995               | 0.986  | 0.932               | 0.977  | 0.980               | 0.987 | 3.8   |
| 1,2,3-Trichlorobenzene      | 0.870  | 0.862               | 0.864  | 0.798               | 0.837  | 0.862               | 0.849 | 3.3   |
| 1,2-Dichloroethane-d4       | 0.533  | 0.551               | 0.498  | 0.409               | 0.409  | 0.388               | 0.465 | 15.3  |
| Dibromofluoromethane        | 0.292  | 0.291               | 0.289  | 0.241               | 0.242  | 0.233               | 0.265 | 10.8  |
| Toluene-d8                  | 1.242  | 1.188               | 1.195  | 0.890               | 0.892  | 0.859               | 1.044 | 17.3  |
| 4-Bromofluorobenzene        | 0.425  | 0.419               | 0.403  | 0.333               | 0.336  | 0.323               | 0.373 | 12.7  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG No.: 01233  
 Instrument ID: MSVOA\_D Calibration Date/Time: 01/20/2023 11:13  
 Lab File ID: VD075146.D Init. Calib. Date(s): 01/18/2023 01/18/2023  
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:25 13:10  
 GC Column: RTX-VMS ID: 0.18 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Dichlorodifluoromethane        | 0.435 | 0.403  |            | -7.36  | 20    |
| Chloromethane                  | 0.646 | 0.592  | 0.1        | -8.36  | 20    |
| Vinyl Chloride                 | 0.693 | 0.650  |            | -6.2   | 20    |
| Bromomethane                   | 0.471 | 0.446  |            | -5.31  | 20    |
| Chloroethane                   | 0.481 | 0.471  |            | -2.08  | 20    |
| Trichlorofluoromethane         | 0.881 | 0.908  |            | 3.07   | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.544 | 0.579  |            | 6.43   | 20    |
| 1,1-Dichloroethene             | 0.573 | 0.596  |            | 4.01   | 20    |
| Acetone                        | 0.135 | 0.162  |            | 20     | 20    |
| Carbon Disulfide               | 1.857 | 1.937  |            | 4.31   | 20    |
| Methyl tert-butyl Ether        | 1.196 | 1.225  |            | 2.42   | 20    |
| Methyl Acetate                 | 0.366 | 0.370  |            | 1.09   | 20    |
| Methylene Chloride             | 0.849 | 0.689  |            | -18.85 | 20    |
| trans-1,2-Dichloroethene       | 0.637 | 0.665  |            | 4.4    | 20    |
| 1,1-Dichloroethane             | 1.087 | 1.152  | 0.1        | 5.98   | 20    |
| Cyclohexane                    | 1.023 | 1.045  |            | 2.15   | 20    |
| 2-Butanone                     | 0.163 | 0.183  |            | 12.27  | 20    |
| Carbon Tetrachloride           | 0.409 | 0.449  |            | 9.78   | 20    |
| cis-1,2-Dichloroethene         | 0.695 | 0.719  |            | 3.45   | 20    |
| Bromochloromethane             | 0.297 | 0.289  |            | -2.69  | 20    |
| Chloroform                     | 1.084 | 1.111  |            | 2.49   | 20    |
| 1,1,1-Trichloroethane          | 0.934 | 0.984  |            | 5.35   | 20    |
| Methylcyclohexane              | 0.598 | 0.658  |            | 10.03  | 20    |
| Benzene                        | 1.384 | 1.453  |            | 4.99   | 20    |
| 1,2-Dichloroethane             | 0.364 | 0.387  |            | 6.32   | 20    |
| Trichloroethene                | 0.389 | 0.402  |            | 3.34   | 20    |
| 1,2-Dichloropropane            | 0.341 | 0.366  |            | 7.33   | 20    |
| Bromodichloromethane           | 0.446 | 0.473  |            | 6.05   | 20    |
| 4-Methyl-2-Pentanone           | 0.183 | 0.194  |            | 6.01   | 20    |
| Toluene                        | 0.875 | 0.922  |            | 5.37   | 20    |
| t-1,3-Dichloropropene          | 0.422 | 0.441  |            | 4.5    | 20    |
| cis-1,3-Dichloropropene        | 0.516 | 0.551  |            | 6.78   | 20    |
| 1,1,2-Trichloroethane          | 0.238 | 0.249  |            | 4.62   | 20    |
| 2-Hexanone                     | 0.132 | 0.152  |            | 15.15  | 20    |
| Dibromochloromethane           | 0.287 | 0.301  |            | 4.88   | 20    |
| 1,2-Dibromoethane              | 0.227 | 0.233  |            | 2.64   | 20    |
| Tetrachloroethene              | 0.357 | 0.381  |            | 6.72   | 20    |
| Chlorobenzene                  | 1.025 | 1.070  | 0.3        | 4.39   | 20    |
| Ethyl Benzene                  | 1.844 | 1.986  |            | 7.7    | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233  
 Instrument ID: MSVOA\_D Calibration Date/Time: 01/20/2023 11:13  
 Lab File ID: VD075146.D Init. Calib. Date(s): 01/18/2023 01/18/2023  
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:25 13:10  
 GC Column: RTX-VMS ID: 0.18 (mm)

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------------|-------|--------|------------|-------|-------|
| m/p-Xylenes                 | 0.712 | 0.763  |            | 7.16  | 20    |
| o-Xylene                    | 0.668 | 0.712  |            | 6.59  | 20    |
| Styrene                     | 1.111 | 1.175  |            | 5.76  | 20    |
| Bromoform                   | 0.178 | 0.191  | 0.1        | 7.3   | 20    |
| Isopropylbenzene            | 3.949 | 4.198  |            | 6.3   | 20    |
| 1,1,2,2-Tetrachloroethane   | 0.629 | 0.668  | 0.3        | 6.2   | 20    |
| 1,3-Dichlorobenzene         | 1.749 | 1.801  |            | 2.97  | 20    |
| 1,4-Dichlorobenzene         | 1.720 | 1.777  |            | 3.31  | 20    |
| 1,2-Dichlorobenzene         | 1.496 | 1.536  |            | 2.67  | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.096 | 0.100  |            | 4.17  | 20    |
| 1,2,4-Trichlorobenzene      | 0.982 | 0.996  |            | 1.43  | 20    |
| 1,2,3-Trichlorobenzene      | 0.838 | 0.846  |            | 0.95  | 20    |
| 1,2-Dichloroethane-d4       | 0.510 | 0.508  |            | -0.39 | 20    |
| Dibromofluoromethane        | 0.302 | 0.312  |            | 3.31  | 20    |
| Toluene-d8                  | 1.190 | 1.231  |            | 3.44  | 20    |
| 4-Bromofluorobenzene        | 0.378 | 0.382  |            | 1.06  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG No.: 01233  
 Instrument ID: MSVOA\_D Calibration Date/Time: 01/23/2023 10:15  
 Lab File ID: VD075168.D Init. Calib. Date(s): 01/18/2023 01/18/2023  
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:25 13:10  
 GC Column: RTX-VMS ID: 0.18 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Dichlorodifluoromethane        | 0.435 | 0.374  |            | -14.02 | 20    |
| Chloromethane                  | 0.646 | 0.586  | 0.1        | -9.29  | 20    |
| Vinyl Chloride                 | 0.693 | 0.633  |            | -8.66  | 20    |
| Bromomethane                   | 0.471 | 0.424  |            | -9.98  | 20    |
| Chloroethane                   | 0.481 | 0.441  |            | -8.32  | 20    |
| Trichlorofluoromethane         | 0.881 | 0.868  |            | -1.48  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.544 | 0.536  |            | -1.47  | 20    |
| 1,1-Dichloroethene             | 0.573 | 0.563  |            | -1.75  | 20    |
| Acetone                        | 0.135 | 0.184  |            | 36.3   | 20    |
| Carbon Disulfide               | 1.857 | 1.830  |            | -1.45  | 20    |
| Methyl tert-butyl Ether        | 1.196 | 1.193  |            | -0.25  | 20    |
| Methyl Acetate                 | 0.366 | 0.363  |            | -0.82  | 20    |
| Methylene Chloride             | 0.849 | 0.628  |            | -26.03 | 20    |
| trans-1,2-Dichloroethene       | 0.637 | 0.641  |            | 0.63   | 20    |
| 1,1-Dichloroethane             | 1.087 | 1.102  | 0.1        | 1.38   | 20    |
| Cyclohexane                    | 1.023 | 0.970  |            | -5.18  | 20    |
| 2-Butanone                     | 0.163 | 0.195  |            | 19.63  | 20    |
| Carbon Tetrachloride           | 0.409 | 0.430  |            | 5.13   | 20    |
| cis-1,2-Dichloroethene         | 0.695 | 0.694  |            | -0.14  | 20    |
| Bromochloromethane             | 0.297 | 0.283  |            | -4.71  | 20    |
| Chloroform                     | 1.084 | 1.088  |            | 0.37   | 20    |
| 1,1,1-Trichloroethane          | 0.934 | 0.931  |            | -0.32  | 20    |
| Methylcyclohexane              | 0.598 | 0.611  |            | 2.17   | 20    |
| Benzene                        | 1.384 | 1.437  |            | 3.83   | 20    |
| 1,2-Dichloroethane             | 0.364 | 0.373  |            | 2.47   | 20    |
| Trichloroethene                | 0.389 | 0.385  |            | -1.03  | 20    |
| 1,2-Dichloropropane            | 0.341 | 0.349  |            | 2.35   | 20    |
| Bromodichloromethane           | 0.446 | 0.463  |            | 3.81   | 20    |
| 4-Methyl-2-Pentanone           | 0.183 | 0.190  |            | 3.83   | 20    |
| Toluene                        | 0.875 | 0.902  |            | 3.09   | 20    |
| t-1,3-Dichloropropene          | 0.422 | 0.432  |            | 2.37   | 20    |
| cis-1,3-Dichloropropene        | 0.516 | 0.533  |            | 3.3    | 20    |
| 1,1,2-Trichloroethane          | 0.238 | 0.242  |            | 1.68   | 20    |
| 2-Hexanone                     | 0.132 | 0.157  |            | 18.94  | 20    |
| Dibromochloromethane           | 0.287 | 0.288  |            | 0.35   | 20    |
| 1,2-Dibromoethane              | 0.227 | 0.226  |            | -0.44  | 20    |
| Tetrachloroethene              | 0.357 | 0.356  |            | -0.28  | 20    |
| Chlorobenzene                  | 1.025 | 1.046  | 0.3        | 2.05   | 20    |
| Ethyl Benzene                  | 1.844 | 1.911  |            | 3.63   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233  
 Instrument ID: MSVOA\_D Calibration Date/Time: 01/23/2023 10:15  
 Lab File ID: VD075168.D Init. Calib. Date(s): 01/18/2023 01/18/2023  
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:25 13:10  
 GC Column: RTX-VMS ID: 0.18 (mm)

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------------|-------|--------|------------|-------|-------|
| m/p-Xylenes                 | 0.712 | 0.735  |            | 3.23  | 20    |
| o-Xylene                    | 0.668 | 0.697  |            | 4.34  | 20    |
| Styrene                     | 1.111 | 1.138  |            | 2.43  | 20    |
| Bromoform                   | 0.178 | 0.184  | 0.1        | 3.37  | 20    |
| Isopropylbenzene            | 3.949 | 3.967  |            | 0.46  | 20    |
| 1,1,2,2-Tetrachloroethane   | 0.629 | 0.641  | 0.3        | 1.91  | 20    |
| 1,3-Dichlorobenzene         | 1.749 | 1.721  |            | -1.6  | 20    |
| 1,4-Dichlorobenzene         | 1.720 | 1.704  |            | -0.93 | 20    |
| 1,2-Dichlorobenzene         | 1.496 | 1.503  |            | 0.47  | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.096 | 0.097  |            | 1.04  | 20    |
| 1,2,4-Trichlorobenzene      | 0.982 | 0.896  |            | -8.76 | 20    |
| 1,2,3-Trichlorobenzene      | 0.838 | 0.768  |            | -8.35 | 20    |
| 1,2-Dichloroethane-d4       | 0.510 | 0.517  |            | 1.37  | 20    |
| Dibromofluoromethane        | 0.302 | 0.318  |            | 5.3   | 20    |
| Toluene-d8                  | 1.190 | 1.250  |            | 5.04  | 20    |
| 4-Bromofluorobenzene        | 0.378 | 0.377  |            | -0.26 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG No.: 01233  
 Instrument ID: MSVOA\_N Calibration Date/Time: 01/20/2023 10:27  
 Lab File ID: VN076337.D Init. Calib. Date(s): 01/16/2023 01/16/2023  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 10:54 13:17  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Dichlorodifluoromethane        | 0.512 | 0.472  |            | -7.81  | 20    |
| Chloromethane                  | 0.531 | 0.468  | 0.1        | -11.86 | 20    |
| Vinyl Chloride                 | 0.650 | 0.587  |            | -9.69  | 20    |
| Bromomethane                   | 0.533 | 0.487  |            | -8.63  | 20    |
| Chloroethane                   | 0.488 | 0.443  |            | -9.22  | 20    |
| Trichlorofluoromethane         | 0.895 | 0.825  |            | -7.82  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.507 | 0.479  |            | -5.52  | 20    |
| 1,1-Dichloroethene             | 0.479 | 0.448  |            | -6.47  | 20    |
| Acetone                        | 0.195 | 0.184  |            | -5.64  | 20    |
| Carbon Disulfide               | 1.192 | 1.019  |            | -14.51 | 20    |
| Methyl tert-butyl Ether        | 1.559 | 1.530  |            | -1.86  | 20    |
| Methyl Acetate                 | 0.608 | 0.535  |            | -12.01 | 20    |
| Methylene Chloride             | 0.645 | 0.515  |            | -20.16 | 20    |
| trans-1,2-Dichloroethene       | 0.519 | 0.471  |            | -9.25  | 20    |
| 1,1-Dichloroethane             | 0.877 | 0.813  | 0.1        | -7.3   | 20    |
| Cyclohexane                    | 0.860 | 0.733  |            | -14.77 | 20    |
| 2-Butanone                     | 0.274 | 0.255  |            | -6.93  | 20    |
| Carbon Tetrachloride           | 0.496 | 0.476  |            | -4.03  | 20    |
| cis-1,2-Dichloroethene         | 0.605 | 0.578  |            | -4.46  | 20    |
| Bromochloromethane             | 0.356 | 0.353  |            | -0.84  | 20    |
| Chloroform                     | 0.971 | 0.919  |            | -5.36  | 20    |
| 1,1,1-Trichloroethane          | 0.886 | 0.851  |            | -3.95  | 20    |
| Methylcyclohexane              | 0.543 | 0.530  |            | -2.39  | 20    |
| Benzene                        | 1.329 | 1.262  |            | -5.04  | 20    |
| 1,2-Dichloroethane             | 0.455 | 0.425  |            | -6.59  | 20    |
| Trichloroethene                | 0.376 | 0.353  |            | -6.12  | 20    |
| 1,2-Dichloropropane            | 0.318 | 0.300  |            | -5.66  | 20    |
| Bromodichloromethane           | 0.466 | 0.454  |            | -2.58  | 20    |
| 4-Methyl-2-Pentanone           | 0.340 | 0.325  |            | -4.41  | 20    |
| Toluene                        | 0.889 | 0.843  |            | -5.17  | 20    |
| t-1,3-Dichloropropene          | 0.444 | 0.470  |            | 5.86   | 20    |
| cis-1,3-Dichloropropene        | 0.511 | 0.511  |            | 0      | 20    |
| 1,1,2-Trichloroethane          | 0.340 | 0.322  |            | -5.29  | 20    |
| 2-Hexanone                     | 0.247 | 0.242  |            | -2.02  | 20    |
| Dibromochloromethane           | 0.358 | 0.371  |            | 3.63   | 20    |
| 1,2-Dibromoethane              | 0.339 | 0.338  |            | -0.29  | 20    |
| Tetrachloroethene              | 0.451 | 0.415  |            | -7.98  | 20    |
| Chlorobenzene                  | 1.057 | 1.011  | 0.3        | -4.35  | 20    |
| Ethyl Benzene                  | 1.809 | 1.805  |            | -0.22  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233  
 Instrument ID: MSVOA\_N Calibration Date/Time: 01/20/2023 10:27  
 Lab File ID: VN076337.D Init. Calib. Date(s): 01/16/2023 01/16/2023  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 10:54 13:17  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| m/p-Xylenes                 | 0.725 | 0.724  |            | -0.14  | 20    |
| o-Xylene                    | 0.712 | 0.712  |            | 0      | 20    |
| Styrene                     | 1.137 | 1.175  |            | 3.34   | 20    |
| Bromoform                   | 0.284 | 0.293  | 0.1        | 3.17   | 20    |
| Isopropylbenzene            | 3.868 | 3.769  |            | -2.56  | 20    |
| 1,1,2,2-Tetrachloroethane   | 1.112 | 0.997  | 0.3        | -10.34 | 20    |
| 1,3-Dichlorobenzene         | 1.823 | 1.710  |            | -6.2   | 20    |
| 1,4-Dichlorobenzene         | 1.842 | 1.691  |            | -8.2   | 20    |
| 1,2-Dichlorobenzene         | 1.816 | 1.687  |            | -7.1   | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.193 | 0.185  |            | -4.14  | 20    |
| 1,2,4-Trichlorobenzene      | 0.870 | 0.848  |            | -2.53  | 20    |
| 1,2,3-Trichlorobenzene      | 0.823 | 0.806  |            | -2.07  | 20    |
| 1,2-Dichloroethane-d4       | 0.560 | 0.568  |            | 1.43   | 20    |
| Dibromofluoromethane        | 0.313 | 0.326  |            | 4.15   | 20    |
| Toluene-d8                  | 1.155 | 1.207  |            | 4.5    | 20    |
| 4-Bromofluorobenzene        | 0.377 | 0.435  |            | 15.39  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG No.: 01233  
 Instrument ID: MSVOA\_N Calibration Date/Time: 01/23/2023 10:17  
 Lab File ID: VN076355.D Init. Calib. Date(s): 01/16/2023 01/16/2023  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 10:54 13:17  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Dichlorodifluoromethane        | 0.512 | 0.458  |            | -10.55 | 20    |
| Chloromethane                  | 0.531 | 0.476  | 0.1        | -10.36 | 20    |
| Vinyl Chloride                 | 0.650 | 0.625  |            | -3.85  | 20    |
| Bromomethane                   | 0.533 | 0.520  |            | -2.44  | 20    |
| Chloroethane                   | 0.488 | 0.479  |            | -1.84  | 20    |
| Trichlorofluoromethane         | 0.895 | 0.812  |            | -9.27  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.507 | 0.467  |            | -7.89  | 20    |
| 1,1-Dichloroethene             | 0.479 | 0.428  |            | -10.65 | 20    |
| Acetone                        | 0.195 | 0.177  |            | -9.23  | 20    |
| Carbon Disulfide               | 1.192 | 1.088  |            | -8.73  | 20    |
| Methyl tert-butyl Ether        | 1.559 | 1.510  |            | -3.14  | 20    |
| Methyl Acetate                 | 0.608 | 0.533  |            | -12.34 | 20    |
| Methylene Chloride             | 0.645 | 0.523  |            | -18.92 | 20    |
| trans-1,2-Dichloroethene       | 0.519 | 0.463  |            | -10.79 | 20    |
| 1,1-Dichloroethane             | 0.877 | 0.828  | 0.1        | -5.59  | 20    |
| Cyclohexane                    | 0.860 | 0.717  |            | -16.63 | 20    |
| 2-Butanone                     | 0.274 | 0.251  |            | -8.39  | 20    |
| Carbon Tetrachloride           | 0.496 | 0.489  |            | -1.41  | 20    |
| cis-1,2-Dichloroethene         | 0.605 | 0.565  |            | -6.61  | 20    |
| Bromochloromethane             | 0.356 | 0.387  |            | 8.71   | 20    |
| Chloroform                     | 0.971 | 0.932  |            | -4.02  | 20    |
| 1,1,1-Trichloroethane          | 0.886 | 0.858  |            | -3.16  | 20    |
| Methylcyclohexane              | 0.543 | 0.526  |            | -3.13  | 20    |
| Benzene                        | 1.329 | 1.247  |            | -6.17  | 20    |
| 1,2-Dichloroethane             | 0.455 | 0.430  |            | -5.49  | 20    |
| Trichloroethene                | 0.376 | 0.335  |            | -10.9  | 20    |
| 1,2-Dichloropropane            | 0.318 | 0.300  |            | -5.66  | 20    |
| Bromodichloromethane           | 0.466 | 0.475  |            | 1.93   | 20    |
| 4-Methyl-2-Pentanone           | 0.340 | 0.330  |            | -2.94  | 20    |
| Toluene                        | 0.889 | 0.825  |            | -7.2   | 20    |
| t-1,3-Dichloropropene          | 0.444 | 0.477  |            | 7.43   | 20    |
| cis-1,3-Dichloropropene        | 0.511 | 0.524  |            | 2.54   | 20    |
| 1,1,2-Trichloroethane          | 0.340 | 0.319  |            | -6.18  | 20    |
| 2-Hexanone                     | 0.247 | 0.234  |            | -5.26  | 20    |
| Dibromochloromethane           | 0.358 | 0.384  |            | 7.26   | 20    |
| 1,2-Dibromoethane              | 0.339 | 0.327  |            | -3.54  | 20    |
| Tetrachloroethene              | 0.451 | 0.385  |            | -14.63 | 20    |
| Chlorobenzene                  | 1.057 | 0.969  | 0.3        | -8.32  | 20    |
| Ethyl Benzene                  | 1.809 | 1.778  |            | -1.71  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233  
 Instrument ID: MSVOA\_N Calibration Date/Time: 01/23/2023 10:17  
 Lab File ID: VN076355.D Init. Calib. Date(s): 01/16/2023 01/16/2023  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 10:54 13:17  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| m/p-Xylenes                 | 0.725 | 0.699  |            | -3.59  | 20    |
| o-Xylene                    | 0.712 | 0.691  |            | -2.95  | 20    |
| Styrene                     | 1.137 | 1.135  |            | -0.18  | 20    |
| Bromoform                   | 0.284 | 0.309  | 0.1        | 8.8    | 20    |
| Isopropylbenzene            | 3.868 | 3.758  |            | -2.84  | 20    |
| 1,1,2,2-Tetrachloroethane   | 1.112 | 1.010  | 0.3        | -9.17  | 20    |
| 1,3-Dichlorobenzene         | 1.823 | 1.657  |            | -9.11  | 20    |
| 1,4-Dichlorobenzene         | 1.842 | 1.609  |            | -12.65 | 20    |
| 1,2-Dichlorobenzene         | 1.816 | 1.613  |            | -11.18 | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.193 | 0.183  |            | -5.18  | 20    |
| 1,2,4-Trichlorobenzene      | 0.870 | 0.821  |            | -5.63  | 20    |
| 1,2,3-Trichlorobenzene      | 0.823 | 0.789  |            | -4.13  | 20    |
| 1,2-Dichloroethane-d4       | 0.560 | 0.565  |            | 0.89   | 20    |
| Dibromofluoromethane        | 0.313 | 0.314  |            | 0.32   | 20    |
| Toluene-d8                  | 1.155 | 1.173  |            | 1.56   | 20    |
| 4-Bromofluorobenzene        | 0.377 | 0.403  |            | 6.9    | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: louie01  
 Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG No.: 01233  
 Instrument ID: MSVOA\_Y Calibration Date/Time: 01/23/2023 12:09  
 Lab File ID: VY012307.D Init. Calib. Date(s): 01/16/2023 01/16/2023  
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:17 14:34  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Dichlorodifluoromethane        | 0.437 | 0.387  |            | -11.44 | 20    |
| Chloromethane                  | 0.403 | 0.377  | 0.1        | -6.45  | 20    |
| Vinyl Chloride                 | 0.471 | 0.462  |            | -1.91  | 20    |
| Bromomethane                   | 0.364 | 0.354  |            | -2.75  | 20    |
| Chloroethane                   | 0.307 | 0.305  |            | -0.65  | 20    |
| Trichlorofluoromethane         | 0.919 | 0.884  |            | -3.81  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.447 | 0.440  |            | -1.57  | 20    |
| 1,1-Dichloroethene             | 0.459 | 0.454  |            | -1.09  | 20    |
| Acetone                        | 0.139 | 0.146  |            | 5.04   | 20    |
| Carbon Disulfide               | 1.420 | 1.365  |            | -3.87  | 20    |
| Methyl tert-butyl Ether        | 1.280 | 1.242  |            | -2.97  | 20    |
| Methyl Acetate                 | 0.312 | 0.292  |            | -6.41  | 20    |
| Methylene Chloride             | 0.762 | 0.518  |            | -32.02 | 20    |
| trans-1,2-Dichloroethene       | 0.512 | 0.505  |            | -1.37  | 20    |
| 1,1-Dichloroethane             | 0.785 | 0.778  | 0.1        | -0.89  | 20    |
| Cyclohexane                    | 0.729 | 0.682  |            | -6.45  | 20    |
| 2-Butanone                     | 0.154 | 0.146  |            | -5.2   | 20    |
| Carbon Tetrachloride           | 0.581 | 0.569  |            | -2.07  | 20    |
| cis-1,2-Dichloroethene         | 0.563 | 0.549  |            | -2.66  | 20    |
| Bromochloromethane             | 0.171 | 0.159  |            | -7.02  | 20    |
| Chloroform                     | 0.927 | 0.919  |            | -0.86  | 20    |
| 1,1,1-Trichloroethane          | 0.918 | 0.900  |            | -1.96  | 20    |
| Methylcyclohexane              | 0.550 | 0.560  |            | 1.82   | 20    |
| Benzene                        | 1.299 | 1.299  |            | 0      | 20    |
| 1,2-Dichloroethane             | 0.413 | 0.406  |            | -1.7   | 20    |
| Trichloroethene                | 0.388 | 0.383  |            | -1.29  | 20    |
| 1,2-Dichloropropane            | 0.271 | 0.274  |            | 1.11   | 20    |
| Bromodichloromethane           | 0.457 | 0.456  |            | -0.22  | 20    |
| 4-Methyl-2-Pentanone           | 0.197 | 0.186  |            | -5.58  | 20    |
| Toluene                        | 0.865 | 0.875  |            | 1.16   | 20    |
| t-1,3-Dichloropropene          | 0.478 | 0.466  |            | -2.51  | 20    |
| cis-1,3-Dichloropropene        | 0.514 | 0.511  |            | -0.58  | 20    |
| 1,1,2-Trichloroethane          | 0.258 | 0.247  |            | -4.26  | 20    |
| 2-Hexanone                     | 0.150 | 0.147  |            | -2     | 20    |
| Dibromochloromethane           | 0.342 | 0.329  |            | -3.8   | 20    |
| 1,2-Dibromoethane              | 0.249 | 0.239  |            | -4.02  | 20    |
| Tetrachloroethene              | 0.456 | 0.461  |            | 1.1    | 20    |
| Chlorobenzene                  | 1.023 | 1.005  | 0.3        | -1.76  | 20    |
| Ethyl Benzene                  | 1.828 | 1.852  |            | 1.31   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: loui01  
 Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG No.: O1233  
 Instrument ID: MSVOA\_Y Calibration Date/Time: 01/23/2023 12:09  
 Lab File ID: VY012307.D Init. Calib. Date(s): 01/16/2023 01/16/2023  
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:17 14:34  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------------|-------|--------|------------|-------|-------|
| m/p-Xylenes                 | 0.723 | 0.728  |            | 0.69  | 20    |
| o-Xylene                    | 0.690 | 0.693  |            | 0.44  | 20    |
| Styrene                     | 1.154 | 1.171  |            | 1.47  | 20    |
| Bromoform                   | 0.247 | 0.234  | 0.1        | -5.26 | 20    |
| Isopropylbenzene            | 3.605 | 3.679  |            | 2.05  | 20    |
| 1,1,2,2-Tetrachloroethane   | 0.568 | 0.540  | 0.3        | -4.93 | 20    |
| 1,3-Dichlorobenzene         | 1.726 | 1.693  |            | -1.91 | 20    |
| 1,4-Dichlorobenzene         | 1.721 | 1.678  |            | -2.5  | 20    |
| 1,2-Dichlorobenzene         | 1.540 | 1.496  |            | -2.86 | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.117 | 0.110  |            | -5.98 | 20    |
| 1,2,4-Trichlorobenzene      | 0.987 | 0.948  |            | -3.95 | 20    |
| 1,2,3-Trichlorobenzene      | 0.849 | 0.816  |            | -3.89 | 20    |
| 1,2-Dichloroethane-d4       | 0.465 | 0.425  |            | -8.6  | 20    |
| Dibromofluoromethane        | 0.265 | 0.258  |            | -2.64 | 20    |
| Toluene-d8                  | 1.044 | 0.958  |            | -8.24 | 20    |
| 4-Bromofluorobenzene        | 0.373 | 0.357  |            | -4.29 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.