



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VD075146.D Date Analyzed: 01/20/2023
Instrument ID: MSVOA_D Time Analyzed: 11:13
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	87549	7.88	150535	8.78	130836	11.59
UPPER LIMIT	175098	8.375	301070	9.281	261672	12.087
LOWER LIMIT	43774.5	7.375	75267.5	8.281	65418	11.087
EPA SAMPLE NO.						
B-P36B	4925 *	7.88	11060 *	8.78	10688 *	11.59
B-P38A	3464 *	7.88	6979 *	8.78	6961 *	11.58
B-P38B	41902 *	7.88	88407	8.78	88309	11.59
VD0120SBL04	86545	7.88	180550	8.78	168873	11.58
VD0120SBS02	78332	7.88	142964	8.78	128592	11.59

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.