



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: louie01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VN076337.D Date Analyzed: 01/20/2023
Instrument ID: MSVOA_N Time Analyzed: 10:27
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	383049	8.23	610968	9.11	549444	11.87
UPPER LIMIT	766098	8.73	1221940	9.606	1098890	12.365
LOWER LIMIT	191525	7.73	305484	8.606	274722	11.365
EPA SAMPLE NO.						
FB01	370105	8.23	595933	9.11	523917	11.87
VN0120WBL01	490775	8.23	773684	9.11	657515	11.87
VN0120WBS01	434458	8.23	686884	9.11	590812	11.87
VN0120WBSD01	359856	8.23	567895	9.11	510094	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.