



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233
Lab File ID: VN076355.D Date Analyzed: 01/23/2023
Instrument ID: MSVOA_N Time Analyzed: 10:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	360494	8.23	576114	9.11	518968	11.87
UPPER LIMIT	720988	8.73	1152230	9.606	1037940	12.37
LOWER LIMIT	180247	7.73	288057	8.606	259484	11.37
EPA SAMPLE NO.						
B-P38C	380260	8.22	594043	9.10	544762	11.87
B-P38CDL	384506	8.23	612820	9.11	549267	11.87
VN0123MBL01	303559	8.23	491398	9.11	403787	11.87
VN0123MBS01	381993	8.23	609908	9.11	527333	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.