



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VY012307.D Date Analyzed: 01/23/2023
Instrument ID: MSVOA_Y Time Analyzed: 12:09
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	179464	7.79	268117	8.69	243165	11.49
UPPER LIMIT	358928	8.289	536234	9.191	486330	11.99
LOWER LIMIT	89732	7.289	134059	8.191	121583	10.99
EPA SAMPLE NO.						
B-P37B	35937 *	7.79	57297 *	8.69	55952 *	11.49
B-P38ARE	4898 *	7.80	8005 *	8.69	7433 *	11.50
VY0123SBL01	162197	7.79	255207	8.69	227538	11.49
VY0123SBS01	173816	7.79	263834	8.69	239765	11.49

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.