



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VD075146.D Date Analyzed: 01/20/2023
Instrument ID: MSVOA_D Time Analyzed: 11:13
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	59366	13.522				
UPPER LIMIT	118732	14.022				
LOWER LIMIT	29683	13.022				
EPA SAMPLE NO.						
B-P36B	5310 *	13.52				
B-P38A	2980 *	13.52				
B-P38B	40736	13.52				
VD0120SBL04	72499	13.52				
VD0120SBS02	58018	13.52				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.