



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VD075168.D Date Analyzed: 01/23/2023
Instrument ID: MSVOA_D Time Analyzed: 10:15
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	56972	13.522				
UPPER LIMIT	113944	14.022				
LOWER LIMIT	28486	13.022				
EPA SAMPLE NO.						
B-P34A	54879	13.52				
B-P34B	41757	13.52				
B-P36A	64243	13.52				
B-P36BRE	1552 *	13.52				
B-P37A	63432	13.52				
B-P38BRE	699 *	13.53				
VD0123SBL01	71751	13.52				
VD0123SBS01	55819	13.52				
VD0123SBSD01	53241	13.52				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.