



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: louie01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VN076337.D Date Analyzed: 01/20/2023
Instrument ID: MSVOA_N Time Analyzed: 10:27
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	272155	13.794				
UPPER LIMIT	544310	14.294				
LOWER LIMIT	136078	13.294				
EPA SAMPLE NO.						
FB01	212240	13.79				
VN0120WBL01	251353	13.79				
VN0120WBS01	268276	13.79				
VN0120WBSD01	238179	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.