



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VN076355.D Date Analyzed: 01/23/2023
Instrument ID: MSVOA_N Time Analyzed: 10:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	248688	13.794				
UPPER LIMIT	497376	14.294				
LOWER LIMIT	124344	13.294				
EPA SAMPLE NO.						
B-P38C	318779	13.79				
B-P38CDL	255428	13.79				
VN0123MBL01	158993	13.79				
VN0123MBS01	238215	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.