



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: O1233 SAS No.: O1233 SDG NO.: O1233
Lab File ID: VY012307.D Date Analyzed: 01/23/2023
Instrument ID: MSVOA_Y Time Analyzed: 12:09
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	123436	13.428				
UPPER LIMIT	246872	13.928				
LOWER LIMIT	61718	12.928				
EPA SAMPLE NO.						
B-P37B	24351 *	13.43				
B-P38ARE	2570 *	13.43				
VY0123SBL01	110055	13.43				
VY0123SBS01	122385	13.43				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.