



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233
Lab File ID: VY012193.D BFB Injection Date: 01/16/2023
Instrument ID: MSVOA_Y BFB Injection Time: 09:00
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.8
75	30.0 - 60.0% of mass 95	54.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1 (1.1) 1
174	50.0 - 100.0% of mass 95	88.4
175	5.0 - 9.0% of mass 174	6.7 (7.6) 1
176	95.0 - 101.0% of mass 174	85.2 (96.4) 1
177	5.0 - 9.0% of mass 176	5.9 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY012194.D	01/16/2023	12:17
VSTDIC010	VSTDIC010	VY012195.D	01/16/2023	13:03
VSTDIC020	VSTDIC020	VY012196.D	01/16/2023	13:26
VSTDIC050	VSTDIC050	VY012197.D	01/16/2023	13:48
VSTDIC100	VSTDIC100	VY012198.D	01/16/2023	14:11
VSTDIC150	VSTDIC150	VY012199.D	01/16/2023	14:34