



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: loui01
Lab Code: CHEM Case No.: 01233 SAS No.: 01233 SDG NO.: 01233
Lab File ID: VY012306.D BFB Injection Date: 01/23/2023
Instrument ID: MSVOA_Y BFB Injection Time: 10:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.7
75	30.0 - 60.0% of mass 95	54.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1 (1.1) 1
174	50.0 - 100.0% of mass 95	90.6
175	5.0 - 9.0% of mass 174	6.3 (7) 1
176	95.0 - 101.0% of mass 174	87.1 (96.1) 1
177	5.0 - 9.0% of mass 176	5.8 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY012307.D	01/23/2023	12:09
VY0123SBL01	VY0123SBL01	VY012308.D	01/23/2023	13:04
VY0123SBS01	VY0123SBS01	VY012309.D	01/23/2023	13:27
B-P37B	O1233-07	VY012315.D	01/23/2023	15:48
B-P38ARE	O1233-08RE	VY012316.D	01/23/2023	16:11