



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG No.: O2213

Calibration Sequence : FE040523		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	138705834	138706	FE041180.D
500	71225054	142450	FE041181.D
200	28701045	143505	FE041182.D
100	15839209	158392	FE041183.D
50	7706171	154123	FE041184.D
AVG RF : 147435		% RSD : 5.688	AVG RT : 14.942