



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG No.: O2213

Calibration Sequence : FF031623		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	128171793	128172	FF012284.D
500	61098639	122197	FF012285.D
200	24884867	124424	FF012286.D
100	13267007	132670	FF012287.D
50	6297439	125949	FF012288.D
AVG RF : 126682		% RSD : 3.153	AVG RT : 15.0264