



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02  
ProjectID: LOD-LOQ Study  
Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG No.: O2213

| Calibration Sequence : FG031523 |            | Test : Diesel Range Organics |                  |
|---------------------------------|------------|------------------------------|------------------|
| Concentration (PPM)             | Area Count | Reference Factor             | File ID          |
| 1000                            | 109865168  | 109865                       | FG011595.D       |
| 500                             | 56627452   | 113255                       | FG011596.D       |
| 200                             | 22275081   | 111375                       | FG011597.D       |
| 100                             | 12582389   | 125824                       | FG011598.D       |
| 50                              | 6222261    | 124445                       | FG011599.D       |
| AVG RF : 116953                 |            | % RSD : 6.482                | AVG RT : 15.0538 |