



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/13/2023 Initial Calibration Date(s): 04/12/2023 04/12/2023

Continuing Calib Time: 01:18 Initial Calibration Time(s): 12:07 13:44

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	6.63	6.63	6.53	6.73	0.00
2,4-DCAA	6.48	6.48	6.38	6.58	0.00
DICHLORPROP	7.25	7.25	7.15	7.35	0.01
2,4-D	7.45	7.45	7.35	7.55	0.00
2,4,5-TP(Silvex)	8.22	8.22	8.12	8.32	0.00
2,4,5-T	8.48	8.48	8.38	8.58	0.00
2,4-DB	8.99	8.99	8.89	9.09	0.00
Dinoseb	10.07	10.07	9.97	10.17	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/13/2023 Initial Calibration Date(s): 04/12/2023 04/12/2023

Continuing Calib Time: 01:18 Initial Calibration Time(s): 12:07 13:44

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.04	7.04	6.94	7.14	0.00
2,4-DCAA	6.87	6.87	6.77	6.97	0.00
DICHLORPROP	7.66	7.66	7.56	7.76	0.00
2,4-D	7.95	7.95	7.85	8.05	0.00
2,4,5-TP(Silvex)	8.74	8.74	8.64	8.84	0.00
2,4,5-T	9.12	9.12	9.02	9.22	0.00
2,4-DB	9.63	9.63	9.53	9.73	0.00
Dinoseb	9.96	9.96	9.86	10.06	0.00