



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/12/2023 04/12/2023

Client Sample No.: CCAL01 Date Analyzed: 04/12/2023

Lab Sample No.: HSTDCCC750 Data File : PS022669.D Time Analyzed: 14:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.477	8.376	8.576	721.070	712.500	1.2
2,4,5-TP(Silvex)	8.217	8.117	8.317	720.070	712.500	1.1
2,4-D	7.446	7.345	7.545	715.470	705.000	1.5
2,4-DB	8.991	8.891	9.091	722.590	712.500	1.4
2,4-DCAA	6.484	6.384	6.584	763.390	750.000	1.8
DICAMBA	6.634	6.533	6.733	729.750	705.000	3.5
DICHLORPROP	7.245	7.145	7.345	714.990	705.000	1.4
Dinoseb	10.065	9.966	10.166	718.770	705.000	2.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/12/2023 04/12/2023

Client Sample No.: CCAL01 Date Analyzed: 04/12/2023

Lab Sample No.: HSTDCCC750 Data File : PS022669.D Time Analyzed: 14:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.122	9.019	9.219	747.330	712.500	4.9
2,4,5-TP(Silvex)	8.746	8.644	8.844	746.490	712.500	4.8
2,4-D	7.952	7.850	8.050	736.700	705.000	4.5
2,4-DB	9.630	9.527	9.727	748.640	712.500	5.1
2,4-DCAA	6.872	6.771	6.971	776.190	750.000	3.5
DICAMBA	7.040	6.939	7.139	742.110	705.000	5.3
DICHLORPROP	7.663	7.561	7.761	731.540	705.000	3.8
Dinoseb	9.960	9.859	10.059	739.440	705.000	4.9