



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/12/2023 04/12/2023

Client Sample No.: CCAL02 Date Analyzed: 04/12/2023

Lab Sample No.: HSTDCCC750 Data File : PS022681.D Time Analyzed: 20:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.476	8.376	8.576	730.710	712.500	2.6
2,4,5-TP(Silvex)	8.216	8.117	8.317	735.170	712.500	3.2
2,4-D	7.446	7.345	7.545	722.080	705.000	2.4
2,4-DB	8.990	8.891	9.091	720.820	712.500	1.2
2,4-DCAA	6.484	6.384	6.584	770.220	750.000	2.7
DICAMBA	6.634	6.533	6.733	736.910	705.000	4.5
DICHLORPROP	7.245	7.145	7.345	724.160	705.000	2.7
Dinoseb	10.065	9.966	10.166	720.880	705.000	2.3



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/12/2023 04/12/2023

Client Sample No.: CCAL02 Date Analyzed: 04/12/2023

Lab Sample No.: HSTDCCC750 Data File : PS022681.D Time Analyzed: 20:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.121	9.019	9.219	754.820	712.500	5.9
2,4,5-TP(Silvex)	8.745	8.644	8.844	752.920	712.500	5.7
2,4-D	7.951	7.850	8.050	743.840	705.000	5.5
2,4-DB	9.629	9.527	9.727	758.500	712.500	6.5
2,4-DCAA	6.871	6.771	6.971	783.200	750.000	4.4
DICAMBA	7.040	6.939	7.139	749.630	705.000	6.3
DICHLORPROP	7.662	7.561	7.761	739.580	705.000	4.9
Dinoseb	9.959	9.859	10.059	746.170	705.000	5.8