



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/03/2023 04/04/2023

Continuing Calib Time: 08:28 Initial Calibration Time(s): 18:36 02:18

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1016-2	(2)	5.62	5.62	5.52	5.72	0.00
Aroclor-1016-3	(3)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-4	(4)	5.78	5.78	5.68	5.88	0.00
Aroclor-1016-5	(5)	6.08	6.08	5.98	6.18	0.00
Aroclor-1260-1	(1)	7.21	7.22	7.12	7.32	0.01
Aroclor-1260-2	(2)	7.47	7.48	7.38	7.58	0.01
Aroclor-1260-3	(3)	7.84	7.84	7.74	7.94	0.00
Aroclor-1260-4	(4)	8.06	8.07	7.97	8.17	0.01
Aroclor-1260-5	(5)	8.39	8.39	8.29	8.49	0.00
Tetrachloro-m-xylene		4.42	4.42	4.32	4.52	0.00
Decachlorobiphenyl		10.24	10.25	10.15	10.35	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/03/2023 04/04/2023

Continuing Calib Time: 08:28 Initial Calibration Time(s): 18:36 02:18

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.71	4.71	4.61	4.81	0.00
Aroclor-1016-2	(2)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-3	(3)	4.90	4.91	4.81	5.01	0.01
Aroclor-1016-4	(4)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-5	(5)	5.16	5.16	5.06	5.26	0.00
Aroclor-1260-1	(1)	6.19	6.20	6.10	6.30	0.01
Aroclor-1260-2	(2)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-3	(3)	6.54	6.54	6.44	6.64	0.01
Aroclor-1260-4	(4)	7.01	7.01	6.91	7.11	0.00
Aroclor-1260-5	(5)	7.25	7.26	7.16	7.36	0.01
Tetrachloro-m-xylene		3.62	3.63	3.53	3.73	0.01
Decachlorobiphenyl		8.64	8.65	8.55	8.75	0.01