



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/06/2023 04/07/2023

Continuing Calib Time: 11:38 Initial Calibration Time(s): 17:37 01:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.61	5.61	5.51	5.71	0.00
Aroclor-1248-2 (2)	5.88	5.89	5.79	5.99	0.01
Aroclor-1248-3 (3)	6.09	6.09	5.99	6.19	0.00
Aroclor-1248-4 (4)	6.50	6.50	6.40	6.60	0.01
Aroclor-1248-5 (5)	6.53	6.54	6.44	6.64	0.01
Tetrachloro-m-xylene	4.43	4.44	4.34	4.54	0.01
Decachlorobiphenyl	10.27	10.27	10.17	10.37	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/06/2023 04/07/2023

Continuing Calib Time: 11:38 Initial Calibration Time(s): 17:37 01:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1248-2 (2)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-3 (3)	5.04	5.05	4.95	5.15	0.01
Aroclor-1248-4 (4)	5.22	5.22	5.12	5.32	0.00
Aroclor-1248-5 (5)	5.62	5.62	5.52	5.72	0.01
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.75	8.75	8.65	8.85	0.01