



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/06/2023 04/07/2023

Continuing Calib Time: 14:39 Initial Calibration Time(s): 17:37 01:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.61	5.61	5.51	5.71	0.00
Aroclor-1016-2	(2)	5.63	5.64	5.54	5.74	0.01
Aroclor-1016-3	(3)	5.69	5.70	5.60	5.80	0.01
Aroclor-1016-4	(4)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-5	(5)	6.09	6.09	5.99	6.19	0.00
Aroclor-1260-1	(1)	7.22	7.23	7.13	7.33	0.01
Aroclor-1260-2	(2)	7.48	7.49	7.39	7.59	0.01
Aroclor-1260-3	(3)	7.84	7.85	7.75	7.95	0.01
Aroclor-1260-4	(4)	8.07	8.08	7.98	8.18	0.01
Aroclor-1260-5	(5)	8.40	8.40	8.30	8.50	0.00
Tetrachloro-m-xylene		4.43	4.44	4.34	4.54	0.01
Decachlorobiphenyl		10.27	10.28	10.18	10.38	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/06/2023 04/07/2023

Continuing Calib Time: 14:39 Initial Calibration Time(s): 17:37 01:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.00	5.01	4.91	5.11	0.01
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.26	6.27	6.17	6.37	0.01
Aroclor-1260-2	(2)	6.45	6.46	6.36	6.56	0.01
Aroclor-1260-3	(3)	6.61	6.61	6.51	6.71	0.00
Aroclor-1260-4	(4)	7.09	7.09	6.99	7.19	0.00
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.66	3.67	3.57	3.77	0.01
Decachlorobiphenyl		8.75	8.75	8.65	8.85	0.00