



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/06/2023 04/07/2023

Continuing Calib Time: 20:25 Initial Calibration Time(s): 17:37 01:03

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1016-1       | (1) | 5.61       | 5.61      | 5.51                   | 5.71  | 0.00       |
| Aroclor-1016-2       | (2) | 5.63       | 5.64      | 5.54                   | 5.74  | 0.01       |
| Aroclor-1016-3       | (3) | 5.69       | 5.70      | 5.60                   | 5.80  | 0.01       |
| Aroclor-1016-4       | (4) | 5.79       | 5.80      | 5.70                   | 5.90  | 0.01       |
| Aroclor-1016-5       | (5) | 6.09       | 6.09      | 5.99                   | 6.19  | 0.00       |
| Aroclor-1260-1       | (1) | 7.22       | 7.23      | 7.13                   | 7.33  | 0.01       |
| Aroclor-1260-2       | (2) | 7.48       | 7.49      | 7.39                   | 7.59  | 0.01       |
| Aroclor-1260-3       | (3) | 7.84       | 7.85      | 7.75                   | 7.95  | 0.01       |
| Aroclor-1260-4       | (4) | 8.07       | 8.08      | 7.98                   | 8.18  | 0.01       |
| Aroclor-1260-5       | (5) | 8.40       | 8.40      | 8.30                   | 8.50  | 0.00       |
| Tetrachloro-m-xylene |     | 4.43       | 4.44      | 4.34                   | 4.54  | 0.01       |
| Decachlorobiphenyl   |     | 10.27      | 10.28     | 10.18                  | 10.38 | 0.01       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/07/2023 Initial Calibration Date(s): 04/06/2023 04/07/2023

Continuing Calib Time: 20:25 Initial Calibration Time(s): 17:37 01:03

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.76       | 4.77      | 4.67      | 4.87 | 0.01       |
| Aroclor-1016-2       | (2) | 4.78       | 4.79      | 4.69      | 4.89 | 0.01       |
| Aroclor-1016-3       | (3) | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1016-4       | (4) | 5.00       | 5.01      | 4.91      | 5.11 | 0.01       |
| Aroclor-1016-5       | (5) | 5.22       | 5.22      | 5.12      | 5.32 | 0.00       |
| Aroclor-1260-1       | (1) | 6.26       | 6.27      | 6.17      | 6.37 | 0.01       |
| Aroclor-1260-2       | (2) | 6.45       | 6.46      | 6.36      | 6.56 | 0.01       |
| Aroclor-1260-3       | (3) | 6.61       | 6.61      | 6.51      | 6.71 | 0.00       |
| Aroclor-1260-4       | (4) | 7.08       | 7.09      | 6.99      | 7.19 | 0.01       |
| Aroclor-1260-5       | (5) | 7.33       | 7.33      | 7.23      | 7.43 | 0.00       |
| Tetrachloro-m-xylene |     | 3.66       | 3.67      | 3.57      | 3.77 | 0.01       |
| Decachlorobiphenyl   |     | 8.74       | 8.75      | 8.65      | 8.85 | 0.01       |