



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/10/2023 Initial Calibration Date(s): 03/27/2023 03/27/2023

Continuing Calib Time: 09:01 Initial Calibration Time(s): 10:20 16:54

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.51       | 4.51      | 4.41                   | 4.61 | 0.00       |
| Aroclor-1016-2       | (2) | 4.53       | 4.53      | 4.43                   | 4.63 | 0.00       |
| Aroclor-1016-3       | (3) | 4.59       | 4.58      | 4.48                   | 4.68 | -0.01      |
| Aroclor-1016-4       | (4) | 4.68       | 4.68      | 4.58                   | 4.78 | 0.00       |
| Aroclor-1016-5       | (5) | 4.97       | 4.96      | 4.86                   | 5.06 | -0.01      |
| Aroclor-1260-1       | (1) | 6.07       | 6.07      | 5.97                   | 6.17 | 0.00       |
| Aroclor-1260-2       | (2) | 6.33       | 6.33      | 6.23                   | 6.43 | 0.00       |
| Aroclor-1260-3       | (3) | 6.68       | 6.68      | 6.58                   | 6.78 | 0.00       |
| Aroclor-1260-4       | (4) | 6.90       | 6.90      | 6.80                   | 7.00 | 0.00       |
| Aroclor-1260-5       | (5) | 7.21       | 7.21      | 7.11                   | 7.31 | 0.00       |
| Tetrachloro-m-xylene |     | 3.41       | 3.41      | 3.31                   | 3.51 | 0.00       |
| Decachlorobiphenyl   |     | 8.59       | 8.59      | 8.49                   | 8.69 | 0.00       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/10/2023 Initial Calibration Date(s): 03/27/2023 03/27/2023

Continuing Calib Time: 09:01 Initial Calibration Time(s): 10:20 16:54

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 3.77       | 3.77      | 3.67      | 3.87 | 0.00       |
| Aroclor-1016-2       | (2) | 3.78       | 3.79      | 3.69      | 3.89 | 0.01       |
| Aroclor-1016-3       | (3) | 3.94       | 3.95      | 3.85      | 4.05 | 0.01       |
| Aroclor-1016-4       | (4) | 3.99       | 4.00      | 3.90      | 4.10 | 0.01       |
| Aroclor-1016-5       | (5) | 4.19       | 4.19      | 4.09      | 4.29 | 0.00       |
| Aroclor-1260-1       | (1) | 5.18       | 5.19      | 5.09      | 5.29 | 0.01       |
| Aroclor-1260-2       | (2) | 5.37       | 5.38      | 5.28      | 5.48 | 0.01       |
| Aroclor-1260-3       | (3) | 5.51       | 5.52      | 5.42      | 5.62 | 0.01       |
| Aroclor-1260-4       | (4) | 5.97       | 5.98      | 5.88      | 6.08 | 0.01       |
| Aroclor-1260-5       | (5) | 6.22       | 6.23      | 6.13      | 6.33 | 0.01       |
| Tetrachloro-m-xylene |     | 2.76       | 2.76      | 2.66      | 2.86 | 0.00       |
| Decachlorobiphenyl   |     | 7.54       | 7.55      | 7.45      | 7.65 | 0.01       |