



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/11/2023 Initial Calibration Date(s): 03/27/2023 03/27/2023

Continuing Calib Time: 10:17 Initial Calibration Time(s): 10:20 16:54

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.51	4.51	4.41	4.61	0.00
Aroclor-1016-2	(2)	4.52	4.53	4.43	4.63	0.01
Aroclor-1016-3	(3)	4.58	4.58	4.48	4.68	0.00
Aroclor-1016-4	(4)	4.67	4.68	4.58	4.78	0.01
Aroclor-1016-5	(5)	4.96	4.96	4.86	5.06	0.00
Aroclor-1260-1	(1)	6.06	6.07	5.97	6.17	0.01
Aroclor-1260-2	(2)	6.32	6.33	6.23	6.43	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	6.89	6.90	6.80	7.00	0.01
Aroclor-1260-5	(5)	7.21	7.21	7.11	7.31	0.01
Tetrachloro-m-xylene		3.41	3.41	3.31	3.51	0.01
Decachlorobiphenyl		8.59	8.59	8.49	8.69	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/11/2023 Initial Calibration Date(s): 03/27/2023 03/27/2023

Continuing Calib Time: 10:17 Initial Calibration Time(s): 10:20 16:54

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.77	3.77	3.67	3.87	0.00
Aroclor-1016-2	(2)	3.78	3.79	3.69	3.89	0.01
Aroclor-1016-3	(3)	3.94	3.95	3.85	4.05	0.01
Aroclor-1016-4	(4)	3.99	4.00	3.90	4.10	0.01
Aroclor-1016-5	(5)	4.19	4.19	4.09	4.29	0.00
Aroclor-1260-1	(1)	5.18	5.19	5.09	5.29	0.01
Aroclor-1260-2	(2)	5.37	5.38	5.28	5.48	0.01
Aroclor-1260-3	(3)	5.51	5.52	5.42	5.62	0.01
Aroclor-1260-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1260-5	(5)	6.22	6.23	6.13	6.33	0.01
Tetrachloro-m-xylene		2.76	2.76	2.66	2.86	0.00
Decachlorobiphenyl		7.54	7.55	7.45	7.65	0.01