



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/08/2023 Initial Calibration Date(s): 04/07/2023 04/07/2023

Continuing Calib Time: 05:30 Initial Calibration Time(s): 12:55 15:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-2	(2)	4.94	4.95	4.85	5.05	0.01
Aroclor-1016-3	(3)	5.01	5.01	4.91	5.11	0.00
Aroclor-1016-4	(4)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-5	(5)	5.43	5.43	5.33	5.53	0.00
Aroclor-1260-1	(1)	6.65	6.66	6.56	6.76	0.01
Aroclor-1260-2	(2)	6.94	6.94	6.84	7.04	0.01
Aroclor-1260-3	(3)	7.33	7.33	7.23	7.43	0.00
Aroclor-1260-4	(4)	7.57	7.58	7.48	7.68	0.01
Aroclor-1260-5	(5)	7.91	7.92	7.82	8.02	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		9.65	9.66	9.56	9.76	0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Continuing Calib Date: 04/08/2023 Initial Calibration Date(s): 04/07/2023 04/07/2023

Continuing Calib Time: 05:30 Initial Calibration Time(s): 12:55 15:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.01	4.01	3.91	4.11	0.00
Aroclor-1016-2 (2)	4.03	4.03	3.93	4.13	0.00
Aroclor-1016-3 (3)	4.20	4.20	4.10	4.30	0.00
Aroclor-1016-4 (4)	4.26	4.26	4.16	4.36	0.00
Aroclor-1016-5 (5)	4.47	4.47	4.37	4.57	0.00
Aroclor-1260-1 (1)	5.56	5.56	5.46	5.66	0.01
Aroclor-1260-2 (2)	5.76	5.76	5.66	5.86	0.00
Aroclor-1260-3 (3)	5.92	5.92	5.82	6.02	0.00
Aroclor-1260-4 (4)	6.42	6.42	6.32	6.52	0.00
Aroclor-1260-5 (5)	6.69	6.69	6.59	6.79	0.01
Tetrachloro-m-xylene	2.90	2.90	2.80	3.00	0.00
Decachlorobiphenyl	8.13	8.13	8.03	8.23	0.00