



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/03/2023 04/03/2023

Client Sample No.: CCAL01 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO093919.D Time Analyzed: 12:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.596	5.499	5.699	463.730	500.000	-7.3
Aroclor-1016-2	5.619	5.521	5.721	447.500	500.000	-10.5
Aroclor-1016-3	5.681	5.583	5.783	455.200	500.000	-9.0
Aroclor-1016-4	5.780	5.682	5.882	463.330	500.000	-7.3
Aroclor-1016-5	6.078	5.980	6.180	487.270	500.000	-2.5
Aroclor-1260-1	7.214	7.117	7.317	428.120	500.000	-14.4
Aroclor-1260-2	7.473	7.377	7.577	442.370	500.000	-11.5
Aroclor-1260-3	7.835	7.740	7.940	407.390	500.000	-18.5
Aroclor-1260-4	8.063	7.967	8.167	433.270	500.000	-13.3
Aroclor-1260-5	8.390	8.294	8.494	514.770	500.000	3.0
Decachlorobiphenyl	10.243	10.151	10.351	48.190	50.000	-3.6
Tetrachloro-m-xylene	4.420	4.322	4.522	41.140	50.000	-17.7



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/03/2023 04/03/2023

Client Sample No.: CCAL01 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO093919.D Time Analyzed: 12:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.708	4.610	4.810	500.500	500.000	0.1
Aroclor-1016-2	4.726	4.629	4.829	489.000	500.000	-2.2
Aroclor-1016-3	4.903	4.805	5.005	481.460	500.000	-3.7
Aroclor-1016-4	4.945	4.848	5.048	448.680	500.000	-10.3
Aroclor-1016-5	5.158	5.061	5.261	484.650	500.000	-3.1
Aroclor-1260-1	6.194	6.097	6.297	441.220	500.000	-11.8
Aroclor-1260-2	6.383	6.287	6.487	463.200	500.000	-7.4
Aroclor-1260-3	6.536	6.440	6.640	413.350	500.000	-17.3
Aroclor-1260-4	7.010	6.914	7.114	447.070	500.000	-10.6
Aroclor-1260-5	7.253	7.158	7.358	507.700	500.000	1.5
Decachlorobiphenyl	8.644	8.551	8.751	44.710	50.000	-10.6
Tetrachloro-m-xylene	3.625	3.527	3.727	44.820	50.000	-10.4