



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/03/2023 04/03/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO093930.D Time Analyzed: 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.595	5.499	5.699	484.320	500.000	-3.1
Aroclor-1016-2	5.618	5.521	5.721	467.640	500.000	-6.5
Aroclor-1016-3	5.680	5.583	5.783	470.000	500.000	-6.0
Aroclor-1016-4	5.779	5.682	5.882	476.720	500.000	-4.7
Aroclor-1016-5	6.076	5.980	6.180	456.400	500.000	-8.7
Aroclor-1260-1	7.213	7.117	7.317	451.330	500.000	-9.7
Aroclor-1260-2	7.472	7.377	7.577	460.720	500.000	-7.9
Aroclor-1260-3	7.835	7.740	7.940	435.920	500.000	-12.8
Aroclor-1260-4	8.062	7.967	8.167	461.230	500.000	-7.8
Aroclor-1260-5	8.389	8.294	8.494	524.900	500.000	5.0
Decachlorobiphenyl	10.242	10.151	10.351	50.620	50.000	1.2
Tetrachloro-m-xylene	4.420	4.322	4.522	43.200	50.000	-13.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/03/2023 04/03/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO093930.D Time Analyzed: 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.707	4.610	4.810	523.470	500.000	4.7
Aroclor-1016-2	4.725	4.629	4.829	521.110	500.000	4.2
Aroclor-1016-3	4.902	4.805	5.005	504.570	500.000	0.9
Aroclor-1016-4	4.944	4.848	5.048	479.810	500.000	-4.0
Aroclor-1016-5	5.157	5.061	5.261	515.010	500.000	3.0
Aroclor-1260-1	6.193	6.097	6.297	482.180	500.000	-3.6
Aroclor-1260-2	6.382	6.287	6.487	499.160	500.000	-0.2
Aroclor-1260-3	6.535	6.440	6.640	454.790	500.000	-9.0
Aroclor-1260-4	7.008	6.914	7.114	487.590	500.000	-2.5
Aroclor-1260-5	7.253	7.158	7.358	547.240	500.000	9.4
Decachlorobiphenyl	8.643	8.551	8.751	48.350	50.000	-3.3
Tetrachloro-m-xylene	3.624	3.527	3.727	47.930	50.000	-4.1