



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/03/2023 04/03/2023

Client Sample No.: CCAL03 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO093941.D Time Analyzed: 19:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.595	5.499	5.699	486.720	500.000	-2.7
Aroclor-1016-2	5.618	5.521	5.721	470.860	500.000	-5.8
Aroclor-1016-3	5.680	5.583	5.783	476.040	500.000	-4.8
Aroclor-1016-4	5.779	5.682	5.882	480.340	500.000	-3.9
Aroclor-1016-5	6.076	5.980	6.180	462.710	500.000	-7.5
Aroclor-1260-1	7.213	7.117	7.317	456.090	500.000	-8.8
Aroclor-1260-2	7.472	7.377	7.577	462.990	500.000	-7.4
Aroclor-1260-3	7.835	7.740	7.940	443.530	500.000	-11.3
Aroclor-1260-4	8.062	7.967	8.167	464.900	500.000	-7.0
Aroclor-1260-5	8.389	8.294	8.494	525.120	500.000	5.0
Decachlorobiphenyl	10.241	10.151	10.351	50.950	50.000	1.9
Tetrachloro-m-xylene	4.419	4.322	4.522	43.720	50.000	-12.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/03/2023 04/03/2023

Client Sample No.: CCAL03 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PO093941.D Time Analyzed: 19:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.707	4.610	4.810	531.630	500.000	6.3
Aroclor-1016-2	4.725	4.629	4.829	518.950	500.000	3.8
Aroclor-1016-3	4.901	4.805	5.005	513.480	500.000	2.7
Aroclor-1016-4	4.944	4.848	5.048	492.370	500.000	-1.5
Aroclor-1016-5	5.157	5.061	5.261	516.240	500.000	3.2
Aroclor-1260-1	6.192	6.097	6.297	486.560	500.000	-2.7
Aroclor-1260-2	6.381	6.287	6.487	506.680	500.000	1.3
Aroclor-1260-3	6.534	6.440	6.640	466.050	500.000	-6.8
Aroclor-1260-4	7.008	6.914	7.114	502.210	500.000	0.4
Aroclor-1260-5	7.252	7.158	7.358	558.520	500.000	11.7
Decachlorobiphenyl	8.643	8.551	8.751	49.190	50.000	-1.6
Tetrachloro-m-xylene	3.623	3.527	3.727	48.110	50.000	-3.8