



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/06/2023 04/06/2023

Client Sample No.: CCAL01 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PP056709.D Time Analyzed: 11:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.608	5.512	5.712	446.290	500.000	-10.7
Aroclor-1016-2	5.630	5.535	5.735	449.140	500.000	-10.2
Aroclor-1016-3	5.693	5.598	5.798	435.760	500.000	-12.8
Aroclor-1016-4	5.792	5.697	5.897	430.390	500.000	-13.9
Aroclor-1016-5	6.090	5.994	6.194	431.400	500.000	-13.7
Aroclor-1260-1	7.223	7.128	7.328	442.280	500.000	-11.5
Aroclor-1260-2	7.482	7.386	7.586	460.550	500.000	-7.9
Aroclor-1260-3	7.843	7.747	7.947	457.210	500.000	-8.6
Aroclor-1260-4	8.070	7.975	8.175	480.590	500.000	-3.9
Aroclor-1260-5	8.396	8.301	8.501	508.210	500.000	1.6
Decachlorobiphenyl	10.267	10.175	10.375	50.500	50.000	1.0
Tetrachloro-m-xylene	4.433	4.337	4.537	45.160	50.000	-9.7



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/06/2023 04/06/2023

Client Sample No.: CCAL01 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PP056709.D Time Analyzed: 11:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	468.380	500.000	-6.3
Aroclor-1016-2	4.781	4.685	4.885	469.950	500.000	-6.0
Aroclor-1016-3	4.960	4.864	5.064	475.590	500.000	-4.9
Aroclor-1016-4	5.002	4.906	5.106	438.710	500.000	-12.3
Aroclor-1016-5	5.218	5.123	5.323	484.990	500.000	-3.0
Aroclor-1260-1	6.264	6.169	6.369	430.850	500.000	-13.8
Aroclor-1260-2	6.453	6.358	6.558	440.610	500.000	-11.9
Aroclor-1260-3	6.609	6.514	6.714	426.360	500.000	-14.7
Aroclor-1260-4	7.085	6.990	7.190	417.560	500.000	-16.5
Aroclor-1260-5	7.327	7.232	7.432	434.880	500.000	-13.0
Decachlorobiphenyl	8.744	8.651	8.851	45.690	50.000	-8.6
Tetrachloro-m-xylene	3.660	3.565	3.765	47.340	50.000	-5.3