



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/06/2023 04/06/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PP056722.D Time Analyzed: 14:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.608	5.512	5.712	481.000	500.000	-3.8
Aroclor-1016-2	5.632	5.535	5.735	491.240	500.000	-1.8
Aroclor-1016-3	5.694	5.598	5.798	487.530	500.000	-2.5
Aroclor-1016-4	5.793	5.697	5.897	485.850	500.000	-2.8
Aroclor-1016-5	6.090	5.994	6.194	490.530	500.000	-1.9
Aroclor-1260-1	7.224	7.128	7.328	489.610	500.000	-2.1
Aroclor-1260-2	7.481	7.386	7.586	495.680	500.000	-0.9
Aroclor-1260-3	7.843	7.747	7.947	494.830	500.000	-1.0
Aroclor-1260-4	8.070	7.975	8.175	506.060	500.000	1.2
Aroclor-1260-5	8.397	8.301	8.501	510.610	500.000	2.1
Decachlorobiphenyl	10.268	10.175	10.375	51.630	50.000	3.3
Tetrachloro-m-xylene	4.434	4.337	4.537	48.640	50.000	-2.7



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/06/2023 04/06/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PP056722.D Time Analyzed: 14:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	478.270	500.000	-4.3
Aroclor-1016-2	4.782	4.685	4.885	483.520	500.000	-3.3
Aroclor-1016-3	4.960	4.864	5.064	491.190	500.000	-1.8
Aroclor-1016-4	5.002	4.906	5.106	457.060	500.000	-8.6
Aroclor-1016-5	5.218	5.123	5.323	518.650	500.000	3.7
Aroclor-1260-1	6.264	6.169	6.369	480.160	500.000	-4.0
Aroclor-1260-2	6.453	6.358	6.558	487.570	500.000	-2.5
Aroclor-1260-3	6.609	6.514	6.714	474.290	500.000	-5.1
Aroclor-1260-4	7.085	6.990	7.190	481.750	500.000	-3.7
Aroclor-1260-5	7.327	7.232	7.432	488.750	500.000	-2.3
Decachlorobiphenyl	8.746	8.651	8.851	50.870	50.000	1.7
Tetrachloro-m-xylene	3.661	3.565	3.765	48.470	50.000	-3.1