



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/06/2023 04/06/2023

Client Sample No.: CCAL03 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PP056733.D Time Analyzed: 17:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.608	5.512	5.712	488.310	500.000	-2.3
Aroclor-1016-2	5.630	5.535	5.735	492.840	500.000	-1.4
Aroclor-1016-3	5.693	5.598	5.798	491.900	500.000	-1.6
Aroclor-1016-4	5.792	5.697	5.897	489.890	500.000	-2.0
Aroclor-1016-5	6.089	5.994	6.194	493.460	500.000	-1.3
Aroclor-1260-1	7.223	7.128	7.328	499.560	500.000	-0.1
Aroclor-1260-2	7.481	7.386	7.586	505.310	500.000	1.1
Aroclor-1260-3	7.844	7.747	7.947	500.760	500.000	0.2
Aroclor-1260-4	8.070	7.975	8.175	514.070	500.000	2.8
Aroclor-1260-5	8.396	8.301	8.501	517.420	500.000	3.5
Decachlorobiphenyl	10.267	10.175	10.375	52.000	50.000	4.0
Tetrachloro-m-xylene	4.433	4.337	4.537	48.690	50.000	-2.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/06/2023 04/06/2023

Client Sample No.: CCAL03 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PP056733.D Time Analyzed: 17:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.761	4.666	4.866	477.410	500.000	-4.5
Aroclor-1016-2	4.781	4.685	4.885	488.020	500.000	-2.4
Aroclor-1016-3	4.959	4.864	5.064	493.600	500.000	-1.3
Aroclor-1016-4	5.001	4.906	5.106	455.870	500.000	-8.8
Aroclor-1016-5	5.218	5.123	5.323	482.160	500.000	-3.6
Aroclor-1260-1	6.263	6.169	6.369	481.830	500.000	-3.6
Aroclor-1260-2	6.452	6.358	6.558	490.840	500.000	-1.8
Aroclor-1260-3	6.609	6.514	6.714	486.240	500.000	-2.8
Aroclor-1260-4	7.084	6.990	7.190	490.140	500.000	-2.0
Aroclor-1260-5	7.326	7.232	7.432	484.730	500.000	-3.1
Decachlorobiphenyl	8.744	8.651	8.851	51.150	50.000	2.3
Tetrachloro-m-xylene	3.661	3.565	3.765	48.400	50.000	-3.2