



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060837.D Time Analyzed: 14:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.506	4.407	4.607	460.870	500.000	-7.8
Aroclor-1016-2	4.525	4.426	4.626	452.310	500.000	-9.5
Aroclor-1016-3	4.583	4.484	4.684	456.830	500.000	-8.6
Aroclor-1016-4	4.675	4.576	4.776	462.560	500.000	-7.5
Aroclor-1016-5	4.962	4.864	5.064	439.080	500.000	-12.2
Aroclor-1260-1	6.063	5.966	6.166	473.620	500.000	-5.3
Aroclor-1260-2	6.323	6.226	6.426	477.300	500.000	-4.5
Aroclor-1260-3	6.677	6.580	6.780	557.720	500.000	11.5
Aroclor-1260-4	6.894	6.797	6.997	525.550	500.000	5.1
Aroclor-1260-5	7.205	7.107	7.307	532.390	500.000	6.5
Decachlorobiphenyl	8.584	8.486	8.686	43.860	50.000	-12.3
Tetrachloro-m-xylene	3.406	3.307	3.507	43.820	50.000	-12.4



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060837.D Time Analyzed: 14:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.767	3.670	3.870	438.080	500.000	-12.4
Aroclor-1016-2	3.783	3.685	3.885	428.210	500.000	-14.4
Aroclor-1016-3	3.943	3.846	4.046	433.280	500.000	-13.3
Aroclor-1016-4	3.993	3.895	4.095	425.530	500.000	-14.9
Aroclor-1016-5	4.188	4.090	4.290	427.260	500.000	-14.5
Aroclor-1260-1	5.182	5.086	5.286	432.680	500.000	-13.5
Aroclor-1260-2	5.372	5.276	5.476	433.570	500.000	-13.3
Aroclor-1260-3	5.514	5.418	5.618	427.280	500.000	-14.5
Aroclor-1260-4	5.976	5.881	6.081	486.000	500.000	-2.8
Aroclor-1260-5	6.222	6.127	6.327	486.870	500.000	-2.6
Decachlorobiphenyl	7.540	7.446	7.646	40.950	50.000	-18.1
Tetrachloro-m-xylene	2.764	2.664	2.864	41.060	50.000	-17.9