



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL01 Date Analyzed: 04/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060871.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.510	4.407	4.607	488.590	500.000	-2.3
Aroclor-1016-2	4.529	4.426	4.626	482.240	500.000	-3.6
Aroclor-1016-3	4.587	4.484	4.684	485.230	500.000	-3.0
Aroclor-1016-4	4.678	4.576	4.776	490.210	500.000	-2.0
Aroclor-1016-5	4.966	4.864	5.064	465.490	500.000	-6.9
Aroclor-1260-1	6.068	5.966	6.166	493.150	500.000	-1.4
Aroclor-1260-2	6.328	6.226	6.426	501.730	500.000	0.3
Aroclor-1260-3	6.682	6.580	6.780	583.140	500.000	16.6
Aroclor-1260-4	6.899	6.797	6.997	552.650	500.000	10.5
Aroclor-1260-5	7.211	7.107	7.307	563.030	500.000	12.6
Decachlorobiphenyl	8.591	8.486	8.686	49.420	50.000	-1.2
Tetrachloro-m-xylene	3.410	3.307	3.507	48.160	50.000	-3.7



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL01 Date Analyzed: 04/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060871.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.766	3.670	3.870	466.100	500.000	-6.8
Aroclor-1016-2	3.781	3.685	3.885	460.620	500.000	-7.9
Aroclor-1016-3	3.941	3.846	4.046	461.000	500.000	-7.8
Aroclor-1016-4	3.991	3.895	4.095	451.220	500.000	-9.8
Aroclor-1016-5	4.186	4.090	4.290	455.480	500.000	-8.9
Aroclor-1260-1	5.180	5.086	5.286	455.940	500.000	-8.8
Aroclor-1260-2	5.370	5.276	5.476	465.880	500.000	-6.8
Aroclor-1260-3	5.512	5.418	5.618	450.260	500.000	-9.9
Aroclor-1260-4	5.974	5.881	6.081	514.100	500.000	2.8
Aroclor-1260-5	6.221	6.127	6.327	510.310	500.000	2.1
Decachlorobiphenyl	7.541	7.446	7.646	42.130	50.000	-15.7
Tetrachloro-m-xylene	2.762	2.664	2.864	44.540	50.000	-10.9