



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL02 Date Analyzed: 04/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060882.D Time Analyzed: 12:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.505	4.407	4.607	448.010	500.000	-10.4
Aroclor-1016-2	4.524	4.426	4.626	454.980	500.000	-9.0
Aroclor-1016-3	4.582	4.484	4.684	463.660	500.000	-7.3
Aroclor-1016-4	4.674	4.576	4.776	464.580	500.000	-7.1
Aroclor-1016-5	4.961	4.864	5.064	443.490	500.000	-11.3
Aroclor-1260-1	6.063	5.966	6.166	476.990	500.000	-4.6
Aroclor-1260-2	6.323	6.226	6.426	476.860	500.000	-4.6
Aroclor-1260-3	6.676	6.580	6.780	560.730	500.000	12.1
Aroclor-1260-4	6.893	6.797	6.997	528.990	500.000	5.8
Aroclor-1260-5	7.204	7.107	7.307	542.770	500.000	8.6
Decachlorobiphenyl	8.584	8.486	8.686	47.770	50.000	-4.5
Tetrachloro-m-xylene	3.406	3.307	3.507	45.090	50.000	-9.8



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL02 Date Analyzed: 04/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060882.D Time Analyzed: 12:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.766	3.670	3.870	446.710	500.000	-10.7
Aroclor-1016-2	3.782	3.685	3.885	441.890	500.000	-11.6
Aroclor-1016-3	3.942	3.846	4.046	443.870	500.000	-11.2
Aroclor-1016-4	3.991	3.895	4.095	430.980	500.000	-13.8
Aroclor-1016-5	4.186	4.090	4.290	438.510	500.000	-12.3
Aroclor-1260-1	5.181	5.086	5.286	439.300	500.000	-12.1
Aroclor-1260-2	5.371	5.276	5.476	471.660	500.000	-5.7
Aroclor-1260-3	5.512	5.418	5.618	430.360	500.000	-13.9
Aroclor-1260-4	5.975	5.881	6.081	495.190	500.000	-1.0
Aroclor-1260-5	6.221	6.127	6.327	494.970	500.000	-1.0
Decachlorobiphenyl	7.540	7.446	7.646	41.580	50.000	-16.8
Tetrachloro-m-xylene	2.762	2.664	2.864	42.350	50.000	-15.3