



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL02 Date Analyzed: 04/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060894.D Time Analyzed: 14:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.506	4.407	4.607	472.360	500.000	-5.5
Aroclor-1016-2	4.525	4.426	4.626	468.380	500.000	-6.3
Aroclor-1016-3	4.583	4.484	4.684	471.320	500.000	-5.7
Aroclor-1016-4	4.675	4.576	4.776	476.470	500.000	-4.7
Aroclor-1016-5	4.962	4.864	5.064	456.480	500.000	-8.7
Aroclor-1260-1	6.063	5.966	6.166	493.090	500.000	-1.4
Aroclor-1260-2	6.324	6.226	6.426	514.710	500.000	2.9
Aroclor-1260-3	6.677	6.580	6.780	576.240	500.000	15.2
Aroclor-1260-4	6.894	6.797	6.997	548.110	500.000	9.6
Aroclor-1260-5	7.206	7.107	7.307	560.440	500.000	12.1
Decachlorobiphenyl	8.585	8.486	8.686	49.510	50.000	-1.0
Tetrachloro-m-xylene	3.406	3.307	3.507	46.150	50.000	-7.7



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2023 03/27/2023

Client Sample No.: CCAL02 Date Analyzed: 04/10/2023

Lab Sample No.: AR1660CCC500 Data File : PQ060894.D Time Analyzed: 14:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.768	3.670	3.870	467.970	500.000	-6.4
Aroclor-1016-2	3.783	3.685	3.885	462.350	500.000	-7.5
Aroclor-1016-3	3.943	3.846	4.046	467.100	500.000	-6.6
Aroclor-1016-4	3.993	3.895	4.095	461.330	500.000	-7.7
Aroclor-1016-5	4.188	4.090	4.290	463.400	500.000	-7.3
Aroclor-1260-1	5.182	5.086	5.286	462.320	500.000	-7.5
Aroclor-1260-2	5.372	5.276	5.476	463.380	500.000	-7.3
Aroclor-1260-3	5.514	5.418	5.618	451.490	500.000	-9.7
Aroclor-1260-4	5.976	5.881	6.081	518.040	500.000	3.6
Aroclor-1260-5	6.223	6.127	6.327	516.070	500.000	3.2
Decachlorobiphenyl	7.542	7.446	7.646	43.950	50.000	-12.1
Tetrachloro-m-xylene	2.763	2.664	2.864	43.930	50.000	-12.1