



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/07/2023 04/07/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PR060737.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.922	4.823	5.023	515.650	500.000	3.1
Aroclor-1016-2	4.945	4.846	5.046	510.340	500.000	2.1
Aroclor-1016-3	5.009	4.911	5.111	513.480	500.000	2.7
Aroclor-1016-4	5.112	5.014	5.214	515.210	500.000	3.0
Aroclor-1016-5	5.431	5.333	5.533	507.960	500.000	1.6
Aroclor-1260-1	6.653	6.556	6.756	511.840	500.000	2.4
Aroclor-1260-2	6.938	6.840	7.040	510.380	500.000	2.1
Aroclor-1260-3	7.331	7.233	7.433	517.120	500.000	3.4
Aroclor-1260-4	7.576	7.477	7.677	516.290	500.000	3.3
Aroclor-1260-5	7.916	7.817	8.017	523.640	500.000	4.7
Decachlorobiphenyl	9.660	9.561	9.761	51.200	50.000	2.4
Tetrachloro-m-xylene	3.688	3.588	3.788	52.290	50.000	4.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/07/2023 04/07/2023

Client Sample No.: CCAL02 Date Analyzed: 04/07/2023

Lab Sample No.: AR1660CCC500 Data File : PR060737.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.009	3.909	4.109	482.180	500.000	-3.6
Aroclor-1016-2	4.027	3.927	4.127	487.710	500.000	-2.5
Aroclor-1016-3	4.204	4.104	4.304	480.530	500.000	-3.9
Aroclor-1016-4	4.255	4.155	4.355	475.120	500.000	-5.0
Aroclor-1016-5	4.471	4.371	4.571	475.470	500.000	-4.9
Aroclor-1260-1	5.555	5.455	5.655	484.670	500.000	-3.1
Aroclor-1260-2	5.760	5.660	5.860	475.190	500.000	-5.0
Aroclor-1260-3	5.917	5.817	6.017	488.010	500.000	-2.4
Aroclor-1260-4	6.421	6.321	6.521	490.040	500.000	-2.0
Aroclor-1260-5	6.687	6.587	6.787	501.050	500.000	0.2
Decachlorobiphenyl	8.129	8.028	8.228	50.320	50.000	0.6
Tetrachloro-m-xylene	2.901	2.800	3.000	48.460	50.000	-3.1