

Extraction Type:

Date Collected:

Date Received:

SDG No.:

Final Vol:

Injection Volume:

Test:

Matrix: % Solid:

04/05/23

04/05/23

O2213

SOIL

100

10000

PCB

Decanted:

иL

Report of Analysis

Client: Chemtech Consulting Group

Project: LOD-LOQ Study

Client Sample ID: LOD-MDL-SOIL-01-QT2-2023

Lab Sample ID: O2213-01

Analytical Method: SW8082A

Sample Wt/Vol: 30.04 Units: g

Soil Aliquot Vol: uL

PH:

• •

GPC Factor: 1.0

Prep Method: SW3541B

 File ID/Qc Batch:
 Dilution:
 Prep Date
 Date Analyzed
 Prep Batch ID

 PR060722.D
 1
 04/06/23 10:10
 04/07/23 17:48
 PB151952

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.60	U	3.60	17.0	ug/kg
11104-28-2	Aroclor-1221	5.90	U	5.90	17.0	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	17.0	ug/kg
53469-21-9	Aroclor-1242	3.10	U	3.10	17.0	ug/kg
12672-29-6	Aroclor-1248	9.60	J	2.80	17.0	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	17.0	ug/kg
37324-23-5	Aroclor-1262	2.70	U	2.70	17.0	ug/kg
11100-14-4	Aroclor-1268	3.30	U	3.30	17.0	ug/kg
11096-82-5	Aroclor-1260	3.30	U	3.30	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.5		40 - 162	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.1		32 - 176	111%	

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit