



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_D Date(s) Analyzed: 03/14/2023 03/14/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.26	6.16	6.36	24634800
		2	7.17	7.07	7.27	60895500
		3	7.27	7.17	7.37	34353200
		4	7.60	7.50	7.70	43930700
		5	7.96	7.86	8.06	46974500



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_D Date(s) Analyzed: 03/14/2023 03/14/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.01	4.91	5.11	9219640
		2	5.69	5.59	5.79	9927610
		3	6.61	6.51	6.71	33104200
		4	6.74	6.64	6.84	42795300
		5	7.05	6.95	7.15	18683200