



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_L Date(s) Analyzed: 03/22/2023 03/22/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.00	5.90	6.10	16667300
		2	6.82	6.72	6.92	64715500
		3	6.90	6.80	7.00	45669700
		4	7.00	6.90	7.10	28095500
		5	7.32	7.22	7.42	33779500



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_L Date(s) Analyzed: 03/22/2023 03/22/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.20	5.10	5.30	10325400
		2	5.55	5.45	5.65	10972000
		3	5.83	5.73	5.93	14167600
		4	6.46	6.36	6.56	41175700
		5	6.59	6.49	6.69	58721500