



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_L Date(s) Analyzed: 05/19/2023 05/19/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.99	5.89	6.09	18290600
		2	6.80	6.70	6.90	69113900
		3	6.89	6.79	6.99	47819000
		4	6.99	6.89	7.09	29436100
		5	7.31	7.21	7.41	33814800



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG NO.: O2213

Instrument ID: ECD_L Date(s) Analyzed: 05/19/2023 05/19/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.17	5.07	5.27	11128400
		2	5.52	5.42	5.62	12117100
		3	5.81	5.71	5.91	14563600
		4	6.43	6.33	6.53	46428000
		5	6.56	6.46	6.66	65148200