



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG No.: O2213

Calibration Sequence : FE040523		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	222897509	131116	FE041180.D
850	114935778	135219	FE041181.D
340	46493183	136745	FE041182.D
170	25719098	151289	FE041183.D
85	12754478	150053	FE041184.D
AVG RF : 140884		% RSD : 6.515	AVG RT : 14.942