



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG No.: O2213

Calibration Sequence : FF031623		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	197266747	116039	FF012284.D
850	93848942	110411	FF012285.D
340	38322160	112712	FF012286.D
170	20527551	120750	FF012287.D
85	9761445	114841	FF012288.D
AVG RF : 114951		% RSD : 3.382	AVG RT : 15.0264