



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O2213 SAS No.: O2213 SDG No.: O2213

Calibration Sequence : FG031523		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	166574937	97985	FG011595.D
850	87230574	102624	FG011596.D
340	34060798	100179	FG011597.D
170	19171505	112774	FG011598.D
85	9573149	112625	FG011599.D
AVG RF : 105237		% RSD : 6.658	AVG RT : 15.0538